

RPR Materials Provided

MATERIALS	25 TEST KIT	100 TEST KIT	500 TEST KIT*	5000 TEST KIT*	10000 TEST KIT*
RPR Carbon Antigen	0.5 ml	2.0 ml	9.0 ml	10 x 9.0 ml	20 x 9.0 ml
Reactive Control	0.5 ml	1.0 ml	2.0 ml	5 x 2.0 ml	5.0 ml
Weak Reactive Control	0.5 ml	1.0 ml	2.0 ml	5 x 2.0 ml	5.0 ml
Nonreactive Control	0.5 ml	1.0 ml	2.0 ml	5 x 2.0 ml	5.0 ml
3 ml Dropping Bottle	1	1	1	10	10
20-ga Dispensing Needle (60 drops/ml)	1	1	1	10	10
RPR Test Card (10-well)	3	10	50	500	n/a
RPR Test Card (30-well)	n/a	n/a	n/a	175	350
0.05 ml Disposable Stirrer Pipets	25	100	500	5000	10000

* Also available without controls

ADDITIONAL MATERIALS REQUIRED

- + Volumetric pipet to deliver 0.05 ml
- + Saline (0.9% NaCl Solution)
- + Serum nonreactive to syphilis, in 0.9% saline, for diluting specimens reactive at the 1:16 dilution in the semiquantitative procedure
- + Mechanical rotator set at 100 ± rpm and circumscribing 3/4 inch diameter, with humidity cover
- + Timing device, minute and second capability
- + Disposable syringe, 1 or 3 ml, accuracy of ± 5%