

DF New York One, Inc. - Jeeter
Order No.: ONYJTR1117-0010385
1050 Lower South St
New York, 10567
lpalazzo@jeeter.com
4424037621
OCM-PROC-24-000153

Sample: SNYJTR1117-PEIP-0021174
Strain: Maui Wowie, Unit Weight: 2.5000g
Batch#: DFNJ-MAUWOW-111025, Batch Size: 1968 Units, LOT ID:
111000000000334
Date Sampled: 11/17/2025 12:01
Sample Received: 11/17/2025 17:53
Report Created: 11/22/2025 17:51
Sampling SOP 204

Infused Baby Jeeter 0.5GX5 (2.5g) Preroll Maui Wowie
Plant, Enhanced/Infused Preroll



Results

538.25 mg/unit
THCa

505.82 mg/unit
D9-THC

ND
Total CBD

977.86 mg/unit
Total THC


1,056.40 mg/unit
Total Cannabinoids

1.92%
Total Terpenes

Tests Summary

Cannabinoids Tested	Moisture Pass	Microbials Pass
Water Activity Pass	Foreign Matter Not Tested	Homogeneity Not Tested
Terpenes Pass	Residual Solvents Pass	Mycotoxins Pass
Heavy Metals Pass	Pesticides Pass	

Dominant Terpenes



DF New York One, Inc. - Jeeter

Order No.: ONYJTR1117-0010385
1050 Lower South St
New York, 10567
lpalazzo@jeeter.com
4424037621
OCM-PROC-24-000153

Sample: SNYJTR1117-PEIP-0021174

Strain: Maui Wowie, Unit Weight: 2.5000g
Batch#: DFNY-MAUWOW-111025, Batch Size: 1968 Units, LOT ID:
111000000000334
Date Sampled: 11/17/2025 12:01
Sample Received: 11/17/2025 17:53
Report Created: 11/22/2025 17:51
Sampling SOP 204

Infused Baby Jeeter 0.5GX5 (2.5g) Preroll Maui Wowie
Plant, Enhanced/Infused Preroll



Cannabinoids

SOP: SOP 801-NY

Instrument: Agilent HPLC 2518 -
2518

Date/Time Tested: 11/18/2025
10:40

Analyte	LOQ (ug/mL)	%	mg/g	mg/unit
CBDV	4,151.74	ND	ND	ND
CBDa	4,055.34	ND	ND	ND
CBGa	4,055.34	< LOQ	< LOQ	< LOQ
CBG	4,055.34	0.49	4.93	12.33
CBD	4,055.34	ND	ND	ND
THCV	4,151.74	< LOQ	< LOQ	< LOQ
CBN	4,366.44	< LOQ	< LOQ	< LOQ
D9-THC	4,151.74	20.23	202.33	505.82
D8-THC	4,366.44	ND	ND	ND
(6aR,9S)-d10-THC	4,366.44	ND	ND	ND
(6aR,9R)-d10-THC	4,055.34	ND	ND	ND
CBC	4,151.74	< LOQ	< LOQ	< LOQ
THCa	4,366.44	21.53	215.30	538.25
Total CBD			ND	ND
Total THC		39.11	391.14	977.86
Total Cannabinoids		42.26	422.56	1,056.40

Report notes:

Total THC = THCa * 0.877 + Δ8-THC + Δ9-THC + (6aR,9S)-d10-THC + (6aR,9R)-d10-THC
Total CBD = CBDa * 0.877 + CBDT
Total Cannabinoids = Sum of all cannabinoids
LOQ = Limit of Quantitation; The reported result is based on a sample weight with the applicable moisture content for that sample; Unless otherwise stated all quality control samples performed within specifications established by the Laboratory. OCMPPCL-2022-00001. Cannabinoid potency values for flower type products are reported by percentage of dry weight determined via loss on drying; Unless otherwise stated all quality control samples performed within specifications established by the Laboratory. All results were generated by ISO certified methods to full state testing requirements. ND = Not Detected; NT = Not Tested; NR = Not Reported

Tested

Microbials

SOP: SOP 401-NY SOP 418-NY

Instrument: BioRad CFX qPCR -
3195 / 35C Incubator - 1032 / 28C
Incubator - 1034

Date/Time Tested: 11/18/2025
14:39

Analyte	LOQ (CFU/g)	Limit (CFU/g)	CFU/g	Status
Aerobic Bacteria	100		ND	Tested
E. Coli		0	ND	Passed
Yeast & Mold	100		3,984	Tested
Salmonella		0	ND	Passed
Aspergillus Flavus		0	ND	Passed
Aspergillus Fumigatus		0	ND	Passed
Aspergillus Niger		0	ND	Passed
Aspergillus Terreus		0	ND	Passed

Report notes:

Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. ND = Not Detected; NT = Not Tested; NR = Not Reported

Pass

Water Activity

SOP: SOP 206-NY

Date/Time Tested: 11/19/2025
13:32

Analyte	Limit (aw)	aw	Status
Water Activity	0.65	0.51	Passed

Pass

Moisture

SOP: SOP 205-NY

Instrument: Shimadzu Moisture
Meter - 2560

Date/Time Tested: 11/19/2025
13:38

Analyte	Limit (%)	%	Status
Moisture	15	6.4	Passed

Pass



DF New York One, Inc. - Jeeter

Order No.: ONYJTR1117-0010385

1050 Lower South St

New York, 10567

lpalazzo@jeeter.com

4424037621

OCM-PROC-24-000153

Sample: SNYJTR1117-PEIP-0021174

Strain: Maui Wowie, Unit Weight: 2.5000g

Batch#: DFNY-MAUWOW-111025, Batch Size: 1968 Units, LOT ID:

11100000000334

Date Sampled: 11/17/2025 12:01

Sample Received: 11/17/2025 17:53

Report Created: 11/22/2025 17:51

Sampling SOP 204

Infused Baby Jeeter 0.5GX5 (2.5g) Preroll Maui Wowie

Plant, Enhanced/Infused Preroll



Terpenes

SOP: SOP 620-NY

Instrument: Agilent GC-FID 1950 - 1950

Date/Time Tested: 11/19/2025 14:34

Pass

Analyte	LOQ (ug/mL)	Limit (ug/mL)	%	Status	Analyte	LOQ (ug/mL)	Limit (ug/mL)	%	Status
Total Terpenes		115,000	1.9202	Passed	Borneol	80		0.0087	Tested
Limonene	80		0.6972	Tested	Camphene	80		< LOQ	Tested
b-Caryophyllene	80		0.3327	Tested	Sabinene	80		ND	Tested
Linalool	80		0.2374	Tested	a-Phellandrene	80		ND	Tested
b-Myrcene	80		0.1025	Tested	d-3-Carene	80		ND	Tested
a-Humulene	80		0.0848	Tested	a-Terpinene	80		ND	Tested
trans-b-Farnesene	80		0.0609	Tested	p-Cymene	80		ND	Tested
Guaiol	80		0.0545	Tested	Eucalyptol	80		ND	Tested
Terpineol	80		0.0481	Tested	Sabinene Hydrate	80		ND	Tested
a-Bisabolol	80		0.0439	Tested	Fenchone	80		ND	Tested
Fenchol	80		0.0386	Tested	Isopulegol	80		ND	Tested
b-Pinene	80		0.0385	Tested	Camphor	80		ND	Tested
cis-Nerolidol	80		0.0378	Tested	Isoborneol	80		ND	Tested
Geraniol	80		0.0329	Tested	DL-Menthol	80		ND	Tested
g-Terpinene	80		0.0238	Tested	Nerol	80		ND	Tested
a-Pinene	80		0.0198	Tested	Geranyl Acetate	80		ND	Tested
Caryophyllene Oxide	80		0.0173	Tested	a-Cedrene	80		ND	Tested
trans-Nerolidol	80		0.0170	Tested	Valencene	80		ND	Tested
Terpinolene	80		0.0135	Tested	Cedrol	80		ND	Tested
Pulegone	80		0.0103	Tested	trans-b-Ocimene	80		ND	Tested

Report notes:

LOQ = Limit of Quantitation. Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. ND = Not Detected; NT = Not Tested; NR = Not Reported

DF New York One, Inc. - Jeeter

Order No.: ONYJTR1117-0010385

1050 Lower South St

New York, 10567

lpalazzo@jeeter.com

4424037621

OCM-PROC-24-000153

Sample: SNYJTR1117-PEIP-0021174

Strain: Maui Wowie, Unit Weight: 2.5000g

Batch#: DFNJ-MAUWOW-111025, Batch Size: 1968 Units, LOT ID:

111000000000334

Date Sampled: 11/17/2025 12:01

Sample Received: 11/17/2025 17:53

Report Created: 11/22/2025 17:51

Sampling SOP 204

Infused Baby Jeeter 0.5GX5 (2.5g) Preroll Maui Wowie

Plant, Enhanced/Infused Preroll



Residual Solvents

SOP: SOP 612-NY

Instrument: Agilent HS-GC-MS 1333 - 1333

Date/Time Tested: 11/19/2025 13:18

Pass

Analyte	LOQ (ug/g)	Limit (ug/g)	ug/g	Status
1,2-Dichloroethane	3	5	ND	Passed
Acetone	2,500	5,000	< LOQ	Passed
Acetonitrile	205	410	ND	Passed
Benzene	1	2	ND	Passed
Butane	2,500	5,000	ND	Passed
Chloroform	30	60	ND	Passed
Ethanol	2,500	5,000	< LOQ	Passed
Ethyl Acetate	2,500	5,000	ND	Passed
Ethyl Ether	2,500	5,000	ND	Passed
DMSO	2,500	5,000	ND	Passed
Heptane	2,500	5,000	ND	Passed
Hexanes	145	290	ND	Passed
Isopropyl Alcohol	2,500	5,000	< LOQ	Passed
Methanol	1,500	3,000	< LOQ	Passed
Methylene Chloride	300	600	ND	Passed
Pentanes	2,500	5,000	ND	Passed
Propane	2,500	5,000	ND	Passed
Toluene	445	890	ND	Passed
Trichloroethane	750	1,500	ND	Passed
Xylenes	1,085	2,170	ND	Passed
1,1,1,2-Tetrafluoroethane (HFC-134a)	500	1,000	ND	Passed

Report notes:

LOQ = Limit of Quantitation; The reported result is based on a sample weight with the applicable moisture content for that sample; Unless otherwise stated all quality control samples performed within specifications established by the Laboratory. All results were generated by ISO certified methods to full state testing requirements. If DMSO and 1,1,1-Trichloroethane are reported, they are tentatively identified, but not quantitatively confirmed. ND = Not Detected; NT = Not Tested; NR = Not Reported.

DF New York One, Inc. - Jeeter

Order No.: ONYJTR1117-0010385

1050 Lower South St

New York, 10567

lpalazzo@jeeter.com

4424037621

OCM-PROC-24-000153

Sample: SNYJTR1117-PEIP-0021174

Strain: Maui Wowie, Unit Weight: 2.5000g

Batch#: DFNY-MAUWOW-111025, Batch Size: 1968 Units, LOT ID:

111000000000334

Date Sampled: 11/17/2025 12:01

Sample Received: 11/17/2025 17:53

Report Created: 11/22/2025 17:51

Sampling SOP 204

Infused Baby Jeeter 0.5GX5 (2.5g) Preroll Maui Wowie

Plant, Enhanced/Infused Preroll



Mycotoxins

SOP: SOP 808-NY

Instrument: Agilent LC-MS-MS 2755 - 2755

Date/Time Tested: 11/19/2025 17:36

Pass

Analyte	LOQ (ng/g)	Limit (ng/g)	ng/g	Status
B1	4.8		ND	Tested
B2	4.8		ND	Tested
G1	4.8		ND	Tested
G2	4.8		ND	Tested
Ochratoxin A	4.8	20.0	ND	Passed
Total Aflatoxins		20.0	ND	Passed
Total Mycotoxins			ND	Tested

Report notes:

LOQ = Limit of Quantitation. Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. ND = Not Detected; NT = Not Tested; NR = Not Reported

Heavy Metals

SOP: SOP 250-NY

Instrument: Agilent 7800 ICP-MS - 2594

Date/Time Tested: 11/19/2025 14:46

Pass

Analyte	LOQ (ug/g)	Limit (ug/g)	ug/g	Status
Antimony	0.181	2.000	ND	Passed
Arsenic	0.181	0.200	ND	Passed
Cadmium	0.181	0.200	ND	Passed
Chromium	0.181	110.000	< LOQ	Passed
Copper	0.217	30.000	5.731	Passed
Mercury	0.043	0.100	ND	Passed
Nickel	0.217	5.000	< LOQ	Passed
Lead	0.181	0.500	ND	Passed

Report notes:

LOQ = Limit of Quantitation. Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. ND = Not Detected; NT = Not Tested; NR = Not Reported

DF New York One, Inc. - Jeeter

Order No.: ONYJTR1117-0010385
 1050 Lower South St
 New York, 10567
 lpalazzo@jeeter.com
 4424037621
 OCM-PROC-24-000153

Sample: SNYJTR1117-PEIP-0021174

Strain: Maui Wowie, Unit Weight: 2.5000g
 Batch#: DFNJ-MAUWOW-111025, Batch Size: 1968 Units, LOT ID:
 111000000000334
 Date Sampled: 11/17/2025 12:01
 Sample Received: 11/17/2025 17:53
 Report Created: 11/22/2025 17:51
 Sampling SOP 204

Infused Baby Jeeter 0.5GX5 (2.5g) Preroll Maui Wowie
 Plant, Enhanced/Infused Preroll



Pesticides

SOP: SOP 814-NY

Instrument: Agilent LCMS 2546 - 2546 / Agilent GCMS 2541 - 2541

Date/Time Tested: 11/19/2025 12:39

Pass

Analyte	LOQ (ug/g)	Limit (ug/g)	ug/g	Status
Abamectin	0.38	0.50	ND	Passed
Acephate	0.10	0.40	ND	Passed
Acequinocyl	0.10	2.00	ND	Passed
Acetamiprid	0.10	0.20	ND	Passed
Aldicarb	0.10	0.40	ND	Passed
Azoxystrobin	0.10	0.20	ND	Passed
Bifenazate	0.10	0.20	ND	Passed
Bifenthrin	0.10	0.20	ND	Passed
Boscalid	0.10	0.40	ND	Passed
Carbaryl	0.10	0.20	ND	Passed
Carbofuran	0.10	0.20	ND	Passed
Chlorantraniliprole	0.10	0.20	ND	Passed
Chlorpyrifos	0.10	0.20	ND	Passed
Clofentezine	0.10	0.20	ND	Passed
Coumaphos	0.10	1.00	ND	Passed
Cyfluthrin	0.48	1.00	ND	Passed
Cypermethrin	0.10	1.00	ND	Passed
Daminozide	0.10	1.00	ND	Passed
Diazinon	0.10	0.20	ND	Passed
Dichlorvos	0.10	1.00	ND	Passed
Dimethoate	0.10	0.20	ND	Passed
Dimethomorph	0.10	1.00	ND	Passed
Ethoprophos	0.10	0.20	ND	Passed
Etofenprox	0.10	0.40	ND	Passed
Etoxazole	0.10	0.20	ND	Passed
Fenhexamid	0.10	1.00	ND	Passed
Fenoxycarb	0.10	0.20	ND	Passed
Fenpyroximate	0.10	0.40	ND	Passed
Fipronil	0.10	0.40	ND	Passed
Flonicamid	0.10	1.00	ND	Passed
Fludioxonil	0.10	0.40	ND	Passed
Hexythiazox	0.10	1.00	ND	Passed
Imazalil	0.10	0.20	ND	Passed
Imidacloprid	0.10	0.40	ND	Passed
Indole-3 Butyric Acid	0.12	1.00	ND	Passed
Kresoxim Methyl	0.10	0.40	ND	Passed
Malathion	0.10	0.20	ND	Passed
Mevinphos	0.10	1.00	ND	Passed
Metalaxyl	0.10	0.20	ND	Passed
Methiocarb	0.10	0.20	ND	Passed

DF New York One, Inc. - Jeeter

Order No.: ONYJTR1117-0010385

1050 Lower South St

New York, 10567

lpalazzo@jeeter.com

4424037621

OCM-PROC-24-000153

Sample: SNYJTR1117-PEIP-0021174

Strain: Maui Wowie, Unit Weight: 2.5000g

Batch#: DFNJ-MAUWOW-111025, Batch Size: 1968 Units, LOT ID:

111000000000334

Date Sampled: 11/17/2025 12:01

Sample Received: 11/17/2025 17:53

Report Created: 11/22/2025 17:51

Sampling SOP 204

Infused Baby Jeeter 0.5GX5 (2.5g) Preroll Maui Wowie

Plant, Enhanced/Infused Preroll



Analyte	LOQ (ug/g)	Limit (ug/g)	ug/g	Status
Methomyl	0.10	0.40	ND	Passed
MGK-264	0.01	0.20	ND	Passed
Myclobutanil	0.10	0.20	ND	Passed
Naled	0.10	0.50	ND	Passed
Oxamyl	0.10	1.00	ND	Passed
Paclobutrazol	0.10	0.40	ND	Passed
Permethrin	0.10	0.20	ND	Passed
Phosmet	0.10	0.20	ND	Passed
Piperonyl Butoxide	0.10	2.00	ND	Passed
Prallethrin	0.10	0.20	ND	Passed
Propiconazole	0.10	0.40	ND	Passed
Propoxur	0.10	0.20	ND	Passed
Pyrethrins	0.07	1.00	ND	Passed
Pyridaben	0.10	0.20	ND	Passed
Spinetoram	0.10	1.00	ND	Passed
Spinosyn AD	0.10	0.20	ND	Passed
Spiromesifen	0.10	0.20	ND	Passed
Spirotetramat	0.10	0.20	ND	Passed
Spiroxamine	0.10	0.20	ND	Passed
Tebuconazole	0.10	0.40	ND	Passed
Thiacloprid	0.10	0.20	ND	Passed
Thiamethoxam	0.10	0.20	ND	Passed
Trifloxystrobin	0.10	0.20	ND	Passed
Captan		1.00	TIC	Passed
Methyl Parathion	0.10	0.20	ND	Passed
Chlordane	0.10	1.00	ND	Passed
Chlorfenapyr	0.10	1.00	ND	Passed
PCNB	0.10	1.00	ND	Passed
Azadirachtin	0.12	1.00	ND	Passed
Chlormequat Chloride	0.02	1.00	ND	Passed

Report notes:

LOQ = Limit of Quantitation. Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. If captan, chlormequat chloride, or MGK-264 are reported, they are tentatively identified, but not quantitatively confirmed. ND = Not Detected; NT = Not Tested; NR = Not Reported. "TIC" means tentatively identified, but not quantitatively confirmed.