

HCR™ Pro RNA-ISH Setup Guide for the BOND RX/RX™

This Setup Guide demonstrates the use of an HCR™ Pro RNA-ISH kit on the BOND RX/RX™ platforms from Leica Biosystems. Reagent preparation steps, including registering individual BOND Containers for their respective reagents, will be described in further detail. Each BOND RX/RX™ run takes approximately 7-10.5 hours followed by a short post-processing of stained slides. This time range depends on the number of slides processed in a single run and whether you're using the BOND RX or the BOND RX™ platform. Processing 1 slide on the BOND RX takes around 7 hours, and processing 30 slides on the BOND RX takes around 9 hours due to the time the BOND RX/RX™ platforms take to disperse the reagents onto each slide. The HCR™ Pro RNA-ISH kit can be used to probe and visualize RNA transcripts in FFPE tissue sections. Please read through the Setup Guide for additional information so that you can easily incorporate the HCR™ Pro RNA-ISH kit into your current workflow.

Table of Contents

HCR™ Pro RNA-ISH Kit Information	3
<i>HCR™ Pro RNA-ISH Starter Kit</i>	3
<i>HCR™ Pro RNA-ISH Kit</i>	5
<i>Required Materials for the BOND RX/RX™</i>	6
<i>Required Materials for Running Matisse® Chromogens on BOND RX/RX™</i>	7
<i>User-Supplied Materials</i>	8
Overall Workflow of the HCR™ Pro RNA-ISH Protocol	9
Set Up BOND RX/RX™ Protocols	10
<i>Register BOND Containers.....</i>	10
<i>Create Staining Protocols for HCR™ Pro RNA-CISH</i>	13
<i>Register HCR™ HiFi Probe Reagents.....</i>	14
<i>Create a Staining Protocol for HCR™ Pro RNA-FISH</i>	16
<i>Post-Processing for AP-Based Detection</i>	18
<i>Post-Processing for HRP-Based Detection.....</i>	18
<i>Post-Processing for Fluorescent Detection.....</i>	18
Reagent Preparation	19
<i>Reagent Transfer and Storage</i>	19
Overall Workflow of the HCR™ Pro RNA-ISH + IHC/IF Protocol	21
Creating an HCR™ Pro RNA-ISH + IHC/IF Co-Detection Protocol	22
<i>Remove Hematoxylin from the HCR™ Pro RNA-CISH Protocol.....</i>	22
<i>Create an IHC/IF Protocol from an Existing IHC Protocol.....</i>	23
<i>Set Up Slides with Three Sequential Staining Protocols</i>	25
Appendix.....	26
<i>Appendix A: Detailed Staining Protocol for HCR™ RNA-Chr ISH Seq_A [Chromogenic Detection]</i>	26
<i>Appendix B: Detailed Staining Protocol for HCR™ RNA-Chr ISH HRP_B and HCR™ RNA-Chr ISH AP_B [Chromogenic Detection]</i>	27
<i>Appendix C: Detailed Staining Protocol for HCR™ RNA-FISH_B [Fluorescent Detection].....</i>	30

Appendix D: Detailed IHC HRP Detection Staining Protocol for HCRTM Pro RNA-CISH + IHC Co-Detection 32

Appendix E: Detailed IHC AP Detection Staining Protocol for HCRTM Pro RNA-CISH + IHC Co-Detection 34

Appendix F: Detailed IF Detection Staining Protocol for HCRTM Pro RNA-FISH + IF Co-Detection 36

Appendix G: Third-Party Recommended Tyramide Dyes 37

Appendix H: Create a Staining Protocol using Matisse[®] Chromogens..... 38

Appendix I: Tips & Tricks..... 42

Appendix J: FAQ..... 44

HCR™ Pro RNA-ISH Kit Information

Upon receiving an HCR™ Pro RNA-ISH kit, please check all reagents and their storage conditions.

HCR™ Pro RNA-ISH Starter Kit

The HCR™ Pro RNA-ISH Starter Kit is provided with enough reagents to perform the assay on 20 slides and includes two HCR™ HiFi Probes and HCR™ Pro Detect reagents. The HCR™ HiFi Probes included in the Starter Kit (*Ppib/dapB*) are provided in volumes sufficient enough to perform the assay on 10 slides each. The Starter Kit also contains an HCR™ Membrane Stain to be used in an HCR™ Pro RNA-ISH + IHC/IF co-detection assay.

HCR™ HiFi Probes

HCR™ Reagents	Amount for an HCR™ Pro RNA-ISH Starter Kit	Storage Temperature
HCR™ HiFi Probe 1A <i>PPIB/Ppib</i> ¹ – Positive Control	4 mL	2 to 8 °C
HCR™ HiFi Probe 1B <i>PPIB/Ppib</i> ¹ – Positive Control	4 mL	2 to 8 °C
HCR™ HiFi Probe 1A <i>dapB</i> – Negative Control	4 mL	2 to 8 °C
HCR™ HiFi Probe 1B <i>dapB</i> – Negative Control	4 mL	2 to 8 °C
HCR™ Membrane Stain ²	3 mL	2 to 8 °C
HCR™ Control Slides ³	2 Slides	2 to 8 °C

¹Upper and lower cases are used to denote human and mouse HCR™ HiFi Probes respectively.

²The HCR™ Membrane Stain's host species is in rabbit and is provided in a volume sufficient to perform the assay on 5 slides. Please reference pages 21-25 for more information on how to perform an HCR™ Pro RNA-ISH + IHC/IF co-detection assay.

³Please allocate one slide for the positive control and one slide for the negative control.

HCR™ Pro Detect

Please note that all HCR™ Pro Detect reagents are provided in amber bottles, as these reagents are sensitive to light.

HCR™ Reagents	Amount for an HCR™ Pro RNA-ISH Starter Kit	Storage Temperature
Pretreat	5 mL	2 to 8 °C
HCR™ Pro Detect A	5 mL	2 to 8 °C
HCR™ Pro Detect B	7 mL	2 to 8 °C
HCR™ Pro Detect C	5 mL	2 to 8 °C
HCR™ Pro Detect D	7 mL	2 to 8 °C
HCR™ Pro Detect E	7 mL	2 to 8 °C
HCR™ Pro Detect F AP/HRP ¹	7 mL	2 to 8 °C
Post-Process A	5 mL	2 to 8 °C
Post-Process B	7 mL	2 to 8 °C

¹HCR™ Pro Detect F AP is included in the HCR™ Pro RNA-ISH AP Starter Kit, and HCR™ Pro Detect F HRP is included in the HCR™ Pro RNA-ISH HRP Starter Kit.

HCR™ Pro RNA-ISH Kit

The HCR™ Pro RNA-ISH kit consists of an HCR™ HiFi Probe and HCR™ Pro Detect reagents. Each HCR™ HiFi Probe includes 2 components: (1) HCR™ HiFi Probe 1A and (2) HCR™ HiFi Probe 1B. The probes are provided ready-to-use and can be transferred directly to open containers for use or storage.

HCR™ HiFi Probe

HCR™ Reagents	Amount for 20 Slides	Amount for 90 Slides	Storage Temperature
HCR™ HiFi Probe 1A ¹	7 mL	29 mL	2 to 8 °C
HCR™ HiFi Probe 1B ¹	7 mL	29 mL	2 to 8 °C

¹HCR™ HiFi Probes are platform-specific. For example, probes designed for the BOND RX autostainer are not compatible with the DISCOVERY ULTRA, and vice versa.

HCR™ Pro Detect

Please note that all HCR™ Pro Detect reagents are provided in amber bottles, as these reagents are sensitive to light.

HCR™ Reagents	Amount for 20 Slides	Amount for 90 Slides	Storage Temperature
Pretreat	5 mL	21 mL	2 to 8 °C
HCR™ Pro Detect A	5 mL	21 mL	2 to 8 °C
HCR™ Pro Detect B	7 mL	29 mL	2 to 8 °C
HCR™ Pro Detect C	5 mL	21 mL	2 to 8 °C
HCR™ Pro Detect D	7 mL	29 mL	2 to 8 °C
HCR™ Pro Detect E	7 mL	29 mL	2 to 8 °C
HCR™ Pro Detect F AP/HRP ¹	7 mL	29 mL	2 to 8 °C
Post-Process A	5 mL	21 mL	2 to 8 °C
Post-Process B ²	7 mL	29 mL	2 to 8 °C

¹HCR™ Pro Detect F AP is included in the HCR™ Pro RNA-ISH AP kit, and HCR™ Pro Detect F HRP is included in the HCR™ Pro RNA-ISH HRP kit.

²Post-Process B is only needed if you are performing an HCR™ Pro RNA-ISH + IHC/IF co-detection assay.

Required Materials for the BOND RX/RX^m

The HCR™ Pro RNA-ISH protocol requires specific materials available only from Leica Biosystems (LBS). It is essential to check the availability of these materials prior to setting up an HCR™ Pro RNA-ISH experiment. For more information, please inquire with your Leica Biosystems representative.

Materials from Leica Biosystems		
	Catalog #	Storage Temperature
BOND 6 mL Titration Kit	OPT9049	RT
BOND 7 mL Open Container	OP79193	RT
BOND 30 mL Open Container	OP309700	RT
BOND Universal Covertiles 160 Pack	S21.4611	RT
BOND Epitope Retrieval Solution 2	AR9640	2 to 8 °C
BOND Dewax Solution	AR9222	2 to 26 °C
BOND Wash Solution 10X Concentrate	AR9590	2 to 8 °C
BOND Polymer Refine Red Detection ¹	DS9390	2 to 8 °C
BOND Polymer Refine Detection ¹	DS9800	2 to 8 °C
BOND Research Detection	DS9455	RT
BOND Research Detection System 2	DS9777	RT
BOND Mixing Stations	S21.1971	RT

¹The HCR™ Pro RNA-ISH protocol uses Fast Red/DAB chromogens and hematoxylin from the BOND Polymer Refine Red Detection Kit or the BOND Polymer Refine Detection Kit for chromogenic deposition and counterstain. The HCR™ Pro RNA-ISH AP kit requires the use of the BOND Polymer Refine Red Detection Kit, and the HCR™ Pro RNA-ISH HRP kit requires the use of the BOND Polymer Refine Detection Kit.

Required Materials for Running Matisse® Chromogens on BOND RX/RX^m

Matisse® Red Chromogen for AP Detection

For those using our HCR™ Pro Detect reagents for AP detection, you can use our Matisse® Red chromogen for brighter and clearer staining. For further instructions on how to incorporate an HCR™ Pro RNA-ISH assay with Matisse® chromogens, reference **Appendix H** on page 38.

HCR™ Reagents	Amount for 20 Slides	Amount for 90 Slides	Storage Temperature
Matisse® Red	1.5 mL	2.5 mL	2 to 8 °C
Matisse® Red Buffer	7 mL	29 mL	2 to 8 °C
Hematoxylin	7 mL	29 mL	2 to 8 °C

Matisse® Brown and Green Chromogens for HRP Detection

For those using our HCR™ Pro Detect reagents for HRP detection, you can use our Matisse® Brown or Matisse® Green chromogens for brighter and clearer staining. For further instructions on how to incorporate an HCR™ Pro RNA-ISH assay with Matisse® chromogens, reference **Appendix H** on page 38.

Matisse® Brown Chromogen for HRP Detection

HCR™ Reagents	Amount for 20 Slides	Amount for 90 Slides	Storage Temperature
Matisse® Brown	2.5 mL	7 mL	2 to 8 °C
Matisse® Brown Buffer	7 mL	29 mL	2 to 8 °C
Hematoxylin	7 mL	29 mL	2 to 8 °C


Matisse® Green Chromogen for HRP Detection

HCR™ Reagents	Amount for 20 Slides	Amount for 90 Slides	Storage Temperature
Matisse® Green	4 mL	12.5 mL	2 to 8 °C
Matisse® Green Buffer	7 mL	29 mL	2 to 8 °C
Hematoxylin	7 mL	29 mL	2 to 8 °C

User-Supplied Materials

Materials from Other Vendors		
	Supplier	Comment
FFPE Sample Slides	Any, provided they meet the specifications mentioned in the BOND RX User Manual	SuperFrost or SuperFrost® Plus slides are recommended for best results
Propar (xylene substitute)	Fisher Scientific	Xylene may be substituted
Drying Oven	Any	Capable of maintaining temperature at ~60 °C
BioCare EcoMount, Leica CV Ultra Mounting Media, or Vectorlabs VectaMount	BioCare, Leica Biosystems, and Vectorlabs	Mounting medium compatible with all LBS chromogens
Cytoseal	Any	Suitable mounting medium for HRP-driven chromogens
ProLong™ Diamond Antifade Mountant (with or without DAPI) or SlowFade™ Diamond Antifade Mountant (with or without DAPI)	Thermo Fisher	Mounting medium compatible with RNA-FISH and IF staining
Cover Glass	Any	Dimension depends on the size of the tissue
Deionized Water	Any	
100% Ethanol	Any	Used for dehydration of tissues after staining only.
1X PBST	Any	For HCR™ Pro RNA-FISH only. 1X BOND Wash Buffer can be used as an alternative.
Tyramide Dyes	Any	See Appendix G for recommendations

Overall Workflow of the HCR™ Pro RNA-ISH Protocol



Each BOND RX/RX^m run takes 8 hours (i.e., 10 slides in one SSA). The timeline above only accounts for 5 hours and 48-53 minutes of this run (depending on whether you're performing the assay for chromogenic or fluorescent detection), as the remaining time comes from the additional BOND RX/RX^m washing steps.

Set Up BOND RX/RX^m Protocols

The HCR[™] Pro RNA-ISH assay utilizes a sequential staining approach, where the first sequence dispenses HCR[™] HiFi Probe A solution, and the second sequence dispenses HCR[™] HiFi Probe B solution, followed by the supporting detection reagents. As a result, both the BOND Refine (or Red) kit and the BOND Research Detection System are required to stain each slide.

The five main steps in setting up a BOND RX/RX^m protocol for an HCR[™] Pro RNA-ISH run are:

- i. Register BOND Containers associated with their respective reagents
- ii. Register the BOND Research Detection System
- iii. Create a staining protocol
- iv. Register probe reagents
- v. Set up slides and print labels

Register BOND Containers

Reagent names need to be **preferred** before BOND containers can be registered to their respective names. To prefer a reagent name, select the reagent name (e.g., *HCR Detect A) under **Reagent Setup**, and check the “**Preferred**” box.

Edit reagent properties

Name:

*HCR Detect A

Abbreviated name:

*HCR Detect A

Type:

Ancillary

Supplier:

Leica Microsystems

Compatible bulks:

*DI

*BWash

☒ Preferred

☐ Hazardous

Save

Cancel

The table below outlines a list of reagents that need to be preferred and registered with individual BOND containers.

BOND Container Registration for HCR™ Pro Reagents	
Reagent ¹	Type of Container ²
*Enzyme 2	7 mL or 30 mL BOND Container
*HCR™ Pro Detect A to E	7 mL or 30 mL BOND Container
*HCR™ Pro Detect F AP/HRP ³	7 mL or 30 mL BOND Container
*Post-Process A	7 mL or 30 mL BOND Container
*Post-Process B ⁴	7 mL or 30 mL BOND Container

¹The reagent names need to be preferred before they can be used for registering BOND Containers.

²We recommend using 7 mL and 30 mL BOND Containers for 20-slide and 90-slide kits, respectively.

³Register *HCR™ Pro Detect F AP and *HCR™ Pro Detect F HRP for AP and HRP detection, respectively.

⁴Post-Process B is only required for running an HCR™ Pro RNA-ISH + IHC/IF co-detection assay.

Register the BOND Research Detection System 2

1. Prefer ***Detection Wash** reagent under **Reagent Setup**.
2. Scan the barcode of your BOND Research Detection System 2.
3. Enter the name and expiration date.
4. Select “***Detection Wash**” in the first position, and scan a 30 mL BOND Open Container.
5. Click **Add**.

×

Add research reagent system

Name:

UPI:

Lot N°:

Expiration date: 

Reagents

Pstn.	UPI	Reagent	Vol. (mL)
1	26011754	*Detection Wash	30.00
2			
3			
4			
5			
6			
7			
8			
9			

[Add reagent](#) | [Remove reagent](#)

Add Cancel

*Note: The ***Detection Wash** Bond Container needs be placed in the first position in the BOND tray to ensure the BOND RX recognizes that the detection system (e.g., OPEN HCR RNA-ISH) is present.*

Create Staining Protocols for HCR™ Pro RNA-CISH

1. First Staining Protocol: From the top navigation bar, click on **Protocol setup** on the top of the screen and select **Staining** for the protocol group and **Leica Microsystem** for **Protocol origin**. Select ***HCR RNA-Chr ISH Seq_A**, and click on the **Copy** button near the top-right corner.



2. Delete the asterisk (*) from the **Name** and **Abbreviated name**, check **Preferred**, and select the name of the previously registered research detection system (e.g., **OPEN HCR RNA-ISH**) as the **Preferred detection system**. Click **Save**.

Edit protocol properties

Name:

Abbreviated name:

Description:

Staining method: ☐ Single ☒ Preliminary ☐ Final ☒ Preferred

[Import protocol](#) Protocol type: IHC staining

Preferred detection system:

Step N°	Wash	Reagent	Supplier	Ambient	Temperature	Inc. (min)	Dispense type
1		*Detection Wash	Leica Microsystems	✓		0:00	150 µL
2		*MARKER	Leica Microsystems		45	1:00	150 µL
3		*MARKER	Leica Microsystems		45	75:00	Intermediate


☐ Show wash steps [Insert wash](#) | [Insert reagent](#) | [Delete step](#)

3. Second Staining Protocol: Ensure the ***HCR RNA-Chr ISH HRP_B** and ***HCR RNA-Chr ISH AP_B** protocols have been “**Preferred**” in the protocol setup screen. These protocols will be used in their default form.

Register HCR™ HiFi Probe Reagents

It is essential to create staining protocols prior to probe registration. Each target probe (e.g., PPIB) requires registration of two probe reagents (e.g., **HCR HiFi Probe 1A PPIB** AND **HCR HiFi Probe 1B PPIB**).

1. Click on **Reagent setup** from the top navigation bar and select **Add**.



The screenshot shows the 'Reagent setup' window with a table of reagents. The 'Add' button is highlighted in the top left corner. The table lists various reagents with columns for Name, Abb. Name, Type, Other, Supplier, and Pref.

Name	Abb. Name	Type	Other	Supplier	Pref
*1.1 Part B	*1B	Ancillary	Other		✓
*1.20 Part A	*20A	Ancillary	Other		✓
*1.20 Part B	*20B	Ancillary	Other		✓
*1.1.60 Part A	*60A	Ancillary	Other		✓
*1.1.60 Part B	*60B	Ancillary	Other		✓
*1.1.60 Part C	*60C	Ancillary	Other		✓
*1.1.60 Part D	*60D	Ancillary	Other		✓
*AccuLyte CTC Antibody Amplification	*AC-Amp	Ancillary	RunCycle		✓
*AccuLyte CTC Enzymatic Biomarker CD45/DAPI	*AC-BMR	Ancillary	RunCycle		✓
*AccuLyte CTC Fluorescence Detection	*AC-Fleur	Ancillary	RunCycle		✓
*AccuLyte CTC Pan-Cytokeratin	*AC-pCK	Primary antibody	RunCycle		✓
*ACD Amp 1	*ACDamp1	Ancillary	Advanced Cell Diagnostics		✓
*ACD Amp 2	*ACDamp2	Ancillary	Advanced Cell Diagnostics		✓
*ACD Amp 3	*ACDamp3	Ancillary	Advanced Cell Diagnostics		✓
*ACD Amp 4	*ACDamp4	Ancillary	Advanced Cell Diagnostics		✓
*ACD Amp 5 Brown	*ACDamp5Br	Ancillary	Advanced Cell Diagnostics		✓
*ACD Amp 6 Red	*ACDamp6Rd	Ancillary	Advanced Cell Diagnostics		✓
*ACD Amp 6 Brown	*ACDamp6Br	Ancillary	Advanced Cell Diagnostics		✓
*ACD Amp 6 Red	*ACDamp6Rd	Ancillary	Advanced Cell Diagnostics		✓
*ACD Blue	*ACDblue	Ancillary	Advanced Cell Diagnostics		✓
*ACD Enzyme	*ACDE	Ancillary	Advanced Cell Diagnostics		✓
*Blue Part A	*BlueA	Ancillary	Leica Microsystems		✓
*Blue Part B	*BlueB	Ancillary	Leica Microsystems		✓

2. Check the **Preferred** box, and fill out the selections by referencing the images below.

Add reagent

Name:

Abbreviated name:

Type:

Supplier:

Staining method:

Single Preliminary Final

Default staining protocol:

Default HIER protocol:

Default enzyme protocol:

Compatible bulks:

*BWash

☒ Preferred ☐ Hazardous

Add reagent

Name:

Abbreviated name:

Type:

Supplier:

Staining method:

Single Preliminary Final

Default staining protocol:

Default HIER protocol:

Default enzyme protocol:

Compatible bulks:

*BWash

☒ Preferred ☐ Hazardous

3. To complete the registration process, default staining protocols for **staining methods** need to be selected as shown below. Click **Save**.
 - a. Note: **default staining protocols** for all **staining methods** need to be defined for the probe to be registered.

Probe	HCR™ HiFi Probe 1A ¹			HCR™ HiFi Probe 1B ¹		
Type	Primary antibody ²			Primary antibody ²		
Staining Methods	Single	Preliminary	Final	Single	Preliminary	Final
Default staining protocol	Any ³	*HCR RNA-Chr ISH Seq_A	Any ³	Any ³	Any ³	*HCR RNA-Chr ISH HRP_B
Default HIER protocol	*- - - -	*HIER 15 min with ER2 (95)	*- - - -	*- - - -	*- - - -	*- - - -
Default enzyme protocol	*- - - -	*Enzyme 2 for 5 min_Amb ⁴	*- - - -	*- - - -	*- - - -	*- - - -

¹Each target probe such as *PPIB* or *dapB* requires registration of two probe reagents.

²Although the reagent type is labeled as “Primary antibody,” it is specifically used for storing probe reagents.

³The selected protocol will not be used, so any existing protocol can be selected. However, the default staining protocol field cannot be left blank.

⁴***Enzyme 2 for 5 min_Amb** needs to be preferred before the selection appears in the dropdown menu.

4. Register your BOND Containers (e.g., 6 mL titration or 7/30 mL Open Containers) with the associated HCR™ HiFi Probe reagent names by scanning the barcode on the container. Select the **Reagent name** that was created previously and add an expiration date far into the future as the BOND Container can be reused. Click **OK**.

Create a Staining Protocol for HCR™ Pro RNA-FISH

The HCR™ Pro RNA-ISH HRP kit also supports tyramide-based fluorescent detection. The HCR™ Pro RNA-FISH assay follows the same workflow as the HCR™ Pro RNA-CISH assay described above, with a sequential stain that requires 2 staining protocols. To switch from a chromogenic to a fluorescent readout, you only need to modify the second staining protocol (i.e., the first staining protocol remains the same). To create the second staining protocol, copy ***HCR RNA-Chr ISH HRP_B**, change the **Name** to **HCR™ RNA-FISH_B**, and enter your preferred **Abbreviated name**. Select the previously registered research detection system (e.g., OPEN HCR RNA-ISH) as the **Preferred detection system**. Then, modify the protocol according to the detailed staining protocol shown on pages 30-31 (**Appendix C**).

Edit protocol properties

Name:

Abbreviated name:

Description:

Staining method: ☒ Single ☒ Preliminary ☒ Final ☒ Preferred

[Import protocol](#) Protocol type: IHC staining

Preferred detection system:

Step N°	Wash	Reagent	Supplier	Ambient	Temperature	Inc. (min)	Dispense type
30		*HCR Detect D	Leica Microsystems	✓		1:00	150 µL
31		*HCR Detect D	Leica Microsystems	✓		15:00	150 µL
40		*HCR Detect E	Leica Microsystems	✓		1:00	150 µL
41		*HCR Detect E	Leica Microsystems	✓		10:00	150 µL
50		*HCR Detect F HRP	Leica Microsystems	✓		1:00	150 µL
51		*HCR Detect F HRP	Leica Microsystems	✓		15:00	150 µL
60		Tyramide Cy5	Molecular Instruments	✓		1:00	150 µL
61		Tyramide Cy5	Molecular Instruments	✓		15:00	150 µL

☐ Show wash steps [Insert wash](#) | [Insert reagent](#) | [Delete step](#)

*Note: To run the HCR™ Pro RNA-FISH protocol, you need to use an HCR™ Pro RNA-ISH HRP kit and provide a third-party tyramide dye. Please see **Appendix G** for our recommended dyes and starting concentrations.*

Set Up Slides for HRP, AP, and Fluorescent-Based Protocols and Print Labels

- From the top navigation bar, click on **Slide setup** on the top of the screen and select **Add study**. Enter **Study ID**, **Study name**, and **Study comments**. Select ***Bake and Dewax** for the **Preparation protocol**. Click **OK**.

Add study

Study ID:

Study name:

Study comments:

Researcher: [Manage researchers](#)

Study N°:

Dispense volume:
☐ 100 µL
☒ 150 µL

Preparation protocol:

OK Cancel

- Follow all the selections shown below. **HCR PPIB Probe A** is selected as the **Marker** for the **First** sequence, and **HCR PPIB Probe B** is selected as the **Marker** for the **Final** sequence. Click on **Add slide**.
 - Please note:
 - PPIB is used here as an example. Other targets can be selected based on your RNA of interest.*
 - To perform AP-based detection, change the **Staining protocol** in the **Final** sequence to **"*HCR RNA-Chr AP_B."** Selections in the **First** sequence remain the same.*
 - To perform HCR™ Pro RNA-CISH with Matisse® chromogens (see **Appendix H**), please select the custom protocol that you created previously (e.g., HCR RNA-Chr ISH AP_B – Matisse Red) in the **Final** sequence. Selections in the **First** sequence remain the same.*
 - To perform fluorescent-based detection, change the **Staining protocol** in the **Final** sequence to **"HCR RNA-FISH_B."** Selections in the **First** sequence remain the same.*

HCR RNA-CISH HRP- PPIB

Tissue type:
☒ Test tissue
☐ Negative tissue
☐ Positive tissue

Dispense volume:
☐ 100 µL
☒ 150 µL

Staining mode:
 Sequential multiplex
 Routine

Stains:
 2

First Final

Process:
☒ IHC
☐ ISH

Marker:
 HCR PPIB Probe A (Molecular Instruments)

Protocols

Staining:
 HCR RNA-Chr ISH Seq_A

Preparation:
 *Bake and Dewax

HIER:
 *HIER 15 min with ER2 (95)

Enzyme:
 *Enzyme 2 for 5 min_Amb

Add slide Close

HCR RNA-CISH HRP- PPIB

Tissue type:
☒ Test tissue
☐ Negative tissue
☐ Positive tissue

Dispense volume:
☐ 100 µL
☒ 150 µL

Staining mode:
 Sequential multiplex
 Routine

Stains:
 2

First Final

Process:
☒ IHC
☐ ISH

Marker:
 HCR PPIB Probe B (Molecular Instruments)

Protocols

Staining:
 *HCR RNA-Chr ISH HRP_B

HIER:
 *HIER 15 min with ER2 (95)

Enzyme:
 *Enzyme 2 for 5 min_Amb

Add slide Close

3. For additional slides, keep this window open, change the parameters as needed, and then click **Add slide**. Continue until all slides have been added, then click **Close**.
4. After all the slides have been added, click on **Print labels**, select **All slide labels not yet printed for the current study**, and click **Print**.
5. Peel off the labels and attach them individually to the top frosted area of the slides.

Post-Processing for AP-Based Detection

After slides are unloaded from the BOND RX/RX^m, we recommend rinsing the slides with Deionized water. Bake the slides for 15 minutes (or until dry) at 60 °C. We recommend using either the EcoMount (Biocare), VectaMount (Vectorlabs), or Leica CV Ultra mounting media for cover-slipping.

Post-Processing for HRP-Based Detection

After slides are unloaded from the BOND RX/RX^m, we recommend rinsing the slides with Milli-Q water. Dehydrate by immersing the slides in 95% ethanol for 3 minutes twice followed by 100% ethanol for 3 minutes. Then, immerse the slides in a xylene (or xylene substitute) solution for 5 minutes and mount one slide at a time with Cytoseal (or any other xylene-based mounting medium). Allow slides to air dry for 5 minutes before imaging.

Post-Processing for Fluorescent Detection

After slides are unloaded from the BOND RX/RX^m, we recommend immersing the slides in 1X PBST (or 1X BOND Wash Solution) for 3 minutes twice. Mount one slide at a time with any suitable mounting medium that has worked for you. For mounting medium and counterstain, we recommend using the ProLong[™] Diamond Antifade Mountant (with or without DAPI) or the SlowFade[™] Diamond Antifade Mountant (with or without DAPI).

Reagent Preparation

Reagent Transfer and Storage


Instructions for reagent transfer can be found in the table below. Please transfer all HCR™ Pro RNA-ISH reagents to their name-associated containers. Please store all BOND Containers in a 4 °C fridge after use.

BOND Container	Instructions	Storage Conditions
*Enzyme 2	Transfer entirety of Pretreat to the BOND Container	Store at 4 °C after use
*Detection Wash	Fill the 30 mL BOND container with 1X BOND Wash Buffer	Store at 4 °C after use
HCR™ HiFi Probe 1A	Transfer the entirety of HCR™ HiFi Probe 1A Solution to the 6 mL, 7 mL, or 30 mL BOND Container	Store at 4 °C after use
HCR™ HiFi Probe 1B	Transfer the entirety of HCR™ HiFi Probe 1B Solution to the 6 mL, 7 mL, or 30 mL BOND Container	Store at 4 °C after use
*HCR™ Pro Detect A	Transfer entirety of HCR™ Pro Detect A to the 7 mL or 30 mL BOND Container	Store at 4 °C after use
*HCR™ Pro Detect B	Transfer entirety of HCR™ Pro Detect B to the 7 mL or 30 mL BOND Container	Store at 4 °C after use
*HCR™ Pro Detect C	Transfer entirety of HCR™ Pro Detect C to the 7 mL or 30 mL BOND Container	Store at 4 °C after use
*HCR™ Pro Detect D	Transfer entirety of HCR™ Pro Detect D to the 7 mL or 30 mL BOND Container	Store at 4 °C after use
*HCR™ Pro Detect E	Transfer entirety of HCR™ Pro Detect E to the 7 mL or 30 mL BOND Container	Store at 4 °C after use
*HCR™ Pro Detect F AP	Transfer entirety of HCR™ Pro Detect F AP to the 7 mL or 30 mL BOND Container	Store at 4 °C after use
*HCR™ Pro Detect F HRP	Transfer entirety of HCR™ Pro Detect F HRP to the 7 mL or 30 mL BOND Container	Store at 4 °C after use
*Post-Process A	Transfer entirety of Post-Process A to the 7 mL or 30 mL BOND Container	Store at 4 °C after use
*Post-Process B	Transfer entirety of Post-Process B to the 7 mL or 30 mL BOND Container	Store at 4 °C after use
HCR™ Membrane Stain	Transfer the entirety of HCR™ Membrane Stain to the 6 mL, 7 mL, or 30 mL BOND Container	Store at 4 °C after use

Other important notes:

1. Please avoid excessive bubbles when transferring HCR[™] HiFi Probe and HCR[™] Pro Detect Solutions to the 7 mL or 30 mL BOND Open Containers.
2. If the reagent-filled BOND Containers have not been used for more than a week, it is highly recommended to gently invert the BOND Containers several times with the lid closed to ensure solution uniformity before next use.

Overall Workflow of the HCR™ Pro RNA-ISH + IHC/IF Protocol



¹This step is NOT needed if HCR™ Pro RNA-CISH HRP detection is followed by IHC AP detection.

²This step is NOT needed if the species of the primary antibody used is rabbit.

Creating an HCR™ Pro RNA-ISH + IHC/IF Co-Detection Protocol

The HCR™ Pro RNA-ISH + IHC/IF Co-Detection Protocols comprises a 3-plex sequential multiplex stain. The first protocol is the same as described previously. The second protocol is a modified version of the HCR™ Pro RNA-ISH protocol, excluding the counterstain reagents. The third protocol involves the IHC/IF assay, incorporating an alternative chromogen or fluorophore. The list below outlines the main steps required to perform an HCR™ Pro RNA-ISH + IHC/IF co-detection assay:

- i. Remove the hematoxylin step from the HCR™ Pro RNA-CISH protocol for chromogenic detection (not applicable for fluorescent detection)
- ii. Create an IHC/IF protocol from an existing IHC protocol
- iii. Set up slides with three sequential staining protocols

Remove Hematoxylin from the HCR™ Pro RNA-CISH Protocol

Since HCR™ Pro RNA-CISH is performed first, it is essential that you remove the hematoxylin and bluing steps at the end of the HCR™ Pro RNA-ISH protocol for chromogenic detection. Start by copying the previously created *HCR™ RNA-Chr ISH HRP/AP_B protocol and editing the **Name** to **HCR RNA-Chr ISH AP** (or HRP) **No Hematoxylin**. Enter your preferred **Abbreviated name**. Highlight the ***Hematoxylin** step and select **Delete step** in the lower right corner of the window. Do the same for all steps thereafter, including the **Post-Process A** step and all of the ***Deionized Water** washing steps. For the **Staining method**, check **Preliminary** and click **Save**.

Edit protocol properties

Name:

Abbreviated name:

Description:

Staining method: ☐ Single ☒ Preliminary ☐ Final ☒ Preferred

[Import protocol](#) Protocol type: IHC staining

Preferred detection system:

Step N°	Wash	Reagent	Supplier	Ambient	Temperature	Inc. (min)	Dispense type
29		*HCR Detect D	Leica Microsystems	✓		1:00	150 µL
30		*HCR Detect D	Leica Microsystems	✓		15:00	150 µL
39		*HCR Detect E	Leica Microsystems	✓		1:00	150 µL
40		*HCR Detect E	Leica Microsystems	✓		10:00	150 µL
49		*HCR Detect F AP	Leica Microsystems	✓		1:00	150 µL
50		*HCR Detect F AP	Leica Microsystems	✓		15:00	150 µL
59		*Mixed Red Refine	Leica Microsystems	✓		1:00	150 µL
60		*Mixed Red Refine	Leica Microsystems	✓		9:00	150 µL

☐ Show wash steps [Insert wash](#) | [Insert reagent](#) | [Delete step](#)

Create an IHC/IF Protocol from an Existing IHC Protocol

IHC Protocol

- To create an IHC protocol, you can use ***IHC Protocol F** for HRP detection and ***IHC Protocol J** for AP detection.
 - Note: Be sure to select the protocol that uses the alternative chromogen corresponding to your second protocol.*

Name	Protocol type	Description	Modified by	Date
*HCR CTC IF Protocol	IHC staining	AccuCyte CTC IF protocol	Leica	3/15/2021
*HCR CTC IHC Protocol	IHC staining	AccuCyte CTC IHC protocol	Leica	3/15/2021
*IF Protocol	IHC staining	IF protocol	Leica	3/15/2021
*HCR Open Dispense Template	IHC staining	IHC template with Open Auxiliary and Chromogen dispenses	Leica	3/15/2021
*HCR Preliminary Blue	IHC staining	Multiplex - Preliminary Blue	Leica	3/15/2021
*HCR Preliminary DAB	IHC staining	Multiplex - Preliminary DAB	Leica	3/15/2021
*HCR Preliminary Red	IHC staining	Multiplex - Preliminary Red	Leica	3/15/2021
*HCR Protocol B	IHC staining	Bond Intense II IHC protocol	Leica	3/15/2021
*HCR Protocol F	IHC staining	Bond Polymer Refine IHC protocol	Leica	3/15/2021
*HCR Protocol F RX 37M	IHC staining	IHC Protocol F with marker step at 37C	Leica	3/15/2021
*HCR Protocol F RX 48M	IHC staining	IHC Protocol F with marker step at 48C	Leica	3/15/2021
*HCR Protocol J	IHC staining	Bond Polymer Refine Red IHC protocol	Leica	3/15/2021
*HCR Protocol J RX 37M	IHC staining	IHC Protocol J with marker step at 37C	Leica	3/15/2021
*HCR Protocol J RX 48M	IHC staining	IHC Protocol J with marker step at 48C	Leica	3/15/2021
*HCR Protocol K	IHC staining	ChromFlex 1 Dual IHC protocol	Leica	3/15/2021
*HCR Protocol K - IS Test	IHC staining	ChromFlex 1 Dual IHC protocol	Leica	3/15/2021
*HCR Protocol N	IHC staining	IHC Protocol using V60 mixed chromogen	Leica	3/15/2021
*HCR Protocol S	IHC staining	IHC HRP PLEX Protocol with Blue Chromogen	Leica	3/15/2021
*HCR Protocol T	IHC staining	IHC HRP PLEX Protocol with Green Chromogen	Leica	3/15/2021
*Opal 7 Color (v4.2 plus)	IHC staining	Parkes Elmer Opal multiplex protocol for v4.2 plus	Leica	3/15/2021
*Tap-up Dispenses Template	IHC staining	Template with multiple dispense types	Leica	3/15/2021
*Ultrascan ULMapper Assay	IHC staining	Ultrascan ULMapper Assay	Leica	3/15/2021
*ACD 2.5 DAB Flow B	ISH detection	ACD 2.5 DAB Flow B protocol	Leica	3/15/2021
*ACD 2.5 Red Flow B	ISH detection	ACD 2.5 Red Flow B protocol	Leica	3/15/2021
*ACD DAB ISH Protocol	ISH detection	ACD RNAscope DAB ISH protocol	Leica	3/15/2021
*ACD Red ISH Protocol	ISH detection	ACD RNAscope Red ISH protocol	Leica	3/15/2021
*FISH A-RJIO Protocol	ISH detection	Bond RJIO FISH A protocol	Leica	3/15/2021
*FISH Protocol C	ISH detection	FISH wash protocol	Leica	3/15/2021
*Flexible ISH Template	ISH detection	ISH template with open dispenses plus heated wash steps	Leica	3/15/2021
*ISH Heated Wash Template	ISH detection	ISH template with multiple wash temperatures	Leica	3/15/2021

- Click on the required protocol and select **copy**. Change the **Name**, **Abbreviated Name**, and **Description**. Modify the protocol by inserting ***Post-Process B** steps, one extra ***MARKER** step, additional washing steps, and ***Post-Process A** after the ***Hematoxylin** step. Check the **Final** box and click **Save**. Please see pages 32-35 (**Appendix D** and **Appendix E**) for the detailed staining protocols for IHC HRP and AP detections.

Edit protocol properties

Name:

Abbreviated name:

Description:

Staining method: ☐ Single ☐ Preliminary ☒ Final ☒ Preferred

BOND RX Import protocol Protocol type: IHC staining

Preferred detection system:

Step N°	Wash	Reagent	Supplier	Ambient	Temperature	Inc. (min)	Dispense type
1		*HCR Post-Process B	Leica Microsystems		42	1.00	150 µL
2		*HCR Post-Process B	Leica Microsystems		42	15.00	150 µL
9		*MARKER	Leica Microsystems	✓		1.00	150 µL
10		*MARKER	Leica Microsystems	✓		30.00	150 µL
17		*Post Primary	Leica Microsystems	✓		8.00	150 µL
24		*Polymer	Leica Microsystems	✓		8.00	150 µL
30		*Mixed Green Chromogen	Leica	✓		1.00	150 µL
31		*Mixed Green Chromogen	Leica	✓		8.00	150 µL

☐ Show wash steps Insert wash | Insert reagent | Delete step

Save Cancel

IF Protocol

1. To create an IF protocol:
 - a. Repeat Steps 1-4 in creating the IHC protocol
 - b. Replace chromogen deposition steps with third-party Tyramide dyes
 - c. Remove **Post-Process A**
2. Please see page 36 (**Appendix F**) for the detailed IF staining protocol.
 - a. *Note: The Tyramide incubation time (step 31) is a parameter that you need to optimize based on the primary antibody and abundance level of the target.*

Edit protocol properties

Name:

Abbreviated name:

Description:

Staining method: ☐ Single ☐ Preliminary ☒ Final ☒ Preferred

[Import protocol](#) Protocol type: IHC staining

Preferred detection system: 

Step N°	Wash	Reagent	Supplier	Ambient	Temperature	Inc. (min)	Dispense type
1		*HCR Post-Process B	Leica Microsystems		42	1:00	150 µL
2		*HCR Post-Process B	Leica Microsystems		42	15:00	150 µL
9		*MARKER	Leica Microsystems	✓		1:00	150 µL
10		*MARKER	Leica Microsystems	✓		30:00	150 µL
17		*Post Primary	Leica Microsystems	✓		8:00	150 µL
24		*Polymer	Leica Microsystems	✓		8:00	150 µL
30		Tyramide Dye	Molecular Instruments	✓		1:00	150 µL
31		Tyramide Dye	Molecular Instruments	✓		8:00	150 µL

☐ Show wash steps [Insert wash](#) | [Insert reagent](#) | [Delete step](#)

Set Up Slides with Three Sequential Staining Protocols

1. From the top navigation bar, click on **Slide setup** on the top of the screen and select **Add study**. Enter **Study ID**, **Study name**, and **Study comments**. Select ***Bake and Dewax** for the **Preparation protocol**. Click **OK**.
2. Click on **Add slide**, change the **Staining mode** to **Sequential multiplex** and choose **3** for **stains**. Under the **First** tab, the protocol selections are identical to performing HCR™ Pro RNA-CISH. Select the new hematoxylin-free HCR™ Pro RNA-ISH protocol that you created previously as the staining protocol under the **Second** tab. Under the **Final** tab, select the previously modified IHC/IF protocol as the staining protocol. Please reference the images down below for proper selections.

The image displays three sequential screenshots of the 'Add slide' configuration window in the software. Each window is titled 'Add slide' and shows the setup for a different staining protocol.

- First Window:** Shows the 'First' tab. The 'Staining mode' is set to 'Sequential multiplex' and 'Stains' is set to '2'. The 'Process' is 'IHC' and the 'Marker' is 'HCR PP1B Probe A (Molecular Instruments)'. The 'Staining' protocol is 'HCR RNA-CISH HRP- PP1B'.
- Second Window:** Shows the 'Second' tab. The 'Staining mode' is set to 'Sequential multiplex' and 'Stains' is set to '3'. The 'Process' is 'IHC' and the 'Marker' is 'HCR PP1B Probe B (Molecular Instruments)'. The 'Staining' protocol is 'HCR RNA-CISH + IHC Duplex'.
- Third Window:** Shows the 'Final' tab. The 'Staining mode' is set to 'Sequential multiplex' and 'Stains' is set to '3'. The 'Process' is 'IHC' and the 'Marker' is 'Primary 1 (M)'. The 'Staining' protocol is 'IHC Green Mouse Primary'.

Each window includes fields for 'Tissue type' (Test tissue, Negative tissue, Positive tissue) and 'Dispense volume' (100 µL, 150 µL). The 'Staining' field is a dropdown menu. The 'Preparation', 'HIER', and 'Enzyme' fields are also dropdown menus.

Note: Leave **HIER** and **Enzyme** blank for **Second** and **Final**, unless the **IHC** targets require additional **HIER** or enzyme digestion for optimal staining. Click **Add slide**.

Important Note: If a red chromogen is used for either ISH or IHC detection, slides cannot be dehydrated in alcohol. The slides need to be baked (~15 mins) or air-dried overnight prior to mounting.

Appendix

Appendix A: Detailed Staining Protocol for HCR™ RNA-Chr ISH Seq_A [Chromogenic Detection]

Step No.	Reagent	Step Type	Incubation Time	Temperature	Dispense Type
1	*Detection Wash	Reagent	0 min	Ambient	150 µL
2	*MARKER ¹	Reagent	1 min	45 °C	150 µL
3	*MARKER ¹	Reagent	75 min	45 °C	Intermediate
4	*BOND Wash Solution	Wash	0 min	45 °C	150 µL
5	*BOND Wash Solution	Wash	1 min	45 °C	150 µL
6	*BOND Wash Solution	Wash	1 min	45 °C	150 µL
7	*BOND Wash Solution	Wash	5 min	45 °C	150 µL
8	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
9	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
10	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
11	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
12	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
13	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
14	*BOND Wash Solution	Wash	0 min	Ambient	150 µL

¹*MARKER refers to the HCR™ HiFi Probe 1A solution.

Appendix B: Detailed Staining Protocol for HCR™ RNA-Chr ISH HRP B and HCR™ RNA-Chr ISH AP B [Chromogenic Detection]

Step No.	Reagent	Step Type	Incubation Time	Temperature	Dispense Type
1	*MARKER ¹	Reagent	1 min	45 °C	150 µL
2	*MARKER ¹	Reagent	75 min	45 °C	Intermediate
3	*BOND Wash Solution	Wash	0 min	45 °C	150 µL
4	*BOND Wash Solution	Wash	1 min	45 °C	150 µL
5	*BOND Wash Solution	Wash	1 min	45 °C	150 µL
6	*BOND Wash Solution	Wash	5 min	45 °C	150 µL
7	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
8	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
9	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
10	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
11	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
12	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
13	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
14	*HCR™ Pro Detect A	Reagent	3 min	42 °C	150 µL
15	*HCR™ Pro Detect B	Reagent	1 min	42 °C	150 µL
16	*HCR™ Pro Detect B	Reagent	60 min	42 °C	Intermediate
17	*BOND Wash Solution	Wash	0 min	42 °C	150 µL
18	*BOND Wash Solution	Wash	1 min	42 °C	150 µL
19	*BOND Wash Solution	Wash	5 min	42 °C	150 µL
20	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
21	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
22	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
23	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
24	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
25	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
26	*BOND Wash Solution	Wash	1 min	Ambient	Open
27	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
28	*HCR™ Pro Detect C	Reagent	5 min	Ambient	150 µL
29	*HCR™ Pro Detect D	Reagent	1 min	Ambient	150 µL
30	*HCR™ Pro Detect D	Reagent	15 min	Ambient	150 µL
31	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
32	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
33	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
34	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
35	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
36	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
37	*BOND Wash Solution	Wash	1 min	Ambient	Open
38	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
39	*HCR™ Pro Detect E	Reagent	1 min	Ambient	150 µL

40	*HCR™ Pro Detect E	Reagent	10 min	Ambient	150 µL
41	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
42	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
43	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
44	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
45	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
46	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
47	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
48	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
49	*HCR™ Pro Detect F HRP [#]	Reagent	1 min	Ambient	150 µL
50	*HCR™ Pro Detect F HRP [#]	Reagent	15 min	Ambient	150 µL
51	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
52	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
53	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
54	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
55	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
56	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
57	*BOND Wash Solution	Wash	1 min	Ambient	Open
58	*Deionized Water	Wash	1 min	Ambient	150 µL
59	*Mixed DAB Refine [#]	Reagent	1 min	Ambient	150 µL
60	*Mixed DAB Refine [#]	Reagent	7 min	Ambient	150 µL
61	*Deionized Water	Wash	0 min	Ambient	150 µL
62	*Deionized Water	Wash	0 min	Ambient	150 µL
63	*Deionized Water	Wash	0 min	Ambient	150 µL
64	*Deionized Water	Wash	0 min	Ambient	150 µL
65	*Deionized Water	Wash	0 min	Ambient	150 µL
66	*Deionized Water	Wash	0 min	Ambient	150 µL
67	*Hematoxylin [#]	Reagent	3 min	Ambient	150 µL
68	*Deionized Water	Wash	0 min	Ambient	150 µL
69	*Deionized Water	Wash	0 min	Ambient	150 µL
70	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
71	*Deionized Water	Wash	0 min	Ambient	150 µL
72	*Deionized Water	Wash	0 min	Ambient	150 µL
73	*Deionized Water	Wash	0 min	Ambient	150 µL
74	*Post-Process A	Reagent	5 min	Ambient	150 µL
75	*Deionized Water	Wash	0 min	Ambient	150 µL
76	*Deionized Water	Wash	0 min	Ambient	150 µL
77	*Deionized Water	Wash	0 min	Ambient	150 µL
78	*Deionized Water	Wash	0 min	Ambient	150 µL
79	*Deionized Water	Wash	0 min	Ambient	150 µL

¹*MARKER refers to the HCR™ HiFi Probe 1B solution.

#For AP-based detection, please make changes to steps 49, 50, 59, and 60 as follows:

49	*HCR [™] Pro Detect F AP [#]	Reagent	1 min	Ambient	150 µL
50	*HCR [™] Pro Detect F AP [#]	Reagent	15 min	Ambient	150 µL

59	*Mixed Red Refine	Reagent	1 min	Ambient	150 µL
60	*Mixed Red Refine	Reagent	9 min	Ambient	150 µL

Appendix C: Detailed Staining Protocol for HCR™ RNA-FISH B [Fluorescent Detection]

Step No.	Reagent	Step Type	Incubation Time	Temperature	Dispense Type
1	*Detection Wash	Wash	0 min	Ambient	150 µL
2	*MARKER ¹	Reagent	1 min	45 °C	150 µL
3	*MARKER ¹	Reagent	75 min	45 °C	Intermediate
4	*BOND Wash Solution	Wash	0 min	45 °C	150 µL
5	*BOND Wash Solution	Wash	1 min	45 °C	150 µL
6	*BOND Wash Solution	Wash	1 min	45 °C	150 µL
7	*BOND Wash Solution	Wash	5 min	45 °C	150 µL
8	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
9	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
10	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
11	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
12	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
13	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
14	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
15	*HCR™ Pro Detect A	Reagent	3 min	42 °C	150 µL
16	*HCR™ Pro Detect B	Reagent	1 min	42 °C	150 µL
17	*HCR™ Pro Detect B	Reagent	60 min	42 °C	Intermediate
18	*BOND Wash Solution	Wash	0 min	42 °C	150 µL
19	*BOND Wash Solution	Wash	1 min	42 °C	150 µL
20	*BOND Wash Solution	Wash	5 min	42 °C	150 µL
21	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
22	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
23	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
24	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
25	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
26	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
27	*BOND Wash Solution	Wash	1 min	Ambient	Open
28	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
29	*HCR™ Pro Detect C	Reagent	5 min	Ambient	150 µL
30	*HCR™ Pro Detect D	Reagent	1 min	Ambient	150 µL
31	*HCR™ Pro Detect D	Reagent	15 min	Ambient	150 µL
32	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
33	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
34	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
35	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
36	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
37	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
38	*BOND Wash Solution	Wash	1 min	Ambient	Open
39	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
40	*HCR™ Pro Detect E	Reagent	1 min	Ambient	150 µL

41	*HCR™ Pro Detect E	Reagent	10 min	Ambient	150 µL
42	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
43	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
44	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
45	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
46	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
47	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
48	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
49	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
50	*HCR™ Pro Detect F HRP [#]	Reagent	1 min	Ambient	150 µL
51	*HCR™ Pro Detect F HRP [#]	Reagent	15 min	Ambient	150 µL
52	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
53	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
54	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
55	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
56	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
57	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
58	*BOND Wash Solution	Wash	1 min	Ambient	Open
59	*Deionized Water	Wash	1 min	Ambient	150 µL
60	Tyramide	Reagent	1 min	Ambient	150 µL
61	Tyramide	Reagent	15 min	Ambient	150 µL
62	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
63	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
64	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
65	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
66	*BOND Wash Solution	Wash	0 min	Ambient	150 µL

¹*MARKER refers to the HCR™ HiFi Probe 1B solution.

Appendix D: Detailed IHC HRP Detection Staining Protocol for HCR™ Pro RNA-CISH + IHC Co-Detection

Step No.	Reagent	Step Type	Incubation Time	Temperature	Dispense Type
1	*Post-Process B	Reagent	1 min	42 °C	150 µL
2	*Post-Process B	Reagent	15 min	42 °C	150 µL
3	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
4	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
5	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
6	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
7	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
8	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
9	*MARKER ¹	Reagent	1 min	Ambient	150 µL
10	*MARKER ¹	Reagent	30 min	Ambient	150 µL
11	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
12	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
13	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
14	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
15	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
16	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
17	*Post Primary ²	Reagent	8 min	Ambient	150 µL
18	*BOND Wash Solution ²	Wash	0 min	Ambient	150 µL
19	*BOND Wash Solution ²	Wash	0 min	Ambient	150 µL
20	*BOND Wash Solution ²	Wash	0 min	Ambient	150 µL
21	*BOND Wash Solution ²	Wash	2 min	Ambient	150 µL
22	*BOND Wash Solution ²	Wash	2 min	Ambient	150 µL
23	*BOND Wash Solution ²	Wash	2 min	Ambient	150 µL
24	*Polymer	Reagent	8 min	Ambient	150 µL
25	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
26	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
27	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
28	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
29	*Deionized Water	Wash	0 min	Ambient	150 µL
30	*Mixed Green Chromogen [#]	Reagent	1 min	Ambient	150 µL
31	*Mixed Green Chromogen [#]	Reagent	8 min	Ambient	150 µL
32	*Deionized Water	Wash	0 min	Ambient	150 µL
33	*Deionized Water	Wash	0 min	Ambient	150 µL
34	*Deionized Water	Wash	0 min	Ambient	150 µL
35	*Deionized Water	Wash	0 min	Ambient	150 µL
36	*Deionized Water	Wash	0 min	Ambient	150 µL
37	*Hematoxylin	Reagent	3 min	Ambient	150 µL
38	*Deionized Water	Wash	0 min	Ambient	150 µL

39	*Deionized Water	Wash	0 min	Ambient	150 µL
40	*Deionized Water	Wash	0 min	Ambient	150 µL
41	*Deionized Water	Wash	0 min	Ambient	150 µL
42	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
43	*Deionized Water	Wash	0 min	Ambient	150 µL
44	*Post-Process A	Reagent	5 min	Ambient	150 µL
45	*Deionized Water	Wash	0 min	Ambient	150 µL
46	*Deionized Water	Wash	0 min	Ambient	150 µL
47	*Deionized Water	Wash	0 min	Ambient	150 µL
48	*Deionized Water	Wash	0 min	Ambient	150 µL
49	*Deionized Water	Wash	0 min	Ambient	150 µL

¹*MARKER refers to the primary antibody.

²These steps may be deleted if you are using a primary antibody that is already cloned from rabbit.

#You can switch to using the DAB chromogen by making changes to steps 30 and 31 as follows:

30	*Mixed DAB Refine	Reagent	0 min	Ambient	150 µL
31	*Mixed DAB Refine	Reagent	10 min	Ambient	150 µL

Appendix E: Detailed IHC AP Detection Staining Protocol for HCR™ Pro RNA-CISH + IHC Co-Detection

Step No.	Reagent	Step Type	Incubation Time	Temperature	Dispense Type
1	Post-Process B	Reagent	1 min	42 °C	150 µL
2	Post-Process B	Reagent	15 min	42 °C	150 µL
3	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
4	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
5	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
6	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
7	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
8	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
9	*MARKER ¹	Reagent	1 min	Ambient	150 µL
10	*MARKER ¹	Reagent	30 min	Ambient	150 µL
11	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
12	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
13	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
14	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
15	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
16	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
17	*Post Primary AP	Reagent	20 min	Ambient	150 µL
18	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
19	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
20	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
21	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
22	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
23	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
24	*Polymer AP	Reagent	30 min	Ambient	150 µL
25	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
26	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
27	*BOND Wash Solution	Wash	5 min	Ambient	Intermediate
28	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
29	*BOND Wash Solution	Wash	0min	Ambient	150 µL
30	*Deionized Water	Wash	0 min	Ambient	150 µL
31	*Mixed Red Refine	Reagent	10 min	Ambient	150 µL
32	*Mixed Red Refine	Reagent	5 min	Ambient	150 µL
33	*Deionized Water	Wash	0 min	Ambient	150 µL
34	*Deionized Water	Wash	0 min	Ambient	150 µL
35	*Deionized Water	Wash	0 min	Ambient	150 µL
36	*Deionized Water	Wash	0 min	Ambient	150 µL
37	*Deionized Water	Wash	0 min	Ambient	150 µL
38	*Hematoxylin	Reagent	3 min	Ambient	150 µL
39	*Deionized Water	Wash	0 min	Ambient	150 µL

40	*Deionized Water	Wash	0 min	Ambient	150 µL
41	*Deionized Water	Wash	0 min	Ambient	150 µL
42	*Deionized Water	Wash	0 min	Ambient	150 µL
43	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
44	*Deionized Water	Wash	0 min	Ambient	150 µL
44	*Post-Process A	Reagent	5 min	Ambient	150 µL
45	*Deionized Water	Wash	0 min	Ambient	150 µL
46	*Deionized Water	Wash	0 min	Ambient	150 µL
47	*Deionized Water	Wash	0 min	Ambient	150 µL
48	*Deionized Water	Wash	0 min	Ambient	150 µL
49	*Deionized Water	Wash	0 min	Ambient	150 µL

¹*MARKER refers to the primary antibody.

Appendix F: Detailed IF Detection Staining Protocol for HCR™ Pro RNA-FISH + IF Co-Detection

Step No.	Reagent	Step Type	Incubation Time	Temperature	Dispense Type
1	*Post-Process B	Reagent	1 min	42 °C	150 µL
2	*Post-Process B	Reagent	15 min	42 °C	150 µL
3	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
4	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
5	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
6	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
7	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
8	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
9	*MARKER ¹	Reagent	1 min	Ambient	150 µL
10	*MARKER ¹	Reagent	30 min	Ambient	150 µL
11	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
12	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
13	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
14	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
15	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
16	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
17	*Post Primary	Reagent	8 min	Ambient	150 µL
18	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
19	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
20	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
21	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
22	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
23	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
24	*Polymer	Reagent	8 min	Ambient	150 µL
25	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
26	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
27	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
28	*BOND Wash Solution	Wash	2 min	Ambient	150 µL
29	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
30	Tyramide Dye	Reagent	1 min	Ambient	150 µL
31	Tyramide Dye	Reagent	8 min	Ambient	150 µL
32	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
33	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
34	*BOND Wash Solution	Wash	0 min	Ambient	150 µL
35	*BOND Wash Solution	Wash	1 min	Ambient	150 µL
36	*BOND Wash Solution	Wash	0 min	Ambient	150 µL

¹*MARKER refers to the primary antibody.

Appendix G: Third-Party Recommended Tyramide Dyes

Validated Tyramide Dyes	Incubation Time	Recommended Starting Concentration	Vendor	Catalog #
CF488A	8-24 min	5 µM	Biotium	92171
CF550R	8-24 min	5 µM	Biotium	96077
CF555	8-24 min	5 µM	Biotium	96021
CF583R	8-24 min	5 µM	Biotium	96085
CF594	8-24 min	5 µM	Biotium	92174
CF640R	8-24 min	5 µM	Biotium	92175
CF740	8-24 min	5 µM	Biotium	96124
Alexa Fluor 488 - Tyramide	8-24 min	1x	ThermoFisher	B40953
Alexa Fluor 546 - Tyramide	8-24 min	1x	ThermoFisher	B40954
Alexa Fluor 594 - Tyramide	8-24 min	1x	ThermoFisher	B40957
iFluor488 Tyramide	8-24 min	1:600	AAT Bioquest	45100
iFluor647 Tyramide	8-24 min	1:200	AAT Bioquest	45110

Appendix H: Create a Staining Protocol using Matisse® Chromogens

The two main steps in setting up an HCRTM Pro RNA-ISH with the Matisse® chromogens are:

1. Register and fill two 7 mL (or 30 mL) Open BOND Containers for the Matisse® chromogen and one 7 mL (or 30 mL) Open BOND Container for the hematoxylin counterstain
2. Create a custom staining protocol with the Matisse® chromogen

Register the Matisse® Chromogen and Hematoxylin Counterstain

Online mixing of chromogens requires registering and filling two 7 mL BOND Open Containers. The following example demonstrates the registration process for the Matisse® Red chromogen; however, these steps can be similarly applied to register other Matisse® chromogens.

1. Register two 7 mL BOND Containers for mixing at a ratio of 20:1 as shown below:

Add open container

Bond Open Container, 7 mL
Catalog N°: OP79172 UPI: 26035976
Supplier: Leica Microsystems

Reagent name: *Mix 1, 20:1 Part 1B

Lot N°:

Expiration date: 4/24/2026

Initial vol. (mL): 7.00

OK Cancel

Add open container

Bond Open Container, 7 mL
Catalog N°: OP79172 UPI: 26035882
Supplier: Leica Microsystems

Reagent name: *Mix 1, 20:1 Part 20A

Lot N°:

Expiration date: 6/6/2026

Initial vol. (mL): 7.00

OK Cancel

2. Transfer 1.5 mL of the Matisse® Red Chromogen Substrate and 7 mL of the Matisse® Red Chromogen Buffer to **Mix 1, 20:1 Part 1B** and **Mix 1, 20:1 Part 20A** BOND Containers, respectively. The table below outlines the required volumes for both 20-slide and 90-slide scales across our range of Matisse® chromogens. Make sure you input the correct mixing ratios specific to the Matisse® chromogen used in your HCRTM Pro RNA-ISH assay.

Matisse® Chromogens	Reagents	Volume Transferred to BOND Containers (20-Slide Kit)	Volume Transferred to BOND Containers (90-Slide Kit)	Name of Registered BOND Containers
Matisse® Red	Matisse® Red	1.5 mL ¹	2.5 mL ¹	*Mix 1, 20:1 Part 1B
	Matisse® Red Buffer	7 mL ¹	29 mL ²	*Mix 1, 20:1 Part 20A
Matisse® DAB (Brown)	Matisse® DAB	2.5 mL ¹	7 mL ¹	*Mix 9, 6:1 Part 1B
	Matisse® DAB Buffer	7 mL ¹	29 mL ²	*Mix 9, 6:1 Part 6A
Matisse® Green	Matisse® Green	4 mL ¹	12.5 mL ²	*Mix 2, 2:1 Part 1B
	Matisse® Green Buffer	7 mL ¹	29 mL ²	*Mix 2, 2:1 Part 2A

¹Use of 7 mL BOND Open Container is recommended.


²Use of 30 mL BOND Open Container is recommended. If 7 mL Containers are used, you can refill with the chromogen and chromogen buffers on an as-need basis.

3. Register another 7 mL (or 30 mL) BOND Open Container and name it **Hematoxylin**. Transfer the entirety of the provided Hematoxylin into this container.

Create a Custom Staining Protocol with the Matisse® Chromogen

To add Matisse® chromogens to the staining protocol, you need to create a custom staining protocol by using a pre-existing HCR™ Pro RNA-ISH protocol (e.g., *HCR RNA-Chr ISH AP_B) as a template:

- From the top navigation bar, click on **Protocol setup** on the top of the screen and select **Staining** for the protocol group and **Leica Microsystem** for **Protocol origin**. Select ***HCR RNA-Chr ISH AP_B** (or ***HCR RNA-Chr ISH HRP_B** if you plan to use HRP-based chromogens), and click on the **Copy** button near the top-right corner.




Protocol setup

Protocol name	Protocol type	
*HCR RNA-Chr ISH AP_B	IHC staining	HCR RNA-Chromogenic ISH with AP_2nd probe
*HCR RNA-Chr ISH HRP_B	IHC staining	HCR RNA-Chromogenic ISH with HRP_2nd probe

- Modify the protocol:

- Select the same preferred detection system (e.g., OPEN HCR RNA-ISH) as you created previously under the “Create Staining Protocols for HCR™ Pro RNA-CISH” section on page 13.
- Add one “*Detection Wash” step as the first step.
- Replace Leica Chromogens (e.g., *Mixed Red Refine) with user-defined mixing steps (e.g., *Mixed_1, 20A:1B (10')).
 - Note:** Make sure mixing steps such as *Mixed_1, 20A:1B (10') are preferred before they can be used in the staining protocol.
- Replace *Hematoxylin from LBS with Hematoxylin from Molecular Instruments.

The staining protocol should resemble the images shown below:



Please note that Matisse® Red chromogens are typically only stable within 30 minutes after mixing. Thus, it is important to select **Mixed_1, 20A:1B (10')** as opposed to **Mixed_1, 20A:1B (300')**. Refer to the

table below for detailed information on incubation times and names of the chromogenic deposition steps.

Chromogen	Recommended Chromogen Incubation Time (2 steps)	Name of the Chromogenic Deposition Steps
Matisse® Red	1 min	Mixed_1 20A:1B (10')
	13 min	
Matisse® DAB (Brown)	1 min	Mixed_1 6A:1B (10') or Mixed_1 6A:1B (300')
	13 min	
Matisse® Green	1 min	Mixed_2 2A:1B (10')

Appendix I: Tips & Tricks

Adjusting Signal Intensity with HCR™ Pro RNA-ISH

The HCR™ Pro RNA-ISH signal intensity can be adjusted using various parameters to either decrease or increase the strength of the signal.

If you would like to tune down the signal intensity to facilitate data interpretation (e.g., dot counting), we recommend the following options:

Level of Tuning	Guidelines
Coarse-tuning¹	To significantly reduce signal intensity, you can reduce the incubation time of HCR™ Pro Detect E (Step #53 as written in Appendix A and B) from 10 minutes to 5 minutes .
Fine-tuning²	To finely reduce signal intensity, you can reduce the incubation time of HCR™ Pro Detect B (Step #29 as written in Appendix A and B) from 60 minutes to 45 minutes .

Note: If further signal reduction is desired, you can also try reducing the **chromogen deposition** (Step #73 as written in Appendix A) time from 9 minutes to **7 minutes**. (This only applies if you are using a red chromogen.)

If you would like to increase the signal intensity, you can try the following:

Level of Tuning	Guidelines
Coarse-tuning¹	Increase the Antigen Retrieval time from 15 minutes to up to 30 minutes or increase the incubation time of HCR™ Pro Detect E (Step #53 as written in Appendix A and B) from 10 minutes to 15-20 minutes .
Fine-tuning²	Increase the incubation time of HCR™ Pro Detect B (Step #29 as written in Appendix A and B) from 60 minutes to 120 minutes .

Note: If further signal enhancement is desired, you can increase the incubation time of **HCR™ Pro Detect D** and **HCR™ Pro Detect F** from 15 minutes to **20 minutes** (Steps #43 and #63, respectively, as written in Appendix A and B).

¹Use coarse-tuning if the signal needs to be adjusted significantly.

²Use fine-tuning if the signal only needs to be tuned up/down slightly.

Increasing RNA Signal Performance

We have three recommendations for increasing the performance of RNA signal with your HCR[™] HiFi Probe:

1. To open up the tissue more (i.e., break the cross-linkage caused by formalin fixation) and help accessibility of the HCR[™] HiFi Probe and Detect reagents to the target, increase the **Antigen Retrieval** time from 15 to up to **30 minutes**.
2. Though we recommend a 1.25-hour *Marker incubation (the probe hybridization time), this may not be optimal across all tissue types. The incubation time for this step (Steps #3 and #16 as written in Appendix A and B) can be increased from 1.25 hours to **2 hours** to allow sufficient time for probes to diffuse in and bind to their targets.
3. A hallmark feature of the HCR[™] Pro RNA-ISH assay is that it has been optimized to not require protease digestion. However, there are rare cases in which protease can help increase the accessibility of the HCR[™] HiFi Probe to the target molecule. In these cases, incorporating a mild protease step can result in improvements, but could negatively impact co-detection capabilities. We suggest using the BOND Enzyme Pretreatment kit [AR9551] with the following parameters: **0.5 µg/mL, 1:34000 at 37 °C for 15 minutes**.

Appendix J: FAQ

1. **How do I incorporate HCR[™] Pro RNA-ISH into existing IHC/IF assays?**
 - a. Please check out this [blog post](#) for a general overview of HCR[™] Pro RNA-ISH and IHC/IF co-detection. For the IHC/IF portion of the protocol, you can use any off-the-shelf primary antibody and detection kit you have already validated, such as Leica's [BOND Polymer Refine Detection Kit](#), from Leica Biosystems. There are no additional reagents that are required from MI to perform an IHC/IF co-detection assay with HCR[™] Pro RNA-ISH.
2. **Which chromogens do you recommend for an HCR[™] Pro RNA-ISH + IHC co-detection assay?**
 - a. We recommend using a red chromogen for RNA-ISH and a green chromogen for IHC (e.g., BOND [Green Chromogen, from Leica Biosystems \[DS9914\]](#)). This combination provides a nice contrast between targets, but you are not limited to these colors in any technical way.
3. **Which detection systems do you recommend using with the HCR[™] Membrane Stain?**
 - a. For ready-to-use amplified detection, we suggest using the BOND Polymers (either HRP or AP conjugated). For direct-labeled detection, we recommend using either the [Donkey Anti Rabbit \(IgG\) secondary antibody \(Alexa Fluor 647\)](#) or the [Donkey Anti Rabbit \(IgG\) secondary antibody \(Alexa Fluor 555\)](#).
4. **I notice bubbles on my slides that have little to no staining. How do I resolve this?**
 - a. In general, we advise that you follow these steps:
 - i. Check that you programmed each step of our protocol correctly (reference Appendix A-E). Double-check that **Open** is selected as the Dispense Type for **Steps #26, #37, and #57** (as written in Appendix B).
 1. The Open Wash dispense type serves to eliminate bubbles trapped beneath the covertedile, so it is essential that you ensure that the correct dispense type is selected.
 - ii. Contact your Leica FAS/FSE.
 - b. Skip to this section if you noticed bubbles with HCR[™] Pro RNA-ISH + IHC/IF co-detection and observed no impact to your ISH signal.
 - i. In this case, we suggest that you implement an open wash step as the bubbles may have originated from either the Post-Process B reagent or the primary antibody. Since the primary diluent is slightly more viscous, it can lead to additional bubbles during the transfer of the reagent to the BOND Container.
 - ii. First, adjust the dispense type to **Open Wash** for **Steps #7 and #15** (as written in Appendix D and E).
 - iii. Ensure there are no excessive bubbles inside the BOND Container filled with the primary antibody.
 - iv. Rescan and electronically refill BOND Containers to allow the autostainer to remeasure the volumes.

5. How do I lighten the hematoxylin staining on my tissues?

- a. You can try either of the following two options outlined below to lighten hematoxylin staining:
 - i. Decrease the **Hematoxylin** incubation time from 3 minutes to **1 or 2 minutes**.
 - ii. Dilute 1/3 of the **Hematoxylin** reagent with DI water and use this as a counterstain. For this option, please note that you will need to purchase a separate [Hematoxylin](#) reagent from Leica Biosystems and do a 3-fold dilution. You will also need to register a separate BOND Container to use this reagent in the assay.

6. Can I skip Post-Process A (bluing step) after hematoxylin counterstain step?

- a. Yes, you can! The BOND washes towards the end of the protocol can help with tissue bluing, so if you are happy with your results, feel free to omit bluing entirely.

7. Do you use ER1 or ER2?

- a. We use ER2 for HCR[™] Pro RNA-ISH. If your protein target requires ER1, then you can perform the ISH staining first and incorporate a separate ER1 step when conducting the IHC/IF staining.

8. I'd like to incorporate protease in my sample. What are your suggestions?

- a. Our HCR[™] Pro RNA-ISH protocol does not require any protease digestion, thereby supporting native compatibility with IHC/IF and preserving tissue morphology. However, we understand that there are certain situations where you may want to incorporate protease in your sample like over-fixed tissue samples or the need to unmask epitopes for certain IHC/IF targets (e.g., for an ISH + IHC/IF co-detection assay).
- b. In such cases, we recommend that you use a light protease digestion at a diluted level and/or increase both ***Marker incubation (probe hybridization step)** and **amplification** (Step #16 as written in Appendix B) times to **2 hours**. As a starting point, you can choose to dilute your protease to **0.5 µg/mL at 37 °C**, and treat your sample for **5-10 minutes**. The digestion step (e.g., followed by 3-4 BOND Washing steps) can be added in the beginning of the "HCR[™] RNA-Chr ISH Seq_A" protocol (see Appendix A) before the ***Detection Wash step**. Please feel free to adjust the concentration later on as needed.