ALVEO
TECHNOLOGIES, Inc.

Offline Mode Requirements

Offline mode is intended for:

- Unplanned cellular or WiFi disconnection during the testing process.
- Users testing in remote settings that will not have access to cellular or WiFi

This training material applies to customers using Alveo Sense outside of the US.

Product Availability Disclaimer: This product is not available for sales or distribution in the United States. Availability of Alveo Sense Avian Influenza is subject to local regulatory approvals and may vary by country. For inquiries regarding the availability of Alveo Sense Avian Influenza in your country, please contact our Customer Service team at support@alveotechnologies.com.

ALVEO
TECHNOLOGIES, Inc.

The system saves **up to 10 results** that will display in the Mobile App as soon as cellular or WiFi are restored.

Specific setup while connected to wifi or cellular service is required before offline mode can be enabled for testing in remote settings.



TABLE OF CONTENTS

1. User Registration (required only for first login) 3
2. Prepare Analyzers for Offline Mode 6
3. Download Updated Assay 9
4. Cartridge QR Codes 14

1

USER REGISTRATION

(required only for first login).



01. Open App



02. Click **“Register”**



03. Fill in all information



04. Click **"Confirm Information"** once all required fields are filled



05. Scroll to bottom of screen to accept TOU, Click **"I agree"**



06. Confirm your information and click **"Send email"**



07. Go to email inbox to verify



08. Click on the **Confirmation Link**



09. When **“Registration is successful”** screen appears, go back to app



10. Enter registered email & password and click **“Log In”**



11. User is logged in and homepage is displayed

2

PREPARE ANALYZERS FOR OFFLINE MODE.



01. Login to App



02. Click on **Manage Analyzers** and give permission to Bluetooth®



03. Click on **“Pair New Analyzer”**



04. Give Location permissions



05. Click **“Begin Pairing”**



06. Wait for the Analyzers to show



07. Select Analyzer to Pair and Click **“Next Step”**



08. Analyzer is **paired** and **“Offline Enabled”** is shown

3

DOWNLOAD UPDATED ASSAY



01. Open App



02. Click on **“Add Farm”**



03. Fill details and click **“Next”**



04. Confirm the Farm and click **“Save”**



05. Farm is added



06. Select the Farm



07. Enter bird details



08. Select Clinical signs



09. Select Postmortem Findings



10. Select the analyzer to pair OR Pair with a new analyzer



11. Click **“Next”** in the analyzer details screen



12. Click **“Scan Cartridge QR Code”**



13. Allow Camera permissions



14. Scan the sample Cartridge QR code provided on page 14.



15. Cartridge is scanning



16. Confirm the scanned Cartridge details then click **"Yes"**



17. The assay file is automatically downloaded at this step



18. Click on the "**X**" in the right top corner



19. Confirm test cancellation



20. User is taken back to Manage Analyzers screen

The phone is now set up for use in offline mode

Assay file is stored in app. Mobile application and **paired** analyzer are ready to use in offline mode.

4

For OFFLINE MODE
PREPARATION USE ONLY -
DO NOT proceed to cartridge insertion step on app.

H5 H7 Cartridge QR Codes to Scan - DO NOT proceed to cartridge insertion step on app.

For Use with Product: A-FIN-000037
Alveo Sense Poultry Avian Influenza Test Type A H5 H7 Kit (Production)

01008600040245761126042417270424109999999262100001

01008600040245761126042417270424109999999262100002

H5 H9 Cartridge QR Codes to Scan - DO NOT proceed to cartridge insertion step on app.

For Use with Product: A-FIN-000031
Alveo Sense Poultry Avian Influenza Test Type A H5 H9 Kit (Production)

01008600040245521126042417270424109999999262100001

01008600040245521126042417270424109999999262100002