



CERTIFICATE OF CONFORMITY

CE EMC Test Report

For the following information

Ref. File No.: C1M2011083
(C1M1702005)

Product	Medical Grade Mobile Tablet PC
Model No.	MMS0700
Brand Name	MACTRON GROUP
Applicant	MACTRON GROUP CO., LTD.
Test Report Number	EM-E200666
Standards	EN 60601-1-2:2015 (IEC 60601-1-2: 2014) Emission: CISPR 11: 2009 +A1:2010, Class B IEC 61000-3-2:2014 & IEC 61000-3-3:2013 Immunity: IEC 61000-4-2:2008, IEC 61000-4-3:2010, IEC 61000-4-4:2012, IEC 61000-4-5:2014, IEC 61000-4-6:2013, IEC 61000-4-8:2009, IEC 61000-4-11:2004

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the council EMC directive 93/42/EEC. The test data and results are issued on the EMC test report no. EM-E200666.

Signature


Ben Cheng/Manager
Date: 2020. 11. 10

Test Laboratory:
AUDIX Technology Corporation, EMC Department
NVLAP Lab. Code: 200077-0
TAF Accreditation No.: 1724
Web Site: www.audixtech.com

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.