



HCV Test Controls

REF 90-1137

Read this Instructions for Use and the INSTI® HCV Antibody Test Instructions for Use before using this product. Conformance with the test procedure is necessary to ensure accurate results. Before performing the test, all operators must become familiar with Universal Precautions for Prevention of Transmission of Human Immunodeficiency Virus, Hepatitis B Virus, and other Blood-borne Pathogens in Health Care Settings¹

NAME AND INTENDED USE

The INSTI® HCV Test Controls are intended to be used as quality control for the performance of the INSTI® HCV Antibody Test only.

SUMMARY

INSTI® HCV Positive and Negative Controls should be used in conjunction with Good Laboratory Procedures. They should be run under the following circumstances:

- for new INSTI operator verification prior to performing testing on patient specimens.
- when switching to a new lot number of INSTI test kits.
- whenever a new shipment of kits is received.
- when temperature during storage of the INSTI HCV kit falls outside of designated storage temperatures.
- when the temperature of the test area falls outside of 15°-30°C (59°-86°F).
- at regular intervals as determined by the user facility.

PRINCIPLES OF THE PROCEDURE

The INSTI HCV Test Controls have been designed for use with the INSTI HCV Antibody Test to validate the correct performance of the test procedure in the hands of the operator.

The INSTI HCV Positive Controls are prepared from inactivated human plasma. It is negative for HBsAg and anti-HIV by U.S. FDA licensed test procedures.

The Positive Controls have been designed to produce an easily visible but faint blue colour on the INSTI test spot and a darker blue colour on the control spot.

The INSTI HCV Negative Control is prepared from human plasma or serum which is negative for Anti-HIV-1 and Anti-HIV-2, HBsAg, and Anti-HCV. The Negative Control will produce a blue colour on the procedure control spot, but no colour on the test spot, for a Non-Reactive INSTI test result.

REAGENTS:

HCV POSITIVE CONTROL

1 vial containing 1.0 mL of inactivated human plasma. The source material has been heat inactivated at 60°C for 60 minutes. Each vial is sufficient for 20 INSTI tests.

NEGATIVE CONTROL

1 vial containing 1.0 mL of human plasma or serum substitute, non-reactive for antibodies to HIV and HCV and non-reactive for HBsAg. Each vial is sufficient for 20 INSTI tests.

WARNINGS & PRECAUTIONS

- For *in vitro* diagnostic use only. **IVD**

Safety Precautions:

- All specimens should be handled as if capable of transmitting infectious agents.
- Thoroughly wash hands after handling or performing this test.
- Do not smoke, eat, or drink in areas where specimens or reagents are being handled.
- Wear disposable gloves while handling reagents or specimens. Do not pipette by mouth.
- Avoid contact with skin and eyes. If contact occurs, wash affected areas with water.
- Avoid forming aerosols.
- Dispose of all specimens and materials used to perform the test in a biohazard waste container. The preferred method of disposal is sterilization by autoclaving for a minimum of one hour at 121°C. Disposable materials may be incinerated. Liquid waste may be mixed with sodium hypochlorite (bleach) in volumes such that the final mixture contains 1.0% sodium hypochlorite (using a freshly prepared solution containing 10% household bleach). Allow at least 60 minutes for decontamination to be completed. **Do not autoclave solutions that contain bleach.**
- Spills should be cleaned up and decontaminated in accordance with the user facility's established procedures for handling biohazardous spills.
- For additional information on bio-safety refer to "Universal Precautions for Prevention of Transmission of Human Immunodeficiency Virus, Hepatitis B Virus, and other Blood-borne Pathogens in Health Care Settings"¹ and "Updated U.S. Public Health Service Guidelines for the Management of Occupational Exposures to HBV, HCV, and HIV and Recommendations for Postexposure Prophylaxis"².

Handling Precautions:

- Do not use INSTI HCV Test Controls beyond the expiration date.

INDICATIONS OF REAGENT INSTABILITY OR DETERIORATION

Positive or Negative Controls that are visibly turbid and/or contain particulate matter should not be used and should be discarded in accordance with safety precautions.

STORAGE INSTRUCTIONS

- The INSTI HCV Test Controls are shipped without temperature control. Upon receipt, store frozen at -20°C (-4°F) or refrigerate at 2°C (36°F) to 8°C (46°F) until expiry date indicated on the outer label.
- Once the controls are thawed, continue storage at refrigeration temperatures (2-8°C). They remain stable until the expiry date indicated on the outer label. Do not re-freeze once the vials have been opened.

- The shelf life is dictated by the printed expiry date on the INSTI HCV Test Controls.
- It is recommended to store the vials in an upright position.
- Fluctuations in temperature causing freeze-thaw cycles may affect the performance of the INSTI Test Controls.

PROCEDURE

Materials Required but not Provided

- Pipette capable of delivering 50 µL of specimen.
- INSTI HCV Antibody Test Instructions for Use.

Instructions for use

- Read the INSTI HCV Test Controls Instructions for Use prior to using the INSTI HCV Test Controls.
- Remove from storage at 2 - 8°C (36 - 46°F) or -20°C (-4°F) and allow the Controls to reach room temperature before testing with INSTI. Return Controls to refrigeration storage at 2-8°C after use.
- Mix the Controls by gentle inversion before use.
- Uncap the HCV Positive or Negative Control vial. Using a 50 µL pipette, collect 50 µL of the Control.
- Transfer the Control sample held in the pipette to the INSTI Sample Diluent vial (Solution 1). Recap the vial and mix by inversion.
- Follow the INSTI test procedure as described in the TEST PROCEDURE section of the INSTI HCV Antibody Test Instructions for Use.
- All Controls should be tested in the same manner as patient samples.
- The HCV Positive Control and the Negative Control are to be run on separate Membrane Units.

INTERPRETATION OF RESULTS

Follow the interpretation guidelines provided in the INTERPRETATION OF RESULTS section of the INSTI HCV Antibody Test Instructions for Use.

- Reactive Result:** Both the control spot and the test spot show blue colour.
- Non-Reactive Result:** Only the control spot shows blue colour.
- Invalid Result:** The test is invalid if any of the following occurs:
 - There is no blue colour on both the control spot and the test spot
 - There is blue colour on the test spot but not on the control spot
 - Uniform tint across the membrane
 - Only blue specks appear on the membrane

LIMITATIONS OF THE PROCEDURE

The INSTI HCV Test Controls are only validated for use with the INSTI HCV Antibody Test.

- The TEST PROCEDURE and INTERPRETATION OF RESULTS sections of the INSTI HCV Antibody Test Instructions for Use must be adhered to when testing the INSTI HCV Test Controls.
- Deviations from the procedure outlined in the INSTI HCV Antibody Test Instructions for Use may produce unreliable results.
- Do not dilute the INSTI HCV Test Controls. The INSTI HCV Test Controls are intended for use in undiluted form.
- Adverse shipping and storage conditions or use of expired reagents may produce erroneous results.

EXPECTED RESULTS

The HCV Positive Control must be Reactive with INSTI, and the Negative Control must be Non-Reactive with INSTI. Controls that produce incorrect or invalid results must be re-tested with INSTI.

Contact bioLytical Laboratories' Technical Support if the INSTI HCV Test Controls do not produce the expected results.

REFERENCES

- CDC Universal precautions for prevention of transmission of human immunodeficiency virus, hepatitis B virus, and other bloodborne pathogens in health care settings. *MMWR* 1988; 37(24):377-388
- CDC Updated U.S. Public Health Service Guidelines for the Management of Occupational Exposures to HBV, HCV, and HIV and Recommendations for Postexposure Prophylaxis. *MMWR* 2001; 50(RR-11):1-42.

GLOSSARY OF SYMBOLS



Store at -20°C or 2-8°C



In Vitro diagnostic medical device



Consult Instructions for use



Catalogue Number



Do not use if Damaged



Manufacturer



Caution
Harmful if swallowed
Infection risk



Use By Date



Control material

Manufactured by:



bioLytical Laboratories, Inc.
406 - 13251 Delf Place
Richmond, BC,
V6V 2A2, Canada
Phone: +1 604-204-6784
Fax: +1 604-244-8399
www.biolytical.com

51-1555A 07-Nov-2024
© Copyright 2024, All rights reserved.