

检测报告
Certificate of Analysis

材料牌号 (Brand)	MBA-40E 0175-1506	材料名称 (Description)	熔喷聚乳酸无纺布 Meltblown Polylactic acid Nonwoven
发货日期(Inspection Date)	2024/3/27	材料批号 (LOT No.)	25022978

检测项目Items	单位 Unit	执行标准Method/Standard	测试条件 Test Condition	检测结果 Results
克重 (Basis weight)	g/m ²	GB/T 24218	10cm*10cm	39.3
纵向拉力(MD Tensile)	N	GB/T 24218	100mm/min	16.5
横向拉力(CD Tensile)	N	GB/T 24218	100mm/min	10.6
颗粒物过滤效率效果(PFE (Particle Filtration Efficiency))	%	YY 0469-2011	gas flow: 32L/min	96.90
细菌过滤效率 (BFE (Bacterial Filtration Efficiency))	%	YY 0469-2011	gas flow: 32L/min	99.5
压差 (Differential Pressure)	Pa/cm ²	YY 0469-2011	gas flow:8L/min	35.3
外观(Appearance)	---	KFB202A.277.A3-2013	-	合格
结论 (Conclusion) : 合格 (OK)				

附件(Accessory):/

检测员 (Inspector) : 谢炜

审核 (Checker) : 刘显曦

