



# User Manual

# SST1 2 & 3-Year

# Model

# Single Gas Device

**SST1-2-3-YEAR-MAN-ENG-v1.2**  
**28.10.2025**

# Contents

## Safety Information - Read First

### 1. Introduction

- [1.1. SST12 and 3-Year Model Product Description](#)
- [1.2. Configuration Mobile Application](#)
- [1.3. Sensor Poisons and Contaminants](#)

### 2. Overview of Detector

- [2.1. How to Tap a Unit](#)
- [2.2. Overview of the SST1](#)
- [2.3. Display Elements of the SST1](#)
- [2.4. Pushbutton](#)

### 3. Activating the Detector

- [3.1. Start-up Process](#)
- [3.2. Normal Operating Detector Mode](#)
- [3.3. Operating Life Detector Mode](#)
- [3.4. 45 Days or Less Remaining](#)
- [3.5. 24 Hours or Less Remaining](#)
- [3.6. End of Operating Life](#)

### 4. Alarms Indicators and Warnings

- [4.1. Low Alarm](#)
- [4.2. High Alarm](#)
- [4.3. Over-Limit Alarms](#)
- [4.4. Under-Limit Alarms](#)
- [4.5. Factory Alarm Setpoints](#)
- [4.6. Compliance and Non-Compliance](#)
- [4.7. Compliance Interval Exceeded](#)
- [4.8 Calibration Interval Exceeded](#)
- [4.9 Bump Test Interval Exceeded](#)
- [4.10 Alarm Occurred](#)
- [4.11 Zero the Sensor](#)
- [4.12 Automatic Zero Reminder for O2 Models](#)
- [4.13 Zero Procedure](#)

### 5. User Menu

- [5.1. Peak Reading](#)
- [5.2. Real-time Clock](#)
- [5.3. Low Alarm](#)
- [5.4. High Alarm](#)
- [5.5 Clear Peak](#)

### 6. Communicating With the SST1

- [6.1. Communicating With the SST1 in Hibernation](#)

### 7. Device Link

- [7.1. Device Link Main Screen](#)
- [7.2. Device Link Assign/Unassign Function](#)
- [7.3. Device Link Event Logs](#)
- [7.4 Device Link Single Device Setup](#)
- [7.5 Device Link Saved Configurations](#)
- [7.6 Device Link Unassign](#)
- [7.7 Device Link Hibernation](#)
- [7.8 Device Link Maintenance](#)

### 8. Bump Test and Calibration

- [8.1. Manual Bump Test](#)
- [8.2. Manual Calibration](#)

### 9. Firmware Upgrade

### 10. Troubleshooting

### 11. Failure Codes

### 12. Unlocking the SST1 Detector

- [12.1. Unlock](#)
- [12.2. Reset Fault](#)

### 13. SST1 Spare Parts and Accessories Lists

- [13.1. Detector Accessories – External Filter](#)

### 14. SST1 Maintenance

- [14.1. Detector Maintenance – Battery Replacement](#)
- [14.2. Detector Maintenance – Internal Filter](#)

## Replacement

### 14.3. Detector Maintenance – Sensor Replacement

## **15. Certifications and Approvals**

## **16. Contact Details**

## Safety Information - Read First

To ensure personal safety, read **Warning** and the **Cautions** before using the detector.

The SST1 Gas Detectors are personal life safety devices designed to detect the presence of certain toxic gases. The accuracy of ambient gas reading(s) is dependent upon factors such as accuracy of the calibration gas standard used for calibration and frequency of calibration. WatchGas recommends performing a calibration at least once every 180 days (6 months). For safety reasons, users need to be properly trained in the use of the equipment and appropriate actions in the event of an alarm condition.

Use the detector only as specified in this user's manual and the operator's manual, otherwise the usage and protection provided by the detector may be impaired.

Read the **Warning** and **Cautions** below before using the detector.

### **Warning**

- Do not attempt to replace the internal components. This will void the intrinsic safety rating and will void the warranty of the product. Unless replaced with original WatchGas parts.
- Periodically test the response of the sensor by exposing the detector to a target gas concentration that exceeds the alarm setpoint. Manually verify that the audible, vibration and visual alarms are activated.
- Ensure the monitor is switched on, sensor and audible port are clean prior to use.

### **Cautions**

- All inspection should be performed in a clean and hazardous-free environment.
- The detector can be cleaned with a soft damp cloth using a neutral cleaner (e.g ACL Staticide). **NOTE:** Do not use solvents, soaps, or polishes.
- Bump test the response of the sensor by exposing the detector to a target gas concentration that exceeds the alarm setpoint. Manually verify that the audible, vibration and visual alarms are activated.
- This product is a gas detector, not a measuring device.
- Ensure the monitor is switched on, sensor and audible port are clean prior to use.
- For optimal performance, periodically zero the sensor in a normal atmosphere (20.9% v/v O<sub>2</sub>) that is free of hazardous gas.
- Activate the detector one year after purchase at the latest, or validate battery function.
- The equipment is intended for use in hazardous atmosphere in air with normal oxygen content not exceeding 21% v/v.
- The end-user shall contact the equipment manufacturer if the antistatic film on the LCD display is damaged.

## 1. Introduction

### 1.1 SST1 2 and 3-Year Model Product Description

The SST1 single gas detectors provides up to three years maintenance-free operation and is available in H2S, SO2, CO, or O2 models. With its simple design it is perfect for users that are looking for ultimate protection while being easy to use, durable, and cost-effective. Simply turn on the device and it runs 24/7 for 2 to 3 years (based on 2 minutes of alarm per day) — no calibration, no battery replacement, no charging, and no stress.

The device is easy to use and reduces the need for extensive training. The SST1, coupled with the SST Dock, allows for the fastest bump testing on the market and makes fleet management extremely simple.

### 1.2 Configuration Mobile Application

The detector is configured through the WatchGas Device Link Application. The app can be downloaded from [Google Play](#) and the [App Store](#).



## 1.3 Sensor Poisons and Contaminants

Several cleaners, solvents, and lubricants can contaminate and cause permanent damage to the SST1 sensors. Before using cleaners, solvents, and lubricants in close proximity to the SST1 detector sensors, read the following cautions and refer to the lists below.

### Caution

Use only the following WatchGas Technologies recommended products and procedures:

- Use water-based cleaners.
- Use non-alcohol-based cleaners.
- Clean the exterior of the detector with a soft, damp cloth.
- Do not use soaps, polishes, or solvents.

Below are common products **to avoid** using around sensors:

#### 1. Cleaners and Lubricants:

- Brake cleaners
- Lubricants
- Rust inhibitors
- Window and glass cleaners
- Dishsoaps
- Citrus-based cleaners
- Alcohol-based cleaners
- Hand sanitizers
- Anionic detergents
- Methanol (fuels and antifreezes)

#### 2. Silicones:

- Silicone cleaners and protectants
- Silicone-based adhesives, sealants, and gels
- Hand/body and medicinal creams containing silicone
- Tissues containing silicone
- Mold releasing agents
- Polishes

#### 3. Aerosols:

- Bug repellents and sprays
- Lubricants
- Rust inhibitors
- Window cleaners

## 2. Overview of Detector

Configuration of the SST Range products, including the SST1, is done with the WatchGas Device Link Application. The app can be downloaded from [Google Play](#) and the [App Store](#), or by scanning the appropriate QR code in [Section 1.2](#). You can read more about the app [on our website](#).




For detailed information on the Device Link Application, refer to the [Device Link manual](#).

### 2.1 How to Tap a Unit

Ensure that your phone has NFC capability.

The NFC is located on the back of the detector - just below the alligator clip.



Locate the NFC on your phone (normal locations shown above). The phone gives a clear visual indication when the NFC connects.

## 2.2 Overview of the SST1




Figure 6. Overview of the SST1

Symbols	Description
1	Liquid crystal display (LCD)
2	Visual alarm bars
3	Audible alarm
4	Screen scratch protectors
5	Gas identification
6	Pushbutton
7	Sensor
8	Housing
9	Alligator Clip
10	Serial number
11	NFC antenna
12	Certification label

Table 1. Overview of the SST1

## 2.3 Display Elements of the SST1

Symbols	Description
 <b>ALARM</b>	Alarm Notification
<b>LOW</b>	Low Alarm
<b>HIGH</b>	High Alarm
<b>STEL</b>	Short Term Exposure Alarm
<b>TWA</b>	Time Weighted Average Alarm
<b>H<sub>2</sub>S</b>	Hydrogen sulfide Sensor
<b>SO<sub>2</sub></b>	Sulfur dioxide Sensor
<b>CO</b>	Carbon monoxide Sensor
<b>O<sub>2</sub></b>	Oxygen Sensor
 <b>LOCK</b>	Unit Expired / Circuit Fault
 <b>!</b>	Compliance / Diagnostic Failure
 <b>PEAK</b>	Peak alarm seen last 24 hours
 <b>⌚</b>	Product Compliant
 <b>⌚</b>	Product Non-Compliant
<b>CAL</b>	Calibration Required
<b>BUMP</b>	Bump Required
 <b>NFC</b>	In Communication
 <b>!</b>	Button prompt
 <b>%</b>	Percent Volume (O <sub>2</sub> )
<b>PPM</b>	Parts Per Million (H <sub>2</sub> S / CO / SO <sub>2</sub> )
<b>MG/M3</b>	Milligrams per Cubic Meter
 <b>⌚</b>	Real time clock
<b>MM</b>	Months remaining on Device
<b>DD</b>	Days remaining on Device

**Table 2.** SST1 Display Elements

## 2.4 Pushbutton



**To activate the SST1 detector**, press and hold the pushbutton (the big yellow button on the front of the device) and a 5 second countdown will start. Keep pressing the pushbutton until the countdown ends, after which the detector will then turn on.

**To view the peak (maximum) readings**, press the pushbutton twice.

**To clear the peak readings**, press the pushbutton when the LCD displays CLR.

**To activate the NFC**, press the pushbutton for 1,5 seconds until a single beep and the NFC  icon is active.

### 3. Activating the Detector

#### Caution

Only activate the detector in a fresh air environment (with 20.9% oxygen concentration) and in a safe area that is free of any hazardous gases.

#### 3.1. Start-Up Process

Press and hold the pushbutton on the SST1 device until a 5-second countdown appears on the display. Continue holding the button until the countdown reaches zero.

When the countdown is complete, the LCD screen and LEDs will cycle on and off.

The alarm setpoints will then be visibly displayed on the LCD screen.

A countdown for the sensor warm-up will start, with a duration of 20 seconds for Toxic sensors (CO, other gasses) and 2550 seconds (42.5 minutes) for Oxygen sensors, due to it being a biased sensor.

The SST1 detector will then activate its LCD display, LEDs, buzzer, vibrational motor, and internal diagnostics.

Depending on the SST1 model, the detector will proceed to indicate its remaining operational life.

When the start-up has been completed, the detector will remain turned on. Turning the detector off is not possible by means of pressing the pushbutton on the SST1 device. To turn off the detector, refer to the [Hibernation section](#).

#### 3.2 Normal Operating Detector Mode

When the SST1 detector is operating in Detector Mode, the remaining lifetime will be prominently displayed on the LED screen. Simultaneously, the detected concentration of the gas will be continuously shown until any interruption occurs, such as a pushbutton action, a gas alarm triggering, or an error event.




#### 3.3 Operating Life Detector Mode

In Detector Mode, the LED will display the SST1 device's remaining operating life in months. The SST1 detector life is either 24 or 36 months, depending on the model.




## 3.4 45 Days or Less Remaining

When the operating life countdown reaches 45 days or less before expiry, the SST1's display format changes to show the remaining operating life in days. This countdown continues until it reaches 24 hours or less before expiry.




## 3.5 24 Hours or Less Remaining

When the operating life countdown reaches 24 hours or less before expiry, the SST1's display format changes to show the remaining time in hours. Additionally, the warning icon, clock icon, non-compliance symbol, along with a brief audible/visual warning will be provided to alert the user to this.



## 3.6 End of Operating Life

When the operating life countdown ends, the SST1 detector deactivates, and its safety functions are disabled. However, event logs can still be retrieved for a limited time after expiry. The "LOCK" warning message and non-compliance symbol will be displayed for up to 30 days after expiry to ensure continued awareness of the expired status. Re-activation is allowed under some circumstances, refer to the [Unlocking the SST1 Detector section](#).



## 4. Alarms Indicators and Warnings

A gas alarm is triggered when the sensor detects a gas concentration surpassing the predefined alarm setpoints. The alarm remains active until the gas concentration returns to a safe level.

**NOTE:** battery life decreases rapidly during alarm conditions due to heightened power consumption.

### 4.1 Low Alarm

When the gas concentration exceeds the **Low alarm setpoint** it will display LOW and produce the alarm as per below (1x per second):




### 4.2 High Alarm

When the gas concentration exceeds the **High alarm setpoint** it will display HIGH and produce the alarm as per below (2x per second):



### 4.3 Over-Limit Alarms

When the gas concentration exceeds the sensor range it will display OL and produce the alarm as per below (2x per second):



### 4.4 Under-Limit Alarms

When the gas concentration is below the sensor range it will display -OL and produce the alarm as per below (1x per second):



## 4.5 Factory Alarm Setpoints

**NOTE:** Standard factory alarm setpoints may vary by region.

Gas	Low	High
H <sub>2</sub> S	10 ppm	15 ppm
CO	35 ppm	200 ppm
SO <sub>2</sub>	5 ppm	10 ppm
O <sub>2</sub>	19.5% Vol	23.5% Vol

**Table 3.** Factory Alarm Setpoints


## 4.6 Compliance and Non-Compliance

The SST1 will indicate device compliance by showing the happy smiley icon (😊) , and flash the green compliance LED every 5 min, if enabled. The detector will show as compliant when below items are true:

- No active gas alarm
- No gas alarm occurred in last 24 hours
- No faults, failed calibration or failed bump test
- Calibration interval not exceeded (if enabled)
- Bump test interval not exceeded (if enabled)
- Compliance interval not exceeded (if enabled)
- Remaining lifetime more than 24 hours (Detector mode only)



The detector will indicate non-compliance, by means of indicating the sad smiley (😢), showing the 🔞 symbol, and flashing its red LEDs 2x per 30 seconds if one or more of the below situations.



## 4.7 Compliance Interval Exceeded

If enabled, the compliance interval serves to remind users to regularly connect the SST1 detector to a Device Link, SST Kiosk, or SST Dock. When the compliance interval has been exceeded, the detector will indicate non-compliance by displaying the **▲** icon.



## 4.8 Calibration Interval Exceeded

If enabled, the calibration interval is designed to prompt users to conduct an SST1 detector calibration. Upon exceeding the calibration interval, non-compliance indicators will be displayed, accompanied by the **CAL** icon in the lower left corner.

A successful calibration (either manual, or by using the SST Dock) will reset the calibration interval and clear non-compliance.



## 4.9 Bump Test Interval Exceeded

The bump test interval is designed to prompt users to conduct an SST1 detector calibration. Upon surpassing the bump test interval, non-compliance indicators will be displayed, accompanied by the **BUMP** icon in the lower left corner.

A successful bump test or calibration (either manual, or by using the SST Dock) will reset the bump test interval.



## 4.10 Alarm Occurred

The peak readings symbol is displayed when the sensor is exposed to a gas concentration that exceeds the alarm setpoints. It is no longer displayed when more than 24 hours have passed since the last alarm, when the SST1 detector is connected to a SST Kiosk or Device link, or when a successful bump test/calibration is performed via an SST Dock, or manually.

The highest peak reading can be displayed on the SST1 detector by pressing the pushbutton once. The peak can be cleared by pressing and holding the pushbutton for 3 seconds, after which the detector will return to the next screen.



## 4.11 Zero the Sensor


Over time and with usage, the sensor baseline may drift from the original baseline. To ensure optimal performance, WatchGas recommends zeroing the O<sub>2</sub> sensor once every 24 hours, or when the Automatic Zero Reminder is displayed for O<sub>2</sub> models.

For all other models, WatchGas recommends periodic zeroing of the sensor of at least once every 180 days (6 months).

## 4.12 Automatic Zero Reminder for O<sub>2</sub> Models

For **SST1 O<sub>2</sub> models exclusively**, the Automatic Zero Reminder appears when more than 24 hours have passed since the last successful zero procedure was conducted.

It is highly recommended that users perform a Zero Calibration when prompted by the Zero Reminder, in order to ensure accurate readings and optimal device performance.





## 4.13 Zero Procedure

The Zero Procedure can be performed on SST1 O<sub>2</sub> (both detector and monitor mode) and toxic SST1 models (monitor mode only). Here's how to perform the Zero Procedure:

1. Transition to a standard atmosphere containing 20.9% v/v O<sub>2</sub>, ensuring it is free of hazardous gases.
2. Press and hold the pushbutton for 5 seconds until ZERO is displayed, indicating that the Zero Procedure has initiated.
3. Continue to hold down the pushbutton until the 10-seconds countdown has been completed.
4. Upon completion of the countdown, the Zero Procedure has been completed and the result will be displayed.
5. If the Zero Procedure is successful, "PASS" is displayed. The detector will return to measurement mode.
6. If the Zero Procedure fails, the detector will beep and flash the non-compliance LED. It will return to measurement mode, toggling FAIL and display the reading or months remaining. Non-compliance icons are displayed.
7. Repeat the Zero Procedure. For Toxic models, a failed Zero Calibration requires a successful Zero and Span Calibration. If the procedure fails again, contact WatchGas for further assistance.

Repeated failed Zero Calibration will **Lock the unit**.



## 5. User Menu

The SST1 features a user menu that can be entered by pressing the pushbutton once. Upon pressing the pushbutton the next menu screen will be displayed in the following order:

### 5.1 Peak Reading



If the detector has measured a gas concentration above the alarm setpoint within the last 24 hours, the peak reading will be displayed.

Pressing and holding the pushbutton for 3 seconds will clear the peak.

Alternatively, the Peak can be cleared using Device Link or SST Dock.

### 5.2 Real-time Clock



When enabled, the detector will display its current time in a 24-hour format.

The Real-time Clock can be enabled/disabled using Single Device Setup.

Device time can be synchronized using Device Link.

### 5.3 Low Alarm




The detector will display its Low alarm setpoint.

### 5.4 High Alarm



The detector will display its High alarm setpoint.

## 5.5 Clear Peak



The Clear option allows users to clear Peak.

To initiate the Clear, press and hold the pushbutton for 5 seconds. When the data is cleared, the happy smiley icon will be displayed, indicating that the process is completed.

When the pushbutton is pressed again, the SST1 detector will return to measurement mode.

**NOTE:** the SST1 detector will automatically return to measurement mode if no pushbutton is pressed, or when a gas alarm occurs.

## 6. Communicating With the SST1


The SST1 gas detector is equipped with Near Field Communication (NFC). NFC is used to communicate with:

- [WatchGas Device Link](#)
- [WatchGas SST Dock](#)
- [WatchGas SST Kiosk](#)

The NFC chip is located on the back of the detector, under the alligator clip.

To activate NFC, press and hold the pushbutton for 3 seconds until the SST1 detector beeps once. The NFC icon is now displayed in the upper right corner of the device. The SST1 detector is now able to communicate with the WatchGas Device Link, SST Dock, SST Kiosk, and RTR Software.

The NFC will automatically deactivate after a period of 300 seconds.




### 6.1 Communicating With the SST1 in Hibernation

While the SST1 detector is in Hibernation (turned off), the NFC communication can be activated by pressing and holding the pushbutton for 3 seconds. The SST1 detector will display SET, as well as the NFC icon.

The SET mode allows for configuration without the need for powering on the SST1 detector.

The NFC will automatically deactivate after 200 seconds, or by pressing the pushbutton once, after which the SST1 detector will return to hibernation.




## 7. Device Link

The [WatchGas Device Link mobile application](#) can be used to communicate with the SST1 detector and verify detector compliance.

### 7.1 Device Link Main Screen

To connect the SST1 detector to the Device Link:

1. Press and hold the pushbutton for 3 seconds to activate NFC.
2. On the Device Link, press “Tap to activate NFC”.
3. Connect the NFC chip located on the back of the detector to the NFC antenna of the mobile phone.
4. Device Link will show “Compliant” when the SST1 detector is fully compliant.
5. If one or more items are non-compliant, Device Link will show “Not Compliant” and indicate which item is non-compliant.



**NOTE:** Upon connecting with the Device Link, the Compliance Interval will be reset, Peak will be cleared, and the Event Log will be transferred.

### 7.2 Device Link Assign/Unassign Function

The SST1 detector can be designated to a specific user and site through the Device Link mobile application. Enabling the “Assign Check Function” in the Application Settings triggers the Device Link to prompt users to assign any unassigned SST1 detectors to a user and site.

When an unassigned SST1 connects to the Device Link, the user is prompted to input the User ID and Site information. Once entered, these settings can be transmitted to the SST1 using NFC, effectively assigning the User ID and Site to the detector in question.

The Device Link will skip the User ID assignment prompt if the detector already has a User ID assigned.

Alternatively, users can set the User ID and Site using the Single Device Setup feature, detailed in the [Device Link Single Device Setup section](#).

Assign Unit

User ID

Location

 Tap to activate NFC

## 7.3 Device Link Event Logs

The SST1 detector Event Logs, which are transferred upon connecting to the Device Link, can be viewed in the “Event Logs”.

To view an Event Log:

1. Log in using password: user
2. Go to “Event Logs”
3. Select the Event Log to view it

To export an Event Log:

1. Log in using password: user
2. Go to “Event Logs”
3. Select the Event Log file to export (**NOTE:** multiple Event Log files can be selected)
4. Press “Share”
5. Select app to share the Event Log(s) to

Event log

Sort by

Date    Detector type    Serial number

Select all

- SST1-1001 (H<sub>2</sub>S) monitor  
2024.02.15 - SN: 5711330018
- SST1-1004 (CO) monitor  
2023.06.20 - SN: 5741340017
- SST1-1001 (H<sub>2</sub>S) monitor  
2023.07.05 - SN: 5711364577
- SST1-1003 (SO<sub>2</sub>) monitor  
2024.02.09 - SN: 5731330017

When the RTR compliance software is activated, event logs are automatically shared to it.

The SST1 logs the following events in its Event Logs:

- Clearing alarms
- Bump test results
- Calibration results
- Compliance checks, including instances where a non-compliant detector is confirmed by the Device Link User.

SST1-1001 (H <sub>2</sub> S) monitor	
NOT COMPLIANT Not confirmed	15.02.2024, 18:36:12
NOT COMPLIANT Confirmed	15.02.2024, 16:55:36
Bump Test Fail	15.02.2024, 12:57:42
Bump Test Fail	15.02.2024, 12:40:33
Calibration Pass	11.08.2023, 11:20:30
Clear alarm High 217.8 ppm	19.06.2023, 18:11:48
ALARM High 217.8 ppm	19.06.2023, 18:11:48

**100 Event Logs** can be stored in the SST1 detector. When full, the oldest Event Log entry will be overwritten with the newest one.

**NOTE:** Event logs cannot be cleared from the Device Link.

## 7.4 Device Link Single Device Setup

The Single Device Setup in the WatchGas Device Link mobile application can be used to configure the SST1 detector.

The Single Device Setup can be accessed through the Device Link menu, with the password “service”.

This User Options section includes SST1 detector features that can be enabled or disabled. The green checkmark indicates that the option is enabled. Press the checkbox to disable the option.


### Device Identification

- **Model:** SST1 Model type
- **Sensor:** SST1 Sensor type
- **User ID:** User ID assigned to the detector.
- **Location:** Site assigned to the detector.
- **Serial number:** Detector serial number (non-changeable)
- **Firmware version:** Detector firmware version
- **Time:** Detector time
- **Switched on for:** Time the detector has been switched on



### Function Setting

- **Real Time Clock:** Switch on/off displaying the real-time clock on the detector.
- **Green Compliance:** If turned on, a compliant detector will flash a green LED every 5 minutes.
- **Self-test interval:** Interval after which the user is prompted to perform a manual self-test
- **Scan protect:** Enable NFC pin to protect the detector against connecting with an unauthorized Device Link
- **Simple Mode:** When enabled, this option hides the peak indicator and smiley icons from the screen display.
- **Silent Mode:** This function mutes sounds and turns off LED, and only uses the vibrator to indicate alarms.



## Calibration

- **Enable Expiry:** Enable/disable calibration interval.
- **Calibration interval:** Calibration interval in days.

Calibration

ⓘ Enable Expiry

ⓘ Calibration Interval
 days

## Bump

- **Enable Expiry:** Enable/disable calibration interval.
- **Calibration interval:** Calibration interval in days.
- **Bump speed:** Set % of gas applied required to pass bump test.
  - Fast: 50%
  - Medium: 70%
  - Slow: 90%

Bump

ⓘ Enable Expiry

ⓘ Bump Interval
 days

ⓘ Bump speed

▼

**NOTE:** the Bump speed setting does not change the allowed time for a bump test

## Sensor Settings

- **Sensor unit:** unit of measurement
  - Ppm: Parts per Million (default for toxic models)
  - Mg/m3: milligram per cubic meter (toxic models)
  - Vol%: Volume percent (O2 model only)
- **Alarm Low:** Low alarm setpoint
- **Alarm High:** High Alarm setpoint

Sensor setting 1

Sensor CO 1000
 ppm

alarm Low

35.0

alarm High

200.0

alarm STEL

100.0

alarm TWA

20.0

## Health Data

In the Single Device Setup, various parameters indicating detector health can be viewed:

- **Ppm\*hours:** accumulated gas exposure, indicating sensor health.
- **Peak reading:** All-time peak reading registered.
- **Factory calibration:** sensor output during factory calibration
- **Last calibration:** sensor output during last field calibration
- **Sensor SN:** Gas sensor serial number
- **Last calibration date:** Date on which the last successful calibration was performed
- **Last bump date:** Date on which the last successful bump test was performed.
- **IOELV and Daily Peaks:** Press "View" to show the Indicative occupational exposure limit values (IOELV) and daily peaks of the device.

Health data

PPM/hours: 342

Peak reading: 25.1 ppm

Factory calibration: 570.7 ppm

Last calibration: 517.0 ppm

Sensor SN: AA321T0002

Last calibration date: May 31 2023 17:48

Last bump date: May 31 2023 17:48

## Unit Data

In Single Device Setup various parameters related to the detector can be viewed:

- **Switched on for:** total time the SST1 detector has been switched on
- **Min temp:** all-time minimum temperature
- **Max temp:** all-time maximum temperature
- **Battery run time:** total battery runtime
- **Battery SN:** Battery serial number
- **In alarm:** Total minutes the detector has been in alarm

Unit data

Switched on for 259 days 17 hours 36 minutes
Min. temp: 6.7 C° 44.1 F°
Max. temp: 39.7 C° 103.5 F°
Battery run time: 260 days 5 hours 38 minutes
Battery SN: FEB230068
In alarm: 3 minutes

## 7.5 Device Link Saved Configurations

Saved Configurations in the WatchGas Device Link Application can be used to create saved configuration files that can be used to quickly and easily configure multiple detectors.

### Creating Saved Configuration Files

To create a saved configuration file:

1. Log in to the Device Link using password: service
2. Navigate to “Saved Configurations.”
3. Press the “+” icon.
4. Select the SST1 model.
5. Provide a file name and description.
6. Input the configuration settings.
7. Press the “Save” button.

File Function

File name	SST1-1004 (CO) monitor
Description	Default OSHA
<input checked="" type="checkbox"/> Use for compliance <input checked="" type="checkbox"/>	

### Use For Compliance

By selecting this option, the Device Link will utilize the saved configuration file to verify detector alarm setpoints. When a detector with mismatched alarms is linked to the Device Link, the user will be prompted to adjust the alarm setpoints.

### Loading a Saved Configuration to the SST1 Detector


1. Log in to the Device Link using password: service
2. Access “Saved Configurations” and choose the desired configuration file.
3. Manually verify all settings for accuracy.

4. Press the “Save” button.

5. Press “Load settings to device.”

6. Activate NFC on the detector.

7. Connect the detector to the Device Link or SST Kiosk.



**NOTE:** Loading a saved configuration to an SST1 detector also functions when the detector is in SET mode, facilitating rapid configuration updates without requiring the detector to be powered on.

## 7.6 Device Link Unassign

The Unassign function can be used to unassign an SST1 detector. It will clear the User ID in the SST1 detector. Optionally, the Site ID can be cleared as well.

To unassign an SST1 detector:

1. In Device Link, log in with password: service

2. Go to Unassign

3. Optional: select Unassign Site ID

4. Press “Tap to activate NFC”

5. Activate NFC on the detector

6. Connect the detector to the Device Link or SST Kiosk



## 7.7 Device Link Hibernation

The Hibernation function can be used to (temporarily) turn off the SST1 detector.

To enter Hibernation:


1. In Device Link, log in with password: service

2. Go to Hibernation

3. Press “Tap to activate NFC”

4. Activate NFC on the detector

5. Connect the detector to the Device Link or SST Kiosk



**WARNING:** An SST1 detector in Hibernation mode does not have a safety function and is unable to alarm its user when in hazardous atmosphere.


When a detector is turned on after being in Hibernation, sensor warm-up is required.

## 7.8 Device Link Maintenance

The Maintenance menu is used to diagnose and service the SST1 detector.

To enter Maintenance mode:

1. In Device Link, log in with password: service
2. Go to Maintenance
3. Press “Tap to activate NFC”
4. Activate NFC on the detector
5. Connect the detector to the Device Link or SST Kiosk



Upon connecting an SST1 detector to the Maintenance screen, the detector and component serial numbers are displayed.

Below each of the serial numbers, a star rating is displayed, indicating component health:

- **5 stars** mean that the sensor performance is healthy.
- **4.5 stars** mean that the sensor performance is ok.
- **3.5 stars** mean that the sensor performance is poor.
- **2 stars** or fewer mean that the sensor performance is weak.


**NOTE:** Once a sensor's performance falls to 3.5 stars, it needs to be replaced. Refer to the [Sensor Replacement section](#) on how to do this.

The Device serial number and PCB SN cannot be changed.

The Sensor SN and Battery SN can be changed using the “Edit” button.

Changing the Battery SN will reset the following health parameters:

- **Switched on for**
- **Battery run time**
- **In alarm**



Changing the Sensor SN will reset below health parameters:

- **Ppm\*hours**
- **Peak**
- **Factory calibration**



- **Last calibration**

**NOTE:** When changing components, component serial numbers must be changed using Device Link Maintenance, refer to the [SST1 Maintenance section](#).

**WARNING:** Only substitute with original WatchGas parts. Substitution of parts always requires a full calibration prior to use.

## 8. Bump Test and Calibration

WatchGas recommends performing a bump test with a known concentration of gas above the alarm setpoints before each day of use. And, additionally, when a detector is subjected to physical damage, dirt/debris, immersion, under limit alarm, over limit alarm, or if detector performance is in doubt in any way.

To ensure the highest detection accuracy, WatchGas recommends frequent detector calibration.

WatchGas recommends using the SST Dock to perform bump tests and calibrations on the SST1 detector.

WatchGas recommends using calibration gas for both bump tests as well as calibrations as per the table below.

Sensor	Range	Calibration Gas	Calibration Time
CO	0-2000 ppm	100 ppm CO	90 seconds
H2S	0-500 ppm	25 ppm H2S	90 seconds
O2	0-25 vol%	18 vol% O2	90 seconds
SO2	0-100 ppm	10 ppm SO2	120 seconds

**Table 4.** Calibration Gas ranges and time

WatchGas recommends using a stainless steel regulator with a fixed flow of 0.5 L/min, and non-absorbent tubing such as [Last-O-More](#).

### 8.1 Manual Bump Test

When no SST Dock is available, a manual bump test can be performed. To perform a Bump Test on the SST1 detector:

**1.** Transition to a standard ambient atmosphere with 20.9% oxygen concentration, free of any hazardous gases.

**2.** Log in using password: service

**3.** Navigate to “Bump Test.”

**4.** Optionally, add a new calibration gas cylinder.


**5.** Select the gas cylinder from the list.

**6.** Activate NFC on the SST1 detector.

**7.** Connect to the detector Device Link using NFC. A countdown timer will be displayed on both the SST1 detector and the Device Link.

**8.** Connect the gas to the detector using the original calibration cap.

**9.** Start the flow of gas and wait for the Bump Test to finish.



10. Stop the flow of gas.

11. Connect the SST1 detector to the Device Link using NFC.

The Bump Test result is displayed. A successful Bump Test will reset the Bump Test Interval (if this option is enabled).

If the Bump Test fails, retry the test using the steps above.

If the Bump Test fails repeatedly, perform a Calibration. The SST1 detector will automatically lock after 10 consecutive failed Bump Tests.

**NOTE:** The SST1 detector shown on the Device Link is only a simulated example. The real Bump Test result must be viewed on the SST1 detector itself.



## 8.2 Manual Calibration

When no SST Dock is available, a manual Calibration can be performed. To perform a Calibration on the SST1 detector:

1. Transition to a standard ambient atmosphere with 20.9% oxygen concentration, free of any hazardous gases.
2. Log in using password: service
3. Navigate to “Calibration”
4. Optionally, add a new calibration gas cylinder.
5. Select the gas cylinder from the list.
6. Activate NFC on the SST1 detector.
7. Connect to the detector Device Link using NFC. The Fresh Air Calibration will start and a countdown time will be displayed
8. Wait for the Fresh Air Calibration to finish.

9. Connect the gas to the detector using the original calibration cap.

10. Start the flow of gas and wait for the span calibration to finish.

11. Stop the flow of gas.

12. Connect the SST1 detector to the Device Link using NFC.

The Calibration result is displayed. A successful Calibration will reset the Calibration and Bump Test Interval (if this option has been enabled).

A successful Calibration will generate a Calibration Certificate.


The user will be prompted to check the visual, audible, and vibration alarms, as well as to visually inspect the SST1 detector for any damage. The results of these checks must be entered into the Device Link. Additionally, the name of the person carrying out the calibration must be recorded.

Once these steps are completed, the generated Calibration Certificate can be viewed and shared by pressing the “Next” button.

If the Calibration fails, retry the test using the steps above.

If the Calibration fails repeatedly, contact WatchGas. The detector will automatically lock after 10 consecutive failed Calibrations.

**NOTE:** The SST1 detector shown on the Device Link is only a simulated example. The real Calibration result must be viewed on the SST1 detector itself.



## 9. Firmware Upgrade

WatchGas recommends using the latest detector firmware for optimal performance and functionality.

The latest SST1 firmware is available for download at: <https://watchgas.com/datacenter/>

The latest firmware is automatically available in the Device Link or the SST Dock.

To upgrade firmware on the SST1 detector:

1. Log in using password: service
2. Navigate to “Firmware upgrade”
3. Select the correct firmware version and model.
4. Activate NFC on the detector
5. Connect the SST1 detector to the Device Link, ensuring that the connection is not lost.
6. Wait for the NFC file transfer to finalize. This may take several minutes. The SST1 will then reboot upon receiving the firmware file.
7. Verify detector settings in Single Device Setup.

**NOTE:** After conducting a firmware upgrade, it is essential to always perform a configuration check and full calibration of the SST1 to ensure the system is functioning properly.

## 10. Troubleshooting

Problem	Possible Cause	Solution
Time set incorrectly	Time not synchronized	Connect the detector to the Device Link, SST Kiosk, or SST Dock.
Manual zero calibration fails	Sensor drift	Perform a full calibration.
Bump test fails repeatedly	Filter dirty (internal or external) Sensor drift	Replace filter. Perform a full calibration.
Detector displays FAIL/Err2 after warm-up	Sensor insufficiently stabilized	Put the detector in Hibernation mode, then start it again. Perform a full calibration once warm-up has completed.
Detector shows  and  , but no error message/icon is displayed	The compliance interval has been exceeded	Connect the detector to the Device Link, SST Kiosk, or SST Dock to reset its compliance interval.

**Table 5.** Troubleshooting

**NOTE:** This troubleshooting guide assumes that the SST1 detector runs on the latest firmware version.

Refer to the [Firmware Upgrade section](#) on how to upgrade the SST1 firmware version.

## 11. Failure Codes

In the table below, you can find all the failure codes, their most common causes, and how to resolve them:

Code	Reason	Action
BAT	Battery voltage too low	Replace battery
CAL	Calibration fail	Perform successful calibration
BuP	Bump Test fail	Perform successful bump test
TEST	Self-test fail	Perform manual self-test
Err5	Detector life expired (>1 month)	Replace detector
P01	Configuration error	Reconfigure unit, contact WatchGas
Err9	N/A	Unspecified error. Contact WatchGas

Table 6. Failure Codes

When a fault arises in the SST1, the detector will remain operational. The non-compliance signals will promptly alert the user to the issue. Additionally, certain faults may activate the audible alarm to further draw attention to the situation.

Repeated faults will result in the SST1 being locked. A locked SST1 is rendered inoperable and must be handled by authorized personnel, or taken to a WatchGas Service Center for servicing.

To power off a locked SST1 detector, simply press the pushbutton once.

A locked SST1 detector can be unlocked using the Device Link Maintenance, refer to the [Unlocking the SST1 Detector section](#).

Furthermore, the Lock or Fault state can be observed in Device Link Maintenance.



## 12. Unlocking the SST1 Detector

### 12.1 Unlock

When the SST1 is locked, maintenance is required. Refer to the [SST1 Maintenance section](#).

An SST1 detector that has been locked because its activation time has expired can be unlocked by pressing the “Unlock” button.

Service personnel must assess the battery condition by star rating and total minutes in alarm. If the battery's lifespan cannot be guaranteed for the additional detector lifetime, the battery must be replaced.

**NOTE:** The permitted extra lifetime is 12 months.



### 12.2 Reset Fault

After performing the necessary maintenance, a failed SST1 unit can be unlocked by pressing the “Reset Fault” button.

**NOTE:** Unlocking an SST1 unit without proper maintenance may void its safety function and/or intrinsic safety.

## 13. SST1 Spare Parts and Accessories Lists

**NOTE:** SST1 detector maintenance may only be performed by trained/authorized individuals and/or service center personnel.

**NOTE:** SST1 parts may only be replaced by original WatchGas parts, as per the spare parts lists table. Substitution of parts may impair detector function and intrinsic safety.

Part Number	Description
<b>SST1-BAT</b>	WatchGas SST1 Replacement Battery - 3.6V
<b>SST1-BAT-10</b>	WatchGas SST1 Replacement Battery - 3.6V pack of 10
<b>SST-AG-10</b>	WatchGas SST Range Replacement Alligator Clip with screw pack of 10
<b>SST1-CALCAP-10</b>	WatchGas SST1 Calibration Cap (pack of 10)

**Table 7.** SST1 Spare Parts and Accessories List

Part Number	Description	Range
<b>SST-SPARE-H</b>	WatchGas SST sensor H2S 0-500 ppm for All (Hydrogen Sulfide)	0-500 ppm
<b>SST-SPARE-M</b>	WatchGas SST sensor CO 0-2000 ppm for All (Carbon Monoxide)	0-2000 ppm
<b>SST-SPARE-O</b>	WatchGas SST sensor O2 0-25%vol. for All (Oxygen)	0-25% vol
<b>SST-SPARE-S</b>	WatchGas SST sensor SO2 0-100 ppm for All (Sulfur Dioxide)	0-100 ppm


**Table 8.** SST1 Spare Sensors List

## 13.1 Detector Accessories – External Filter

**NOTE:** Only original WatchGas parts may be used as SST1 detector accessories. Substitution of parts may impair detector function and intrinsic safety.

WatchGas strongly recommends utilizing an external filter in applications where dust, dirt, or other debris may be present.

When using an external filter, always ensure that there are no obstructions between the atmosphere and the gas sensor. Regular filter changes are highly recommended to maintain optimal performance.



## 14. SST1 Maintenance

### 14.1 Detector Maintenance – Battery Replacement

**NOTE:** the SST1 battery is non-user replaceable. The battery may only be replaced by authorized personnel and/or WatchGas service centers.

To replace the SST1 battery:

1. Turn off the unit by selecting Hibernation mode in the Device Link.
2. Use a PH00 screwdriver to loosen the four screws on the back of the unit by turning them counter-clockwise, then carefully separate the housing parts.
3. Remove the old battery from the unit. Dispose of the battery as per local regulations.
4. Note the battery batch code for the new battery.
5. Insert the new battery, ensuring that the positive (+) and negative (-) terminals align correctly with the indicators on the unit.
6. Carefully reattach the back cover.
7. Secure the unit by tightening the four screws on the back using a PH00 screwdriver, applying 0.7 N/m torque, and turning the screws clockwise.
8. In the Device Link Maintenance, input the battery batch code. If the unit is locked, unlock it.
9. Perform Calibration on the unit before putting it back into use.

**NOTE:** Changing the Battery SN will reset the following health parameters:

- Switched on for
- Battery run time
- In alarm

### 14.2 Detector Maintenance – Internal Filter Replacement

To replace the SST1 Internal Filter:

1. Turn off the unit by selecting Hibernation mode in the Device Link.
2. Use a PH00 screwdriver to loosen the four screws on the back of the unit by turning them counter-clockwise, then carefully separate the housing parts.
3. Replace the internal filter
4. Carefully reattach the back cover.
5. Secure the unit by tightening the four screws on the back using a PH00 screwdriver, applying 0.7 N/m torque, and turning the screws clockwise.

torque, and turning the screws clockwise.

**6.** Perform calibration on the unit before putting it back into use.

## 14.3 Detector Maintenance – Sensor Replacement

To replace the SST1 gas sensor:

- 1.** Turn off the unit by selecting Hibernation mode in the Device Link.
- 2.** Use a PH00 screwdriver to loosen the four screws on the back of the unit by turning them counter-clockwise, then carefully separate the housing parts.
- 3.** Replace the sensor. Note down the new sensor SN.
- 4.** Carefully reattach the back cover.
- 5.** Secure the unit by tightening the four screws on the back using a PH00 screwdriver, applying 0.7 N/m torque, and turning the screws clockwise.
- 6.** In the Device Link Maintenance, enter the new sensor SN. Input the settings into the SST1 detector.
- 7.** Perform Calibration on the unit before putting it back into use.

**NOTE:** Changing the Sensor SN will reset the following health parameters:

- Ppm\*hours
- Peak

## 15. Certifications and Approvals

### Intrinsic Safety:

II 1G Ex ia IIC T4 Ga  
I M1 Ex ia I Ma  
IECEx EXV 22.0030X ExVeritas 22ATEX1314X  
ExVeritas 22UKEX1315X

EU conformity [www.watchgas.com](http://www.watchgas.com)  
RoHS compliant

### Standard

IEC 60079-11:2013  
IEC 60079-0:2020  
EN IEC 60079-0:2018  
EN 60079-11:2012  
EN 50270:2015

### Inmetro:

ABNT NBR IEC 60079-11:2013  
ABNT NBR IEC 60079-0:2020

### North America:

UL 60079-0 7th Edition  
UL 60079-11 6th  
CSA C22.2 NO. 60079-0:19  
CAN/CSA C22.2 NO. 60079-11:14  
UL 61010-1 Safety Requirements  
CAN/CSA-C22.2 NO. 61010-1-12/A1:18  
Low Voltage Directive 2014/35/EU  
EMC Directive 2014/30/EU  
C II, Zn 0, AEx ia IIC T4 Ga  
Int.Safe: C II Dv 1, Gr A,B,C,D T4  
-40°C ≤ Tamb ≤ +60°C

## 16. Contact Details

### EMEA

Klaverbaan 121  
2908 KD Capelle a/d IJssel  
The Netherlands

[info@watchgas.com](mailto:info@watchgas.com)  
[www.watchgas.com](http://www.watchgas.com)

### Americas

313 N. State Hwy 342  
Red Oak, TX 75154, USA

[info@watchgasusa.com](mailto:info@watchgasusa.com)  
[www.watchgasusa.com](http://www.watchgasusa.com)

### APAC

Woods Square Tower 1,  
12 Woodlands Square,  
#11-71, Singapore 737715

[info@watchgas.com](mailto:info@watchgas.com)  
[www.watchgas.com](http://www.watchgas.com)

### ANZ

[aus@watchgas.com](mailto:aus@watchgas.com)

### SERVICE CENTERS:

**EMEA:** [servicesupportemea@watchgas.com](mailto:servicesupportemea@watchgas.com)

**APAC:** [servicesupportapac@watchgas.com](mailto:servicesupportapac@watchgas.com)

**NSAM:** [servicesupportnsam@watchgas.com](mailto:servicesupportnsam@watchgas.com)

**ANZ:** [servicesupportanz@watchgas.com](mailto:servicesupportanz@watchgas.com)