

Logfile example Oxipack leak tester

Serial	00-OX0000-000					
Date	Time	Parameter	Measure Jump	Measure 1	Measure 2	Result
8-5-2020	08:59:03	Standard	0	0	0	NOK (202)
8-5-2020	09:10:10	Standard	0	0	7	OK
8-5-2020	09:11:30	Standard	0	0	7	OK
8-5-2020	09:13:43	Standard	0	0	6	OK
8-5-2020	09:14:01	Standard	0	0	7	OK
8-5-2020	09:16:33	Standard	0	0	6	OK

Serial:

The serial number of the leak tester is being shown. This number is etched on the back-left corner of the leak tester.

Date:

The actual date of when the measurement was performed.

Time:

The actual time of when the measurement was performed.

Parameter:

The selected recipe which was used to perform the measurement.

Measure Jump:

The first delta of the measurement.

Measure 1:

The second delta of the measurement retrieved in the “measuring 1” phase.

Measure 2:

The third delta of the measurement, retrieved in the “measuring 2” phase

Result:

The actual result of the measurement. Variates between “OK” (good package) and “NOK” (Leaking package)