
### CT General Comments



Diagnostic & Preventative Imaging Center

1. Make reconstructions relative to the patient, not the scanner. (See Fig. A)



- 2. Position the patient to eliminate the need for a gantry tilt, when possible.
- 3. When creating sagittal and coronal reconstructions, do NOT send the additional image showing the plane of the reconstructions to PACS, if possible. If your scanner sends this image automatically, delete from PACS before marking the study as Reviewed. (See Fig. B)






Fig. B

Fig. A

- 4. All Bone Algorithm reconstructions should be created using a sharp bone algorithm, not smooth bone. This may be called Bone Plus, etc. depending on the machine.
- 5. Adult oral contrast for CT should be mixed at a ratio of 50 mL of water soluable contrast (Omnipaque 300 or equivalent water soluble contrast) per 1000 mL of liquid OR 1000mL Ready to Drink (RTD) Omnipaque Oral Solution 12mg/mL 500mL jugs (premixed contrast agent no dilution necessary). CT Barium may be utilized, if indicated.
- 6. Pediatric oral contrast for CT Please refer to CT Oral Contrast Volumes Pediatric Guidelines posted online.

Continued on next page

### CT General Comments Cont.

7. Protocols state Omnipaque for IV contrast. You may substitute other equivalent water soluble contrast brands such as those below.

GE Healthcare	Clinical Equivalent	Bracco
Omnipaque 300 mg/mL (Iohexal)	Clinical Equivalent	Isovue 300 mg/mL (Iopamidol)
Omnipaque 350 mg/mL (Iohexal)	Clinical Equivalent	Isovue 370 mg/mL (Iopamidol)

8. IV contrast for pediatric patients (17 years and younger) should be administered according to the following weight based guidelines: 1 mL/pound, up to a maximum dose of 80 mL

### CT Head



Diagnostic & Preventative Imaging Center

Lateral Canthus

EAC

Anatomical Reference: Orbitomeatal

Patient Position: Supine; Tilt the patient's head so that the orbitomeatal line is perpendicular to the

CT tabletop (see image).

Patient Orientation: Head First

Scan Range: C1 through vertex

Scout: Lateral, AP if necessary

Scan Type: Helical Rotation Time: 0.9

**Thickness:** 5mm, Recon as thin as possible **Interval:** 5mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

MAs:

None

Head

120

Auto mA

Scan Delay: 60 seconds, if contrast used

**Respiration:** None **DFOV:** 25

Recon Algorithm: See comments below

Contrast: 80 mL Omnipaque 300 mg/mL or equivalent water soluble contrast @ 1.5-3

mL/sec

Comments: 1. Send 5 mm Soft Tissue Algorithm Axial Images to PACS

2. Send 1 mm Soft Tissue Algorithm Axial Images to PACS

3. Send 2 mm Bone Algorithm Axial Images to PACS

4. Send 3mm Sagittal and Coronal Soft Tissue Algorithm Reconstructions to

**PACS** 

5. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

Without Contrast: Trauma, Stroke Symptoms, Headache, Seizure, Altered Mental Status

With and Without Contrast: Tumor, Infection AND unable to have MRI, Abnormal Non-Contrast CT

<sup>\*</sup> Image courtesy of University of Wisconsin

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

## CT Perfusion Single Slab



Diagnostic & Preventative Imaging Center

Performed at IMMC & \*ILH Only

	IMMC	ILH
Scanner:	Siemens Edge/Flash	GE Discovery CT750 HD
Detector Width:	38.4	40
Detector Rows:	32 x 1.2 mm	64
Mode:	Helical	Axial
Acq Type:	VPCT	Shuttle
Scan Delay:	5 second	5 second
Tube Voltage (kVp)	70	80
Tube Current (mA)	350	500
Exposure Time:	570	400
Exposure (mAs):	200	200
Interval (mm):	96	40
Slice Thickness:	10	5
Slice Increment:	10	5
Beam Collimation:	38.4	40
Number of Cycles:	44	24
Recon FoV:	220 mm	22 cm
Recon Kernel:	H20f Smooth	Standard

Contrast: IV: 40 mL Omnipaque 350 mg/mL or equivalent water soluble contrast at 5 mL/sec +

40 mL of Saline chase at 5 mL/sec

#### **Wait 2 Minutes**

CTA Head and Neck - Follow posted online protocol

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

# CT Perfusion Dual Slab



Diagnostic & Preventative Imaging Center

Performed at \*ILH & MWH Only

	ILH	MWH
Scanner:	Siemens Edge	Siemens AS 64
Detector Width:	38.4 - per slab	19.2 mm
Detector Rows:	32 x 1.2 mm	16 x 1.2 mm
Mode:	Helical	Axial
Acq Type:	VPCT	NeuroPCT
Scan Delay:	5 second	5 second
Tube Voltage (kVp)	80	80
Tube Current (mA)	350	85
Exposure Time:	570	1 second
Exposure (mAs):	110	85
Interval (mm):	0 mm	0 mm
Slice Thickness:	10	10
Slice Increment:	9.6	4.5
Beam Collimation:	38.4	19.2 mm
Number of Cycles:	44	65
Recon FoV:	220 mm	22 cm
Recon Kernel:	H20f Smooth	H20f Smooth

Contrast: IV: 1st Perfusion Run: 40 mL Omnipaque 350 mg/mL or equivalent water soluble

contrast at 5 mL/sec + 40 mL of Saline chase at 5 mL/sec

#### **Wait 2 Minutes**

**2nd Perfusion Run**: 40 mL Omnipaque 350 mg/mL or equivalent water soluble contrast at 5 mL/sec + 40 mL of Saline chase at 5 mL/sec

#### Wait 2 Minutes

CTA Head and Neck - Follow posted online protocol

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*





Anatomical Reference: Orbitomeatal

Patient Position: Supine; Tilt the patient's head so that the orbitomeatal line is perpendicular to the

CT tabletop (see CT Head protocol).

Patient Orientation: Head First

Scan Range: C-1 through vertex

Scout: Lateral, AP if necessary

Scan Type: Helical Rotation Time: 0.9

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

MAs:

None

Head

120

Auto mA

Scan Delay: Smart Prep (ROI-Aorta) - Start scan 5 seconds after contrast arrival in aorta

Scan a 2nd series immediately after completion of the first series.

Respiration: Inspiration

**DFOV**: 25

Recon Algorithm: Standard

Contrast: 80 mL Omnipaque 300 mg/mL or equivalent water soluble contrast @ 2-3 mL/sec

Comments: 1. Send 5 mm Soft Tissue Algorithm Axial Images to PACS

2. Send 1 mm Soft Tissue Algorithm Axial Images to PACS

3. Send 3mm Sagittal and Coronal Soft Tissue Algorithm Reconstructions to

**PACS** 

4. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With Contrast: Venous Sinus Thrombosis, Headache

### CT Orbits



Diagnostic & Preventative Imaging Center

Anatomical Reference: Orbitomeatal

Patient Position: Supine; Tilt the patient's head so that the orbitomeatal line is perpendicular to the

CT tabletop (see CT Head protocol).

Patient Orientation: Head First

Scan Range: Maxilla through Orbits

Scout: Lateral Scan Type: Helical Rotation Time: 0.9

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

MAs:

None

Head

120

Auto mA

Scan Delay: 60 seconds, if contrast used

**Respiration:** None **DFOV:** 15cm

Recon Algorithm: See comments below

Contrast: 80 mL Omnipaque 300 mg/mL or equivalent water soluble contrast @ 2-3 mL/sec


Comments: 1. Send 1 mm Bone Algorithm Axial Images to PACS

2. Send 3 mm Soft Tissue Algorithm Axial Images to PACS

Send 2 mm Bone Algorithm Sagittal and Coronal Reconstructions to PACS
 Send 2 mm Soft Tissue Algorithm Sagittal and Coronal Reconstructions to

**PACS** 

5. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)



<sup>\*</sup> Image courtesy of University of Wisconsin

### **Indications**

Without Contrast: Injury, Fracture

With Contrast Infection/Cellulitis; If patient cannot have MRI: Tumor/Mass, Proptosis, Grave's

Disease

With and Without Contrast: Rarely indicated - Consult radiologist

\*\*Values will vary between machines. Use your own discretion when selecting these values.\*\*

Reviewed by Dr. Holdsworth 1/2025





Anatomical Reference: Orbitomeatal

Patient Position: Supine; Tilt the patient's head so that the orbitomeatal line is perpendicular to the

CT tabletop (see CT Head protocol).

Patient Orientation: Head First

Scan Range: Bottom of maxillary teeth through top of frontal sinuses

Scout: Lateral Scan Type: Helical Rotation Time: 0.9

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to patient size **Recon Algorithm:** See comments below

Contrast: None

Comments: 1. Send 1 mm Bone Algorithm Axial Images to PACS

2. Send 3 mm Soft Tissue Algorithm Axial Images to PACS

3. Send 2 mm Sagittal and Coronal Bone Algorithm Reconstructions to PACS

4. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

Without Contrast: Sinusitis, Headache

With Contrast: Rarely indicated - Consult radiologist

With and Without Contrast: Rarely indicated - Consult radiologist

### CT Facial Bones



Diagnostic & Preventative Imaging Center

Anatomical Reference: Orbitomeatal

**Patient Position:** Supine; Tilt the patient's head so that the orbitomeatal line is perpendicular to the

CT tabletop (see CT Head protocol).

Patient Orientation: Head First

Scan Range: Bottom of mandible through top of frontal sinuses

Scout: Lateral Scan Type: Helical Rotation Time: 0.9

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

MAS:

None

Head

120

Auto mA

Scan Delay: 60 seconds, if contrast used

Respiration: None

**DFOV:** Adjust to patient size **Recon Algorithm:** See comments below

Contrast: 80 mL Omnipaque 300 mg/mL or equivalent water soluble contrast @ 2-3 mL/sec

Comments: 1. Send 1 mm Bone Algorithm Axial Images to PACS

2. Send 3 mm Soft Tissue Algorithm Axial Images to PACS

3. Send 2 mm Sagittal and Coronal Bone Algorithm Reconstructions to PACS

4. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

Without Contrast: Facial Trauma

With Contrast: Infection

With and Without Contrast: Rarely indicated - Consult radiologist

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

### CT Temporal Bones



Diagnostic & Preventative Imaging Center

Lateral Canthus

Anatomical Reference: Orbitomeatal

Patient Position: Supine; Tilt the patient's head so that the orbitomeatal line is perpendicular to the

CT tabletop (see image).

Patient Orientation: Head First

Scan Range: Mastoid tip through top of petrous bone

Scout:LateralScan Type:HelicalRotation Time:0.9

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 625 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

MAs:

None

Head

120

Auto mA

Scan Delay: 60 seconds, if contrast used

**Respiration:** None **DFOV:** 15cm

Recon Algorithm: See comments below

Contrast: 80 mL Omnipaque 300 mg/mL or equivalent water soluble contrast @ 2-3 mL/sec

Comments: 1. Send 3 mm Full FOV Soft Tissue Axial Images to PACS

2. Send thin (.625 mm, if possible) Bone Algorithm Axial Reconstructions of the

Left Temporal Bone to PACS

3. Send thin (.625 mm, if possible) Bone Algorithm Axial Reconstructions of the Right Temporal Bone to PACS

A O Lui / 205


4. Send thin (.625 mm, if possible) Bone Algorithm Coronal Reconstructions of the Left Temporal Bone to PACS

5. Send thin (.625 mm, if possible) Bone Algorithm Coronal Reconstructions of the

Right Temporal Bone to PACS

6. Send thin Bone and Soft Tissue Algorithm slices to 3D workstation (Vitrea Bridge or TeraRecon)

\* Image courtesy of University of Wisconsin



Example axial field of view and coronal reconstruction plane

### **Indications**

Without Contrast: Hearing Loss, Cholesteatoma, Trauma/fracture, Mastoiditis

With Contrast: Jugular Bulb Anamolies; If patient cannot have MRI: Vestibulitis, Glomus

With and Without Contrast: Rarely indicated - Consult radiologist

Reviewed by Dr. Holdsworth 1/2025

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

### CT Soft Tissue Neck



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Neck
Patient Position: Supine
Patient Orientation: Head First
Scan Range: Orbits to Carina

Scout: AP and Lateral

Scan Type: Helical Rotation Time: 0.6

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None

\*\*If patient has significant dental hardware, scan additional angled axial views at 2

mm for 16 slices at two different angles - See Picture

SFOV: Large kVp: 120 mAs: Auto mA

Scan Delay: 60 seconds, if contrast used

Respiration: None

**DFOV:** Adjust to patient size (Include nose in field)

Recon Algorithm: Standard

Contrast: 80 mL Omnipaque 300 mg/mL or equivalent water soluble contrast @ 2-3 mL/sec

Comments: 1. Send 3 mm Soft Tissue Algorithm Axial Images to PACS


2. Send 2 mm Sagittal and Coronal Soft Tissue Algorithm Reconstructions to

**PACS** 

3. Send Angled Axial Images to PACS, if performed (Do not do recons on these

series)

4. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)



Example of angled axial images, if necessary

### **Indications**

With Contrast: Infection, Mass, Adenopathy

Without Contrast: ONLY perform if patient unable to have contrast

With and Without Contrast: Rarely indicated - Consult radiologist

\*\*Values will vary between machines. Use your own discretion when selecting these values.\*\*

Reviewed by Dr. Holdsworth 1/2025

<sup>\*</sup> Image courtesy of University of Wisconsin

# CT C-Spine



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Neck
Patient Position: Supine
Patient Orientation: Head First
Scan Range: C-1 thru C-7

Scout: Lateral, AP if necessary

Scan Type: Helical Rotation Time: 1.0

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

KVp:

MAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Cone down to only include spine

**Recon Algorithm:** See comments below

Contrast: 80 mL Omnipaque 300 mg/mL or equivalent water soluble contrast @ 2-3 mL/sec

Comments: 1. Send 2 mm Soft Tissue Axial Images to PACS

2. Send 2 mm Bone Algorithm Axial Images to PACS

3. Send 2 mm Bone Algorithm Sagittal and Coronal Reconstructions to PACS

4. Send 2 mm Soft Tissue Algorithm Sagittal Reconstructions to PACS

5. Send Soft Tissue Thin Slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

Without Contrast: Trauma, Fracture, Pain

With Contrast: Infection or Tumor AND patient unable to have MRI

With and Without Contrast: Rarely indicated - Consult radiologist

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

# CT T-Spine



Diagnostic & Preventative Imaging Center

Anatomical Reference: Sternal Notch
Patient Position: Supine
Patient Orientation: Feet First
Scan Range: T-1 thru T-12

Scout: Lateral, AP if necessary

Scan Type: Helical Rotation Time: 1.0

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Cone down to only include spine

Recon Algorithm: See comments below

Contrast: 80 mL Omnipaque 300 mg/mL or equivalent water soluble contrast @ 2-3 mL/sec

Comments: 1. Send 2 mm Soft Tissue Axial Images to PACS

2. Send 2 mm Bone Algorithm Axial Images to PACS

3. Send 2 mm Bone Algorithm Sagittal and Coronal Reconstructions to PACS

4. Send 2 mm Soft Tissue Algorithm Sagittal Reconstructions to PACS

5. Send Soft Tissue Thin Slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

Without Contrast: Trauma, Fracture, Pain

With Contrast: Infection or Tumor AND patient unable to have MRI

With and Without Contrast: Rarely indicated - Consult radiologist

# CT L-Spine



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid
Patient Position: Supine
Patient Orientation: Feet First
Scan Range: L-1 thru S-1

Scout: Lateral, AP if necessary

Scan Type: Helical Rotation Time: 1.0

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

None

Inspiration

**DFOV:** Cone down to only include spine

Recon Algorithm: See comments below

Contrast: 80 mL Omnipaque 300 mg/mL or equivalent water soluble contrast @ 2-3 mL/sec

Comments: 1. Send 2 mm Soft Tissue Axial Images to PACS

2. Send 2 mm Bone Algorithm Axial Images to PACS

3. Send 2 mm Bone Algorithm Sagittal and Coronal Reconstructions to PACS

4. Send 2 mm Soft Tissue Algorithm Sagittal Reconstructions to PACS

5. Send Soft Tissue Thin Slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

Without Contrast: Trauma, Fracture, Pain

With Contrast: Infection or Tumor AND patient unable to have MRI

With and Without Contrast: Rarely indicated - Consult radiologist

# CT Chest Routine



Diagnostic & Preventative Imaging Center

Anatomical Reference: Sternal Notch
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Apices through Adrenal Glands

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

**mAs:** Auto mA (minimum 50-maximum 420)

Scan Delay: Smart Prep (ROI-Aortic Arch, Scan Threshold-70 HU)

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 100 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 3 mL/sec

Comments: 1. Send 3 mm Soft Tissue Recon Algorithm Axial Images to PACS

2. Send 3 mm Lung Recon Algorithm Axial Images to PACS3. Send 2 mm Sagittal and Coronal Reconstructions to PACS

4. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

Without Contrast: Follow-Up Lung Nodule >8mm, Interstitial Lung Disease

With Contrast: Cough, Smoker, Pneumonia, Emphysema, Follow-Up or Staging Lung Cancer,

Follow-Up Long Nodule

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

## CT ION Chest



Diagnostic & Preventative Imaging Center

Performed at IMMC Only

Anatomical Reference: Sternal Notch

Patient Position: Supine; Arms above head

Patient Orientation: Feet First

Scan Range: 2 cm above lung apexes through entire chest 2 cm below costophrenic angles

Scout: AP, Lateral Scan Type: Helical

**Rotation Time:** Fastest to reduce motion

**Thickness:** .5mm-1.0MM, Recon as thin as possible **Interval:** .5mm-.8mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

None

Large

110-140

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: None

**Respiration:** Tech will give Full inspiration instructions

**DFOV:** < 32cm **Recon Algorithm:** Standard

Contrast: Oral: None

Comments: 1. Send .7 mm FOV <32 Soft Tissue Recon Axial Images to Vitrea, Tera, ION

2. Send 3 mm Full FOV Soft Tissue Recon Algorithm Axial Images to PACS

3. Send 3 mm Full FOV Lung Recon Algorithm Axial Images to PACS

4. Send 2 mm Full FOV Sagittal and Coronal Reconstructions to PACS

Scan Patients in the same position that will be used during the procedure





<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

# CT Chest High Resolution

Over 18



Diagnostic & Preventative Imaging Center

Anatomical Reference: Sternal Notch
Patient Position: Supine

Patient Orientation: Feet First

Scan Range: Apices through Lung Base

**Scout:** AP, Lateral if necessary

Scan Type:HelicalRotation Time:0.8Thickness:1 mmInterval:1 mmGantry Tilt:NoneSFOV:LargekVp:120

**mAs:** Auto mA (minimum 50-maximum 420)

Scan Delay: None

**Respiration:** Series 1 - Inspiration Series 2 - Expiration

**DFOV:** Adjust to patient size

Recon Algorithm: Series 1 - Bone Plus (Detailed Lung) Series 2 - Bone Plus (Detailed Lung)

Contrast: Oral: None

IV: None

Comments: History of Cystic Fibrosis - Call Radiologist

1. Series 1 - Inspiration - Scan using High Res technique

- Send 1 mm Inspiration High Res Lung Images to PACS.

- Send 3 mm Mediastinum Images to PACS.

- Send 3 mm Sagittal and Coronal Mediastinum Reconstructions to

PACS.

- Send Thin Inspiration Lung Windows to 3D Workstation (Vitrea

Bridge or TeraRecon)

2. Series 2 - Expiration - Scan using High Res technique

- Send 1 mm Expiration High Res Lung Images to PACS.

### **Indications**

Without Contrast: Interstitial Lung Disease, Bronchiectasis

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

### CT Chest High Resolution



Diagnostic & Preventative Imaging Center

Under 18

Anatomical Reference: Sternal Notch

Patient Position: Prone
Patient Orientation: Feet First

Scan Range: Apices through Lung Base

**Scout:** AP, Lateral if necessary

Scan Type: Axial Rotation Time: 0.8

**Thickness:** 2 mm, Recon as thin as possible **Interval:** 2 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: None Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: None

Comments: 1. Send 2 mm Axial Mediastinum Windows to PACS

2. Send 2 mm Axial High Res Lung Windows to PACS

3. Send 3 mm Sagittal and Coronal Mediastinum Reconstructions to PACS

3. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

Without Contrast: Interstitial Lung Disease, Bronchiectasis

# CT Low Dose Lung Screen



Anatomical Reference: Sternal Notch
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Apices through Lung Base

**Scout:** AP, Lateral if necessary

Scan Type:HelicalRotation Time:0.5Thickness:1 mmInterval:1 mmGantry Tilt:NoneSFOV:LargekVp:120

mAs: Use low dose technique

Scan Delay: None Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: None

Comments: 1. Send 3 mm Mediastinum Algorithm Axial Images to PACS

2. Send 1.5 mm Mediastinum Algorithm Axial Images to PACS

3. Send 3 mm Mediastinum Algorithm Sagittal and Coronal Reconstructions to

**PACS** 

4. Send thin (recon to 0.6mm) Abdomen Algorithm Images to 3D Workstation

(Vitrea Bridge or TeraRecon)

Notes: \*\*\*\* For patients having a lung screening exam which requires a 3 or 6 month

follow up, the follow up exam should be scanned using the low dose

protocol but billed as a CT Chest WO Contrast

\*\*\*\* On follow up exams, be sure to add a comment for the radiologist that the scan is a lung screen F/U so the correct dictation template is used.

### **Indications**

Without Contrast: Screening

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

### CT Cardiac Calcium Score



Diagnostic & Preventative Imaging Center

Anatomical Reference: Top of the Shoulders

Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Carina thru Base of Heart

Scout: AΡ Axial Scan Type: Thickness: 3 mm Interval: 3 mm **Gantry Tilt:** None SFOV: Large kVp: 120 mAs: Auto mA Scan Delay: None Respiration: Inspiration DFOV: Varied **Recon Algorithm:** Mediastinum

Contrast: None

Comments: 1. Send Mediastinum windows to PACS

2. Send Mediastinum windows to 3D Workstation (TeraRecon)

3. Use 4 lead EKG for gating

4. Female patients must remove bra prior to scan

### **Indications**

Without Contrast: Screening, Elevated Cholesterol, Hypertension, Diabetes, Atypical Chest Pain

# Esophagram



Diagnostic & Preventative Imaging Center

Esophageal Leak Study

Anatomical Reference: Mid neck **Patient Position:** Supine **Patient Orientation:** Feet First

Mid Neck (C4/C5) thru Stomach (L2/L3) Scan Range:

Scout: AP, Lateral if necessary

Scan Type: Helical **Rotation Time:** 8.0

Thickness: 3 mm, Recon as thin as possible Interval: 3 mm, Recon as thin as possible

**Gantry Tilt:** None SFOV: Large kVp: 120

Auto mA (minimum 50-maximum 420) mAs:

Scan Delay: None Respiration: Inspiration

DFOV: Adjust to patient size

**Recon Algorithm:** Standard

Contrast: Oral: Dilute 15 mL Omnipaque 350 or equivalent water soluble contrast in

10 oz bottled or sterile water

IV: None

Comments: \*\*Patient needs to be NPO for 2 hours prior to exam

1. Series 1 - Non contrast

2. Series 2 - Oral Contrast - The patient drinks all of the contrast - Scan

immediately

The shorter the time period between swallowing oral contrast and imaging acquisition, the better the diagnostic value

1. Send 3 mm Axial Images to PACS

2. Send 2 mm Sagittal and Coronal Reconstructions to PACS

3. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

Without Contrast: Esophageal leak or rupture

### CT Trauma Chest/Abd/Pelvis



Diagnostic & Preventative Imaging Center

Anatomical Reference: Sternal Notch
Patient Position: Supine
Patient Orientation: Head First

Scan Range: Apices through Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

**mAs:** Auto mA (minimum 50-maximum 420)

**Scan Delay:** 60 seconds **Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 100 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 2-3 mL/sec, preferred

Comments: 1. Send 3 mm Soft Tissue Recon Algorithm Axial Images to PACS

2. Send 3 mm Lung Recon Algorithm Axial Images of entire Chest/Adomen/Pelvis

to PACS

3. Send 2 mm Soft Tissue Sagittal and Coronal Reconstructions to PACS

4. Send 2 mm Bone Recon Algorithm Sagittal and Coronal Reconstructions to

**PACS** 

5. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

6. Delay images only if requested by ER or trauma physician

### **Indications**

With Contrast: Trauma

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

# CT Abd/Pelvis



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm through Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

mAs: Auto mA (minimum 50-maximum 420)

**Scan Delay:** 60 seconds **Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: See notes below - 1000 mL diluted contrast 1-2 hours prior to CT OR

1000 mL Ready to Drink (RTD) Omnipague Oral Solution 1-2 hours

prior to CT

IV: 100 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 2-3 mL/sec

Comments: 1. Send 3 mm Axial Images to PACS

2. Send 2 mm Sagittal and Coronal Reconstructions to PACS

3. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

4. Renal/bladder delays not routinely indicated. Contact the on site radiologist if

there are questions.

Oral Contrast Notes: 1. Oral contrast not needed for most exams.

2. Give oral contrast for:

a. Any history of pelvic radiation therapy

b. PARTIAL small bowel obstruction (SBO) (Complete or High grade

SBO = NO oral contrast)

c. Recently treated SBO (e.g. floor patients with an NG tube for a

couple days)

Continued on next page

### CT Abdomen/Pelvis - Routine Cont.

Oral Contrast Notes: d. Abscess

e. Symptoms of appendicitis for 3 days or more

f. Low BMI (24 or less)

g. Postop from any major bowel surgery including bariatrics

h. Any Gynecologic Oncology patient of Dr. Elg i. Ovarian pathology / Tubo-ovarian abscess

j. GI fistula

k. Duplication cyst

I. Follow up of gastrointestinal or mesenteric trauma

m. Crohn's Disease

### **Indications**

With Contrast: Routine, Follow-Up, Pain

\*\*Values will vary between machines. Use your own discretion when selecting these values.\*\*

Reviewed by Dr. King 1/2025





Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm through Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

**mAs:** Auto mA (minimum 50-maximum 420)

**Scan Delay:** 60 seconds **Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: See notes below - 1000 mL diluted contrast 1-2 hours prior to CT OR

1000 mL Ready to Drink (RTD) Omnipaque Oral Solution 1-2 hours

prior to CT

IV: 100 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 2-3 mL/sec

Comments: 1. Send 3 mm Axial Images to PACS

2. Send 2 mm Sagittal and Coronal Reconstructions to PACS

3. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

4. Renal/bladder delays not routinely indicated. Contact the on site radiologist if

there are questions.

Oral Contrast Notes: 1. Give oral contrast for:

a. Recent (≤ 2 months) hx of abdominal surgery

b. BMI ≤ 24

c. Any hx of inflammatory bowel disease (i.e. Crohns or ulcerative

d. Any hx major complex oncologic abdominal/pelvic cancer surgery

(i.e. gyn-onc, whipple surgery)

e. Lower abdominal pain > 3 days with concern for abscess

f. Concern for partial SBO (no vomiting, passing gas/stool) should

receive oral contrast, regardless of BMI

Continued on next page

### CT Abdomen/Pelvis - ER Cont.

**Oral Contrast Notes:** 

g. Crohns disease

#### 2. Plain film first:

- a. Surgical abdomen (i.e. rigid abd, generalized peritonitis, concern for perforation/free air)
- b. High grade SBO (i.e. vomiting, obstipation, distension, hx prior abdominal surgeries)

Pending results of Plain Film - Should consult surgery prior to ordering CT

#### 3. Without oral contrast for:

- a. Uncomplicated appendicitis, uncomplicated diverticulitis, uncomplicated colitis or pyelonephritis
- b. Recent trauma

### **Indications**

With Contrast: Adult patients presenting to ER with belly pain

\*\*Values will vary between machines. Use your own discretion when selecting these values.\*\*

Reviewed by Dr. King 1/2025





Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm through Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

**mAs:** Auto mA (minimum 50-maximum 420)

**Scan Delay:** 60 seconds **Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None, except for symptoms of appendicitis for 3 days or more, then

give: 1000 mL diluted contrast 1-2 hours prior to CT OR

1000 mL Ready to Drink (RTD) Omnipaque Oral Solution 1-2 hours

prior to CT

IV: 100 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 2-3 mL/sec

Comments: 1. No preliminary non-contrast scans necessary.

2. For all abdominal CT's in which there is a concern for or history of recent

abscess, oral contrast is needed in addition to IV.

3. Send 3 mm Axial Images to PACS

4. Send 2 mm Sagittal and Coronal Reconstructions to PACS

5. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

6. Renal/bladder delays not routinely indicated. Contact the on site radiologist if

there are questions.

7. For symptoms of appendicitis for 3 days or more, give oral and IV contrast.

### **Indications**

With Contrast: RLQ Pain, Elevated WBC, Fever

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*





Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm through Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

None

Large

120

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: None Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: None

Comments: 1. Send 3 mm Axial Images to PACS


2. Send 2 mm Sagittal and Coronal Reconstructions to PACS3. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

4. Check scans with radiologist before patient leaves to make sure contrast is not

necessary.

### **Indications**

Without Contrast: Hematuria, Flank Pain, Hx Stones





\* Call radiologist before the exam. Radiologist needs to review 100 second acquisition immediately.\*

Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm through Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

None

Large

120

mAs: Auto mA (minimum 50-maximum 420)
Scan Delay: Series 1 - 100 seconds; Series 2 - 4 minutes

Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 150 mL Omnipaque 300 mg/mL or Omnipaque 350 mg/mL or

equivalent water soluble contrast injected at 2-3 mL/sec

**Comments:** 1. Series 1 - 100 second delay - Dome of diaphragm through pubic symphysis

\*Radiologist Review\*

2. Series 2 - 4 minute delays - Dome of diaphragm through pubic symphysis

3. Send 3 mm Axial Images to PACS

4. Send 2 mm Sagittal and Coronal Reconstructions to PACS

5. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

#### **Indications**

With Contrast: Blood Clot in IVC

# CT Liver 3 Phase



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Iliac Crest

Scout: AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

**mAs:** Auto mA (minimum 50-maximum 420)

Scan Delay: Arterial - 30-35 sec; Venous - 70sec; Series 4 - 5 min

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 120 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 5 mL/sec

Comments: 1. Series 1 - Without Contrast

2. Series 2 - Arterial Phase3. Series 3 - Venous Phase4. Series 4 - 5 minute delay

5. OPTIONAL - If performed specifically for hemangioma, add:

Series 5 - 10 minute delay

6. Send 3 mm Axial Images (Series 1-4) to PACS

7. Send 2 mm Sagittal and Coronal Reconstructions (Series 1-3) to PACS

8. Send thin images (Series 1-4) to 3D Workstation (Vitrea Bridge or TeraRecon)

9. A proper hepatic arterial phase is when the hepatic artery is heavily contrasted

and there is a trace amount of contrast in the portal vein.

#### **Indications**

With & Without Contrast: Known Liver Mass/Hepatoma, S/P Radiofrequency Ablation (RFA), S/P

Chemotherapy, Embolization, Follow Up Known Hemangioma

### CT Adrenals



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Iliac Crest

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

mAs: Auto mA (minimum 50-maximum 420)
Scan Delay: Venous - 70 sec, Series 3 - 15 minutes

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 130 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 2 mL/sec

Comments: 1. Series 1 - Without Contrast to Localize Adrenals

2. Series 2 - Venous Phase - Dome of Diaphragm to the Iliac Crest, or through

Pubic Symphysis if Pelvis ordered

3. Series 3 - 15 minute delay from Dome of Diaphragm to Iliac Crest, or from

Dome of Diaphragm through Pubic Symphysis if Pelvis ordered

4. Send 3 mm Axial Images (Series 1-3) to PACS

5. Send 2 mm Sagittal and Coronal Reconstructions (Series 1 & 2) to PACS

6. Send thin images (Series 1-3) to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With & Without Contrast: Adrenal Mass, Adrenal F/U, Pheochromocytoma

### CT Pancreas



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Iliac Crest

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

**mAs:** Auto mA (minimum 50-maximum 420)

Scan Delay: Arterial - 25 sec, Venous - 60 sec, Series 4 - 5 min

Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: 200 mL diluted contrast 15-20 minutes prior to scanning

**EXCEPTION**, if recent post-op, following recent abscess in upper abdomen or gastric outlet, or duodenal obstruction, then give: 1000 mL diluted contrast 1-2 hour prior to CT plus 200 mL diluted

contrast just prior to scanning OR

1000 mL Ready to Drink (RTD) Omnipaque Oral Solution 1-2 hours

prior to CT plus 200 mL diluted contrast just prior to scanning

IV: 120 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 4 mL/sec

Comments: All series must include entire pancreas - Scan from top of liver through

bottom of duodenal sweep or iliac crest, whichever is lower.

Series 1 - Without Contrast
 Series 2 - Arterial Phase

3. Series 3 - Venous Phase4. Series 4 - 5 minute delay

5. Send 3 mm Axial Images (Series 1-4) to PACS

6. Send 2 mm Sagittal and Coronal Reconstructions (Series 1-3) to PACS

7. Send thin images (Series 1-4) to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With & Without Contrast: Epigastric Pain, Elevated Amylase-Lipase, Pancreatitis

# CT Kidneys /Mass



Diagnostic & Preventative Imaging Center

\*If ordered as CT Abdomen & Pelvis, see notes below.\*

Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Iliac Crest

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: Arterial - 30 sec; Nephrographic - 60 sec; Delays - 10 mins

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 120 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 4 mL/sec

Comments: 1. Series 1 - Without Contrast - Dome of Diaphragm to the Iliac Crest

2. Series 2 - Arterial Phase - Dome of Diaphragm to the Iliac Crest

3. Series 3 - Nephrographic Phase - Dome of Diaphragm to the Iliac Crest

4. Series 4 - Delayed Kidneys - 10 minute delay

5. Send 3 mm Axial Images (Series 1-4) to PACS

6. Send 2 mm Sagittal and Coronal Reconstructions (Series 1-4) to PACS

7. Send thin images (Series 1-4) to 3D Workstation (Vitrea Bridge or TeraRecon)

8. When a pelvis is also ordered, contact Dr. King or the on site radiologist for pelvis protocol. Default protocol for pelvis portion is a single run in the portal

venous phase.

#### **Indications**

With & Without Contrast: F/U suspected mass on other study, Known renal mass, Follow up abscess, renal

trauma, infarction, laceration, etc.

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*





\*If a CT Urogram is ordered for renal mass, scan according to CT Kidney/Mass protocol (to include pelvis)

Anatomical Reference: Xyphoid
Patient Position: Supine
Patient Orientation: Feet First

**Scan Range:** Dome of Diaphragm to Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical

**Thickness:** Variable, See comments **Interval:** Variable, See comments

Gantry Tilt: None SFOV: Large kVp: 120

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: See below Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: 200 mL water on the CT scan table

IV: 50 mL Omnipaque 300 or Isovue 300 mg/mL + 50 mL Saline flush

injected at 1.5 mL/sec Wait 10 minutes

100 mL Omnipaque 300 or Isovue 300 mg/mL + 50 mL Saline flush

injected at 3 mL/sec

Comments: 1. Patient needs to arrive with a full bladder. Have patient void before bringing to

the scanner.

2. Series 1 - Without Contrast - Dome of Diaphragm to Pubic Symphysis (3 mm slice thickness) - Call radiologist if obvious ureter stone, otherwise proceed.

3. INJECTION - Inject 50 mL of contrast + 50 mL saline flush @ 1.5 mL/second.

Wait 10 minutes, THEN proceed with the next series.

4. Series 2 - Parenchymal Phase - Inject 100 mL of contrast + 50 mL saline flush @ 3mL/second. Start the scan after a 115 second delay (build delay into scan

protocol if possible) - Scan Dome of Diaphragm to Pubic Symphysis

5. Send Axial Images (Series 1 & 2) to PACS

6. Send 2 mm Sagittal & Coronal Reconstructions (Series 1 & 2) to PACS7. Send thin slices (all series) to 3D Workstation (Vitrea Bridge or TeraRecon)

8. Correct Order: CT Abdomen / Pelvis W/WO contrast

### **Indications**

With & Without Contrast: Nonspecific hematuria, Ureter injury

# CT Cystogram



Anatomical Reference: Iliac Crest
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Iliac Crest through Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

**mAs:** Auto mA (minimum 50-maximum 420)

Scan Delay: None Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

Retrograde: 30 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

diluted in 500 mL Saline

Using sterile technique, connect the tubing to the patient's foley catheter and allow the bladder to fill retrograde. If the patient is oriented, have the pt tell you when they are getting uncomfortably full.

If they are unable to tell you, run in about 300 mL.

Comments: 1. Series 1 - Without Contrast

2. Series 2 - Fill bladder and scan

3. Have onsite radiologist check images to determine if additional imaging is

needed.

4. Send 3 mm Axial Images (All series) to PACS

5. Send 2 mm Sagittal and Coronal Reconstructions (All series) to PACS

\*If performed for trauma, include soft tissue and bone algorithm recons on without series. If not for trauma, only include soft tissue algorithm recons on w/o series.

6. Send thin images (All series) to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With & Without Contrast: Trauma, Post Op Bladder/Pelvic Surgery

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

# CT Enterography



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

100

MAs:

Auto mA

Scan Delay:

None

Large

Auto mA

50-60 see

Scan Delay: 50-60 seconds Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: 1350 mL's Volumen @ 60, 45 and 30 minutes prior to scanning.

Full cup of water immediately before scanning.

IV: 120 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 4 mL/sec

Comments: 1. Radiologist to approve prior to scheduling (if ordered for Crohn's, then proceed)

2. KUB before patient starts drinking contrast.

3. Send 3 mm Axial Images to PACS

4. Send 2 mm Sagittal and Coronal Reconstructions to PACS

5. Send thin images to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With Contrast: Crohn's Disease, Stricture, Nonspecific GI Indications, specifically small bowel

GI BLEED WORKUP - USE MULTIPHASE PROTOCOL

# CT Enterography



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

MAS:

None

Large

Auto mA

Scan Delay: Arterial: As Below; Enteric: 20-25 secs; Delayed: 70-75 secs

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: 1350 mL's Volumen @ 60, 45 and 30 minutes prior to scanning.

Full cup of water immediately before scanning.

IV: 120 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 4 mL/sec

Comments: 1. Radiologist to approve prior to scheduling (if ordered for occult GI bleeding, then

proceed).

2. KUB before patient starts drinking contrast.

3. Series 1 - Without contrast

4. Series 2 - Arterial Phase - Bolus triggered w/ ROI on mid aorta. Trigger with 200

HU.

5. Series 3 - Enteric Phase - 20-25 secs after arterial phase6. Series 4 - Delayed Phase - 70-75 secs after arterial phase

7. Send 3 mm Axial Images (Series 1-4) to PACS

8. Send 2 mm Sagittal and Coronal Reconstructions (Series 1-4) to PACS

9. Send thin images (Series 1-4) to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With Contrast: Occult GI Bleed Workup; For Acute GI Bleed-See CTA Protocol

## CT Virtual Colonoscopy



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid

Patient Position: Series 1 - Supine Series 2 - Prone

Patient Orientation: Feet First

**Scan Range:** Dome of Diaphragm to Pubic Symphysis

**Scout:** AP, PA, Lateral if necessary

Helical Scan Type: **Rotation Time:** 8.0 Thickness: 1.25 mm 1.25 mm Interval: **Gantry Tilt:** None SFOV: Large kVp: 120 mAs: Auto mA Scan Delay: None Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Distend colon with CO2

**Comments:** 1. Make sure patient has done the bowel prep.

2. Do not scan until colon is completely distended.

3. Series 1 - Patient is supine.4. Series 2 - Patient is prone.

5. Send 2 mm Sagittal and Coronal Reconstructions (on Series 1 OR 2, whichever

has best coverage) to PACS

6. Send all images to McKesson PACS.

### **Indications**

Without Contrast: Screening

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

## CT Pelvis



Diagnostic & Preventative Imaging Center

Anatomical Reference: Iliac Crest
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Iliac Crest thru Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

Large

Auto mA

75 seconds

Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: See notes below - 1000 mL diluted contrast 1-2 hours prior to CT OR

1000 mL Ready to Drink (RTD) Omnipaque Oral Solution 1-2 hours

prior to CT

IV: 100 mL Omnipaque 300 mg/mL or equivalent water soluble contrast

injected at 2-3 mL/sec

**Comments:** 1. No preliminary non-contrast scans necessary.

2. Send 3 mm Axial Images to PACS

3. Send 2 mm Sagittal and Coronal Reconstructions to PACS

4. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

5. Do delayed bladder on all patients with:

a. Pelvic abcess

b. Recent pelvic surgery

c. Any bladder, prostate or gynecologic surgery

Oral Contrast Notes: 1. Oral contrast not needed for most exams.

2. Give oral contrast for:

a. Any history of pelvic radiation therapy

b. PARTIAL small bowel obstruction (SBO) (Complete or High grade

SBO = NO oral contrast)

Continued on next page

### CT Pelvis Cont.

Oral Contrast Notes: c. Recently treated SBO (e.g. floor patients with an NG tube for a

couple days)

d. Abscess

e. Symptoms of appendicitis for 3 days or more

f. Low BMI (24 or less)

g. Postop from any major bowel surgery including bariatrics

h. Any Gynecologic Oncology patient of Dr. Elg i. Ovarian pathology / Tubo-ovarian abscess

j. GI fistula

k. Duplication cyst

I. Follow up of gastrointestinal or mesenteric trauma

### **Indications**

Without Contrast: Pain, Follow Up

## CT Bony Pelvis



Diagnostic & Preventative Imaging Center

Anatomical Reference: Iliac Crest
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: 2 cm above Iliac Crest thru Proximal Femoral Shaft

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: None

**Comments:** 1. Send 3 mm Soft Tissue Axial Images to PACS

2. Send 3 mm Bone Window Axial Images to PACS


3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal


Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)







<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

Without Contrast: Pain, Injury, Fracture

## CT Hip/Pelvis



Diagnostic & Preventative Imaging Center

Anatomical Reference: Iliac Crest
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: 2 cm above Iliac Crest thru Proximal Femoral Shaft

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images (Pelvis & Affected Hip) to PACS

2. Send 3 mm Bone Window Axial Images (Pelvis & Affected Hip) to PACS


3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal


Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)







<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

Without Contrast: Pain, Injury, Fracture

## CT Hip Arthrogram



Diagnostic & Preventative Imaging Center

Anatomical Reference: Iliac Crest
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Mid Iliac Crest to Subtrochanter (See picture)

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:NoneSFOV:SmallkVp:120mAs:Auto mAScan Delay:NoneRespiration:None

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images (Pelvis & Affected Hip) to PACS


2. Send 3 mm Bone Window Axial Images (Pelvis & Affected Hip) to PACS


3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal


Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)







<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

With Contrast: Pain, Injury, Fracture

### CT SI Joint / Tailbone



Diagnostic & Preventative Imaging Center

Anatomical Reference: Iliac Crest
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: 2 cm above Iliac Crest thru Coccyx

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images to PACS

2. Send 3 mm Bone Window Axial Images to PACS

3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)

\*\*SI Joint Reconstruction:

Align reconstruction plan from the mid sagittal sacral image Oblique Coronal: Angle plane parallel to the long axis of the sacrum Oblique Axial: Angle plane parallel to the OBLIQUE CORONAL reconstucted image

### **Indications**

Without Contrast: Pain, Injury, Fracture

<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)





Anatomical Reference: Mid Knee

Patient Position: Supine; TAPE FEET TOGETHER

Patient Orientation: Feet First Scan Range: Entire Knee

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to Knee **Recon Algorithm:** Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images to PACS


2. Send 3 mm Bone Window Axial Images to PACS

3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)



<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

Without Contrast: Pain, Injury, Fracture

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*

## CT Knee Arthrogram



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Knee

Patient Position: Supine; TAPE FEET TOGETHER

Patient Orientation: Feet First

Scan Range: Entire Knee (See picture)

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to Knee

Recon Algorithm: Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images to PACS


2. Send 3 mm Bone Window Axial Images to PACS


3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)





<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

With Contrast: Pain, Injury, Fracture

\*\*Values will vary between machines. Use your own discretion when selecting these values.\*\*

Reviewed by Dr. Choi 6/2021





Anatomical Reference: Mid Ankle

Patient Position: Supine; TAPE FEET TOGETHER (if possible)

Patient Orientation: Feet First Scan Range: Entire Ankle

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

KVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to Ankle

Recon Algorithm: Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Window Axial Images to PACS

2. Send 3 mm Bone Window Axial Images to PACS


3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal


Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)







<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

Without Contrast: Pain, Injury, Fracture

### CT Subtalar



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Foot

Patient Position: Supine; TAPE FEET TOGETHER (if possible)

Patient Orientation: Feet First Scan Range: Entire Ankle

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

KVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to Foot **Recon Algorithm:** Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images to PACS

2. Send 3 mm Bone Window Axial Images to PACS

3. Send 2 mm thick / 2 mm interval Bone Window Sagittal and Coronal


Reconstructions to PACS


4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)









<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

Without Contrast: Pain, Injury, Fracture





Anatomical Reference: Mid Ankle

Patient Position: Supine; TAPE FEET TOGETHER (if possible)

Patient Orientation: Feet First Scan Range: Entire Ankle

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

KVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to Foot **Recon Algorithm:** Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images to PACS

2. Send 3 mm Bone Window Axial Images to PACS

3. Send 2 mm thick / 2 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)



\*\*Align reconstruction on the 1st TMT Joint.

### **Indications**

Without Contrast: Pain, Injury, Fracture

<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)





Anatomical Reference: Mid Foot

Patient Position: Supine; TAPE FEET TOGETHER (if possible)

Patient Orientation: Feet First Scan Range: Entire Foot

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

KVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to Foot **Recon Algorithm:** Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images to PACS

2. Send 3 mm Bone Window Axial Images to PACS

3. Send 2 mm thick / 2 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)



\*\*Align reconstruction on the metatarsals (MTP joint).

### **Indications**

Without Contrast: Pain, Injury, Fracture

<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

## CT Shoulder



Diagnostic & Preventative Imaging Center

Start above

**AC** Joint

**FOV** to include

Scapula & Humerus

Scan thru

Scapula

Anatomical Reference: Sternal Notch

Patient Position: Supine (off center pt so affected shoulder is completely on table)

Patient Orientation: Head First
Scan Range: Entire Shoulder

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

kVp:

mAs:

Scan Delay:

Respiration:

None

None

Inspiration

**DFOV:** Adjust to Shoulder

Recon Algorithm: Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images to PACS


2. Send 3 mm Bone Window Axial Images to PACS

3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)









### **Indications**

Without Contrast: Pain, Injury, F/U Fracture, Post Arthrogram

<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

## CT Shoulder Arthrogram



Diagnostic & Preventative Imaging Center

Anatomical Reference: Sternal Notch

Patient Position: Supine (off center pt so affected shoulder is completely on table)

Patient Orientation: Head First

Scan Range: Proximal 1/3 Shaft (See picture)

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:NoneSFOV:LargekVp:120mAs:Auto mAScan Delay:NoneRespiration:Inspiration

**DFOV:** Adjust to Shoulder

Recon Algorithm: Standard

Contrast: None

Comments: 1. Series 1 - Scan as a Routine CT Shoulder

2. Series 2 - ABER (Abduction External Rotation), if tolerated - Affected arm up with the patient's hand resting behind their head - Scan through shoulder

3. Send 3 mm Soft Tissue Axial Images (Series 1-2) to PACS4. Send 3 mm Bone Window Axial Images (Series 1-2) to PACS

5. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal

Reconstructions (Series 1-2) to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices (Series 1-2) to 3D Workstation (Vitrea

Bridge or TeraRecon)

### **Indications**

With Contrast: Pain, Injury, F/U Fracture, Post Arthrogram

## CT Elbow



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Elbow

Patient Position: Supine; Arm extended above head

Patient Orientation: Head First Scan Range: Entire Elbow

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

Small

kVp:

120

mAs:

Auto mA

Scan Delay:

Respiration:

None

**DFOV:** Adjust to Elbow

Recon Algorithm: Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images to PACS


2. Send 3 mm Bone Window Axial Images to PACS

3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)



<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

Without Contrast: Pain, Injury, Fracture

## CT Elbow Arthrogram



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Elbow

Patient Position: Supine; Arm extended above head

Patient Orientation: Head First Scan Range: Entire Elbow

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:NoneSFOV:SmallkVp:120mAs:Auto mAScan Delay:NoneRespiration:None

**DFOV:** Adjust to Elbow

Recon Algorithm: Standard

Contrast: None

Comments: 1. Send 3 mm Soft Tissue Axial Images to PACS


2. Send 3 mm Bone Window Axial Images to PACS

3. Send 3 mm thick / 3 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)



<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

With Contrast: Pain, Injury, Fracture

### CT Wrist



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Wrist

Patient Position: Prone, Arm Extended

Patient Orientation: Head First Scan Range: Entire Wrist

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

KVp:

mAs:

Scan Delay:

Respiration:

None

None

**DFOV:** Adjust to Wrist **Recon Algorithm:** Standard

Contrast: None

Comments: 1. Send 2 mm Soft Tissue Axial Images to PACS


2. Send 2 mm Bone Window Axial Images to PACS

3. Send 2 mm thick / 2 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)



<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

Without Contrast: Pain, Injury, Fracture

## CT Scaphoid / Wrist



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Wrist

Patient Position: Prone, Arm Extended

Patient Orientation:Head FirstScan Range:Entire Wrist

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

Small

kVp:

120

mAs:

Auto mA

Scan Delay:

Respiration:

None

**DFOV:** Adjust to Wrist **Recon Algorithm:** Standard

Contrast: None

Comments: 1. Send 2 mm Soft Tissue Axial Images to PACS


2. Send 2 mm Bone Window Axial Images to PACS

3. Send 2 mm thick / 2 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or TeraRecon)





<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### **Indications**

Without Contrast: Pain, Injury, Fracture

### CT Hand



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Hand

Patient Position: Prone, Arm Extended

Patient Orientation: Head First Scan Range: Entire Hand

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.8

**Thickness:** 1.25 mm, Recon as thin as possible **Interval:** 0.6 mm, Recon as thin as possible

Gantry Tilt:

SFOV:

Small

kVp:

120

mAs:

Auto mA

Scan Delay:

Respiration:

None

**DFOV:** Adjust to Hand **Recon Algorithm:** Standard

Contrast: None

Comments: 1. Send 2 mm Soft Tissue Axial Images to PACS

2. Send 2 mm Bone Window Axial Images to PACS

3. Send 2 mm thick / 2 mm interval Bone Window Sagittal and Coronal

Reconstructions to PACS

4. Send Soft Tissue Thin (0.6 mm) Slices to 3D Workstation (Vitrea Bridge or

TeraRecon)

### **Indications**

Without Contrast: Pain, Injury, Fracture

<sup>\*</sup> Images used with permission: Ken L. Schreibman PhD/MD (schreibman@alum.mit.edu)

### CTA Head & Neck



Diagnostic & Preventative Imaging Center

Anatomical Reference: Mid Neck Patient Position: Supine **Patient Orientation: Head First** 

Scan Range: Inferior Border of Aortic Arch through Vertex

Scout: AP and Lateral

Scan Type: Helical **Rotation Time:** 0.6

Thickness: 3 mm, Recon as thin as possible 3 mm, Recon as thin as possible Interval:

**Gantry Tilt:** None SFOV: Large kVp: 120 mAs: Auto mA

Scan Delay: Smart Prep (ROI-Aorta)

Respiration: None

DFOV: Adjust to include all of head

**Recon Algorithm:** Standard

Contrast: 80 mL Omnipaque 350 mg/mL or equivalent water soluble contrast @ 3 mL/sec


Comments: 1. Send thin Soft Tissue Axial images to PACS

2. Send Axial, Coronal and Sagittal Head MIP reformats to PACS, see below

3. Send Coronal and Sagittal Neck MIP reformats to PACS, see below 4. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

5. If only a CTA Head or CTA Neck is ordered, contact a radiologist

MIPs: Head - 10 mm slice thickness; 2.5 mm slice overlap; Window width = 600; Window level = 200 Neck - 10 mm slice thickenss; 5 mm slice overlap; Window width = 600; Window level = 200





### **Indications**

With Contrast: Headache, Vertigo, Syncope, Aneurysm, Stenosis, Stroke, TIA

Spinous process

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*





Anatomical Reference: Sternal Notch

Patient Position: Supine Patient Orientation: Feet First

Scan Range: Apices through Lung Bases

**Scout:** AP, Lateral if necessary

Scan Type: Helical Rotation Time: 0.5

**Thickness:** 3mm, Recon as thin as possible **Interval:** 3mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large

**kVp:** 100, adjust accordingly for very large patients

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: Smart Prep (ROI-Pulmonary Artery, Scan Threshold-120 HU)

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 100 mL Omnipaque 350 or Isovue 370 mg/mL injected at 5 mL/sec

\*May need to adjust scan parameters to use all 100 mL IV contrast

Comments: 1. Start IV in right arm, if possible.

2. Send 3 mm Axial Images to PACS

3. Send 2 mm Sagittal and Coronal Reconstructions to PACS

4. Send 3D Reformatted Images to PACS

5. Send thin slices to PACS

6. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With Contrast: Evaluate for Pulmonary Emboli, Positive DVT, Acute SOB

### CTA Chest



Diagnostic & Preventative Imaging Center

Anatomical Reference: Sternal Notch

Patient Position:SupinePatient Orientation:Feet First

Scan Range: Aortic Arch through Diaphragm

Scout: AP, Lateral if necessary

Scan Type: Helical

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt:NoneSFOV:LargekVp:120

**mAs:** Auto mA (minimum 50-maximum 420)

Scan Delay: Smart Prep (ROI-descending aorta just below aortic arch)

Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 130 mL Omnipaque 350 or Isovue 370 mg/mL injected at 5 mL/sec

Comments: 1. Series 1 - Without Contrast

2. Series 2 - Arterial Phase

3. If scanning for aneurysm or dissection, need to scan through iliacs.

4. Send 3 mm Axial Images to PACS

5. Send 2 mm Sagittal and Coronal Reconstructions to PACS

6. Send 3D Reformatted Images to PACS

7. Send thin slices to PACS

8. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

Use gating for CTA Chest, if available.

### **Indications**

With & Without Contrast: Aneurysm, Dissection, Coarctation

## CTA Chest/Abd/Pelvis



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Thoracic Inlet to Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: Smart Prep (ROI-Descending Aorta just below Arch)

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 130 mL Omnipaque 350 or Isovue 370 mg/mL injected at 5 mL/sec

Comments: 1. Series 1 - Without Contrast

2. Series 2 - Arterial Phase

3. Send 3 mm Axial Images to PACS

4. Send 2 mm Sagittal and Coronal Reconstructions to PACS

5. Send 3D Reformatted Images to PACS

6. Send thin slices to PACS

7. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

Use gating for CTA Chest, if available and scanner allows for diagnostic

imaging.

### **Indications**

With & Without Contrast: Dissection

## CTA Abd/Pelvis



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

**mAs:** Auto mA (minimum 50-maximum 420)

Scan Delay: Smart Prep (ROI-Descending Aorta just below Arch)

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 130 mL Omnipaque 350 or Isovue 370 mg/mL injected at 5 mL/sec

Comments: 1. Series 1 - Without Contrast

2. Series 2 - Arterial Phase

3. Series 3 (perform regional delay if concern for active bleed seen on arterial

phase acquisition) - 2 Min Delayed Phase 3. Send 3 mm Axial Images to PACS

4. Send 2 mm Sagittal and Coronal Reconstructions to PACS

5. Send 3D Reformatted Images to PACS

6. Send thin slices to PACS

7. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With & Without Contrast: AAA





Anatomical Reference: Xyphoid
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

**mAs:** Auto mA (minimum 50-maximum 420)

Scan Delay: Smart Prep (ROI-Descending Aorta just below Arch)

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 130 mL Omnipaque 350 or Isovue 370 mg/mL injected at 5 mL/sec

**Comments:** 1. Series 1 - Without Contrast

2. Series 2 - Arterial Phase

3. Series 3 - Venous Phase - 70 second delay

4. Send 3 mm Axial Images to PACS

5. Send 2 mm Sagittal and Coronal Reconstructions to PACS

6. Send 3D Reformatted Images to PACS

7. Send thin slices to PACS

8. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With & Without Contrast: Mesenteric Ischemic Bowel





Anatomical Reference: Xyphoid
Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 120

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: See below. Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 130 mL Omnipaque 350 or Isovue 370 mg/mL injected at 5 mL/sec

Comments: 1. Series 1 - Without Contrast

2. Series 2 - Enteric Phase - Place ROI on aorta at level of diaphragm. Scan 25

seconds after ROI trigger of 150 HU.

3. Series 3 - 2 minute delays

4. Send 3 mm Axial Images (Series 1-3) to PACS

5. Send 3 mm Sagittal and Coronal Reconstructions (Series 2) to PACS


6. Send 3D Reformatted Images to PACS7. Send thin images (Series 2) to PACS

8. Send thin images (Series 1-3) to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With & Without Contrast: Clinical/laboratory presentation consistent with rapid GI bleed (estimated bleed

rate of ≥ to 0.35 mL/min) which would be detectable with CTA





#### \*To be performed at IMMC only\*

Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Ischial Tuberosities

**Scout:** AP, Lateral if necessary

Scan Type: Helical

**Thickness:** Variable, See comments **Interval:** Variable, See comments

Gantry Tilt: None SFOV: Large kVp: 140

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: Arterial - 25 sec; Venous - 65 sec; Series 4 - 10 min

Respiration: Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 130mL Omnipaque 350 or Isovue 370 mg/ml injected at 5cc/second

Comments: 1. Series 1 - Without Contrast (5mm slice thickness)

2. Series 2 - Arterial Phase - Dome of Diaphragm to Iliac Crest (3 mm slice

thickness)

3. Series 3 - Venous Phase - Dome of Diaphragm to Iliac Crest (3 mm slice

thickness

4. Series 4 - 10 minute Delay - Top of Kidneys through Bladder (3 mm slice

thickness)

5. Send Axial Images to PACS

6. Send 2 mm Sagittal and Coronal Reconstructions (Series 1-3) to PACS

7. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

8. Correct Order: CT CTA Abdomen / Pelvis W/WO contrast

### **Indications**

With & Without Contrast: Examination of Potential Living Donor

\*\*Values will vary between machines. Use your own discretion when selecting these values.\*\*

Reviewed by Drs. King & Smith 10/2024





\* Kidney Transplant Recipient Protocol - For patients already on dialysis and ordered by Drs. Chaudhry and Malik - To be performed at IMMC only\*

Anatomical Reference: **Xyphoid Patient Position:** Supine **Patient Orientation:** Feet First

Scan Range: Dome of Diaphragm to Ischial Tuberosities

AP, Lateral if necessary Scout:

Scan Type: Helical

Thickness: 3 mm, Recon as thin as possible Interval: 3 mm, Recon as thin as possible

**Gantry Tilt:** None SFOV: Large kVp: 120

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: Smart Prep (ROI-Descending Aorta just below Arch)

Respiration: Inspiration

DFOV: Adjust to patient size

**Recon Algorithm:** Standard

Contrast: Oral: None

> IV: 130 mL Omnipaque 350 or Isovue 370 mg/mL injected at 5 mL/sec

1. Series 1 - Without Contrast Comments:

2. Series 2 - Arterial Phase

3. Series 3 - Portal Venous Phase at approximately 1 minute

4. Send 3 mm Axial Images to PACS

5. Send 2 mm Sagittal and Coronal Reconstructions to PACS 6. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With & Without Contrast: Kidney Transplant Recipient - Only used for patients already on dialysis

\*\* For patients not on dialysis - Perform a CT Abd/Pelvis without contrast only\*\*

### CTA Renal Arteries



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid Patient Position: Supine Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Iliac Crest

**Scout:** AP, Lateral if necessary

Scan Type: Helical

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large

kVp: 120 (100 for small patients)

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: Smart Prep (ROI-Descending Aorta just below Arch)

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 130 mL Omnipaque 350 or Isovue 370 mg/mL injected at 5 mL/sec

**Comments:** 1. Series 1 - Without Contrast

2. Series 2 - Arterial Phase

3. Send 3 mm Axial Images to PACS

4. Send 2 mm Sagittal and Coronal Reconstructions to PACS

5. Send 3D Reformatted Images to PACS

6. Send thin slices to PACS

7. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With & Without Contrast: Hypertension, Stenosis

### CTA Runoff



Diagnostic & Preventative Imaging Center

Anatomical Reference: Xyphoid

Patient Position: Supine; TAPE FEET TOGETHER (if possible) to prevent motion

Patient Orientation: Feet First

Scan Range: Dome of Diaphragm to Toes

**Scout:** AP, Lateral if necessary

Scan Type: Helical

**Thickness:** 3 mm, Recon as thin as possible **Interval:** 3 mm, Recon as thin as possible

Gantry Tilt: None SFOV: Large kVp: 140

mAs: Auto mA (minimum 50-maximum 420)

Scan Delay: Smart Prep (ROI-Descending Aorta just below Arch)

**Respiration:** Inspiration

**DFOV:** Adjust to patient size

Recon Algorithm: Standard

Contrast: Oral: None

IV: 150 mL Omnipaque 350 or Isovue 370 mg/mL injected at 5 mL/sec

Comments: 1. Series 1 - Without Contrast - Dome of diaphragm to toes

2. Series 2 - Arterial Phase - Dome of diaphragm to toes

3. Series 3 - Knees to toes - \*ONLY if you see you lost the contrast bolus\* - Set up

prior to series 2 scan just in case you need it

4. Send 3 mm Axial Images to PACS

5. Send 2 mm Sagittal and Coronal Reconstructions (Series 2) to PACS

6. Send 3D Reformatted Images of Aorta and Lower Extremity Arteries to PACS

7. Send thin slices (Series 2 Only) to PACS

8. Send thin slices to 3D Workstation (Vitrea Bridge or TeraRecon)

### **Indications**

With & Without Contrast: Peripheral Arterial Disease, Gangrenous Extremities, Non-existent Distal Pulses

### Coronary CTA



Diagnostic & Preventative Imaging Center

\*\*To be performed at IMMC and Iowa Radiology-Clive only\*\*

Anatomical Reference: Top of shoulders

Patient Position: Supine
Patient Orientation: Feet First

Scan Range: Entire coronary circulation and heart

Scout:APScan Type:HelicalRotation Time:0.33 sec

Acquisition Slice Thickness: Cardiac score=30x0.6 Coronary CTA=64x0.6mm

Pitch: Varied

**Recon Interval (mm):** Cardiac score=3x3 Coronary CTA=0.6mm x 0.3mm

Gantry Tilt: None SFOV: 500 mm kVp: 120

**mAs:** Varied on patient body habitus

Scan Delay: Smart Prep (ROI-Ascending aorta above LCA, Scan Threshold-100 HU)

**Respiration:** Inspiration **DFOV:** Varied

**Recon Algorithm:** B35f heartview medium ca score, mediastinum for cardiac score, I36f heartview

medium ASA, CT angio for coronary CTA (SAFIRE-3)

Contrast: Oral: None

IV: 60 mL Omnipaque 350 mg/mL or equivalent water soluble contrast

injected at 5 mL/sec followed immediately by a mixed bolus of 30mL Omnipaque 350 mg/mL or equivalent water soluble contrast and 30mL

saline injected at 5 mL/sec.

May need to increase contrast amount with larger body habitus.

Injection of 40 mL saline at 5 mL/sec to flush.

Scans: 1. AP Scout

2. Non contrast of entire heart and vessels (calcium score)

3. Monitoring on the ascending aorta just above LCA (Trigger at 100 HU)

4. Gated scan to include entire coronary circulation and heart

**Comments:** 1. All helical scans are gated.

2. Calcium score prior unless indicated otherwise by ordering doctor (PACS)

Qr36/Mediastinum

3. 65% phase (PACS and 3D workstation) 0.6mm x 3mm Bv41/CT Angio 4. 75% phase (PACS and 3D workstation) 0.6mm x 3mm Bv41/CT Angio 5. 80% phase (PACS and 3D workstation) 0.6mm x 3mm Bv41/CT Angio

Continued on next page

### Coronary CTA - Routine Cont.

Comments: 6. 0-90% phase (3D workstation) 1mm x 1mm Bv41/CT Angio

7. Full FOV to include lungs (3D workstation and PACS) 5 mm x 5 mm

Bf39/Mediastinum

8. 40% trigger (PACS and 3D workstation) 0.6mm x 0.3mm Bv41/CT Angio

\*\*\*Manually push ECG gating image to PACS

9. If the ordering office feels the patient's heart rate will run higher than 65 BPM the day of the exam, they can prescribe 50-150 mg of oral Metoprolol for the patient to take 1.5 hours prior to the exam and the evening before the exam. A 0.4 mg Nitroglycerin tablet will be given sublingual 4 minutes prior to imaging.

### **Indications**

With & Without Contrast: Chest Pain, Abnormal Stress Echo, Suspected Congenital Anomalies, Irregular

**Heart Rate** 

Reviewed by Dr. Wolford 1/2025

<sup>\*\*</sup>Values will vary between machines. Use your own discretion when selecting these values.\*\*