

**EU Declaration of Conformity / UKCA
SET 7685 / MSFK B441 / 0.5A** **EU Declaration of Conformity****Safety measuring lead set with fuse**

Order No.	SET 7685 / MSFK B441 / 0.5A
Manufacturer	Schützinger GmbH Eichwiesenring 6 70567 Stuttgart Germany
The sole responsibility for issuing this declaration carries the manufacturer.	
Rated voltage	1000 V CAT IV
Rated current	8 A
Measuring lead	MSFK B441 / 1 / 100 / ..(colour)
Fuse	SICH 6.35x32 1000 V 0.5 A DMI
Description of the product	2 safety measuring leads for fuse and 2 fuse 6.35 x 32 mm
European directives and European regulations	<p>The object of the declaration described above is in conformity with the relevant harmonization legislation of the Union:</p> <ul style="list-style-type: none">• <i>Low Voltage Directive 2014/35/EU</i> <i>published on 29 March 2014 at the EU Official Journal L 96/357-374 Number</i>• <i>RoHS II-Directive 2011/65/EC</i> <i>published on 01 July 2011 at the EU Official Journal L 174/88 Number</i> Applied exemption, Annex III, 6c: copper alloy containing up to 4 % lead by weight• <i>Commission delegated Directive 2015/863/EC</i> <i>published on 06 June 2015 at the EU Official Journal L 137/10 Number</i>• <i>REACH-Regulation No. 1907/2006</i> <i>published on 30 December 2006 at the EU Official Journal L 396/1 Number</i> <i>CAS-Number: 7439-92-1</i> <i>Description: lead, up to 3 %, contained in metal parts</i>

**EU Declaration of Conformity / UKCA
SET 7685 / MSFK B441 / 0.5A**

**Applied harmonized
standards and technical
specifications**

- *IEC/EN 61010-031:2008-08
Sicherheitsbestimmungen für elektrische Mess-, Steuer-,
Regel- und Laborgeräte – Teil 031:
Sicherheitsbestimmungen für handgehaltenes
Messzubehör zum Messen und Prüfen
(IEC 61010-031:2002 + A1:2008)*

Stuttgart, 21.10.2024


.....
Bernhard Schützinger, CEO