


**EU Declaration of Conformity / UKCA
KAB 8214 / ..(colour)**

CE EU Declaration of Conformity

Kelvin adapter box



| | |
|---|--|
| Order No. | KAB 8214 / ..(colour) |
| Manufacturer | Schützinger GmbH Eichwiesenring 6 70567 Stuttgart Germany |
| The sole responsibility for issuing this declaration carries the manufacturer. | |
| Rated voltage | 30 V _{AC} / 60 V _{DC} |
| Rated current | 1 A |
| Description of the product | this kelvin adapter box has been specially developed for four-wire measurement |
| European directives and European regulations | <p>The object of the declaration described above is in conformity with the relevant harmonization legislation of the Union:</p> <ul style="list-style-type: none"> • <i>RoHS II-Directive 2011/65/EC</i> published on 01 July 2011 at the EU Official Journal L 174/88 Number Applied exemption, Annex III, 6c: copper alloy containing up to 4 % lead by weight • <i>Commission delegated Directive 2015/863/EC</i> published on 06 June 2015 at the EU Official Journal L 137/10 Number • <i>REACH-Regulation No. 1907/2006</i> published on 30 December 2006 at the EU Official Journal L 396/1 Number, note: CAS-Number: 7439-92-1 Description: lead, up to 3 %, contained in metal parts |
| Applied harmonized standards and technical specifications | <ul style="list-style-type: none"> • <i>IEC/EN 61010-1:2011-07</i> Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements (IEC 61010-1:2010 + Cor.: 2011) |

EU Declaration of Conformity / UKCA
KAB 8214 / ..(colour)

Stuttgart, 16.10.2024


.....
Bernhard Schützing, CEO

EU Declaration of Conformity / UKCA
KAB 8214 / ..(colour)



Kelvin adapter box

| | |
|---|---|
| Order No. | KAB 8214 / ..(colour) |
| Manufacturer | Schützinger GmbH Eichwiesenring 6 70567 Stuttgart Germany |
| The sole responsibility for issuing this declaration carries the manufacturer. | |
| Rated voltage | 30 V _{AC} / 60 V _{DC} |
| Rated current | 1 A |
| Description of the product | this kelvin adapter box has been specially developed for four-wire measurement |
| UK Statutory requirements | The product is in conformity with the relevant UK Statutory Instruments (and their amendments): <ul style="list-style-type: none">• <i>The Restriction of the Use of certain hazardous substances in Electrical and Electronic Equipment Regulations 2012</i> |

Stuttgart, 16.10.2024


.....
Bernhard Schützinger, CEO