


# Owner's Manual

Infrared Ear Thermometer

Model ST-103(DET-103)



Document No.: JDET-0304-005  
Article No.: 007  
Version: Z  
Date of Issue: 2023.05

## Contents

Introduction .....	02
Product Description .....	03
Lcd Display Introduction .....	04
Basic Functions .....	05
Ear Thermometer Advantages .....	06
Body Temperature .....	07
Real Time Clock Setting .....	08-09
Illustration For Use .....	10-13
Temperature Taking Guidance .....	14-15
Memory Mode .....	16
Selecting A Temperature Scale .....	17
Care And Cleaning .....	18-19
Battery Replacement .....	20
Specifications .....	21
Troubleshooting .....	22-23
Calibration .....	24
Service .....	25
Warranty .....	26
Electromagnetic Compatibility Information .....	27-34

## Introduction

The DET-103 thermometer is specifically designed for safe use in the eardrum. The Infrared Ear Thermometer is a device capable of measuring people's body temperature by detecting the intensity of infrared light emitted from the ear canal of humans. It converts the measured heat into a temperature reading displayed on the LCD. When properly used, it will quickly assess your temperature in an accurate manner. The infrared thermometer is intended for the intermittent measurement of human body temperature by people of all ages.





**Please read all instructions carefully and thoroughly before using this product.**

### Warning:

1. This thermometer is intended for home use only. It's not meant to replace a visit to the doctor. Please consult with doctor if you have health concerns.
2. There is no gender or age limitation for using infrared ear thermometer.
3. To avoid inaccurate readings, always attach a new, clean probe cover for each temperature reading.
4. Do not expose the thermometer to temperature extremes (below -25°C/-13°F or over 55°C/131°F) nor excessive humidity (>95%RH).
5. This thermometer must only be used with Seagull Healthcare thermometer probe covers.
6. Keep the battery away from children.
7. Remove batteries from the device when not in operation for a long time.


## Product Description

1. Probe Cover
2. Probe
3. TEST Button
4.  Button
5. Battery Cover
6. Eject button



3

## LCD Display Introduction



4

## Basic Functions

Real Time Clock	The real time clock will be recorded with the memory function and help you to recognize each measurement result → Please see the Real time clock setting section to learn how to setup the time in the first use
Ear Temperature Mode	The thermometer has been designed for practical use. It's not meant to replace a visit to the doctor. Please also remember to compare the measurement result to your regular body temperature → Please see the Illustration For Use section to learn how to measure the body temperature
Probe cover detection and eject function	The thermometer has been designed for use with probe cover Users can press the eject button to remove the probe cover
Object Temperature Mode	The object mode shows the actual, unadjusted surface temperatures, which is different from the body temperature. It can help you to monitor if the object temperature is suitable for the baby or patient, for example the baby's milk → Please see the Illustration For Use section to learn how to measure the object temperature
Alarm	If thermometer detects a body temperature $\geq 37.8^{\circ}\text{C}$ ( $100.0^{\circ}\text{F}$ ) under ear mode, there will be a long beep sound followed by three short beep sound to warn the user
Memory Mode	There are each 10 sets memories for ear and object measurements Each memory also records the measurement date/time/mode icon
°C/°F Switch	Please see the Selecting A Temperature Scale section to learn how to change between Celsius and Fahrenheit

5

## Ear Thermometer Advantages

Infrared Ear Thermometer measures core body temperature, which is the temperature of a body's vital organs. (See Figure 1) Ear temperatures accurately reflect core body temperatures since the eardrum is in the cranial cavity and is not affected by the environment. The eardrum shares blood supply with the hypothalamus (part of the brain.) Therefore, body temperature changes are reflected sooner in the ear than they are in other parts of the body.




Figure 1

6

## Body Temperature

The temperature of a healthy person is affected by various factors: the person's individual metabolism, their age (body temperature is higher in babies and toddlers and decreases with age. Greater temperature fluctuations occur faster and more often in children, e.g. due to growth spurts), their clothing, the ambient temperature, the time of day (body temperature is lower in the morning and increases throughout the day towards evening), the preceding physical and, to a lesser extent, mental activity.

It is suggested that users develop the habit of daily measurement, and use this product to establish a normal temperature range for themselves and their families as a reference for measurement.

7

## Real Time Clock Setting

When using thermometer for the first time, please set the parameters of the thermometer. With the thermometer off, press and hold **⏻** *BUTTON* to enter into setting mode for 3 seconds.



① Set the time format  
The device can display the time in either an AM/PM (12-hour) or a 24:00 (24-hour) format. Press and release *TEST BUTTON* to select the format. With the preferred time format on the display, press **⏻** *BUTTON*, and the Hour symbol will flash automatically.



② Set the hour  
Press and release the *TEST BUTTON* to advance one hour until the correct hour appears. After the hour is set, press **⏻** *BUTTON*, and the Minute symbol will flash automatically.



③ Set the minute  
Press and release the *TEST BUTTON* to advance one minute until the correct minute appears. After the minute is set, press **⏻** *BUTTON*, the Year figure is flashing automatically.

8

## Real Time Clock Setting



④ Set the year  
Press and release the *TEST BUTTON* to advance one year until the correct year appears. After the year is set, press **⏻** *BUTTON*, and the Month symbol will then appear.



⑤ Set the month  
Press and release the *TEST BUTTON* to advance one month until the correct month appears. After the month is set, press **⏻** *BUTTON*, and the Date symbol will now flash.



⑥ Set the date  
Press and release the *TEST BUTTON* to advance one day until the correct month appears. After the day is set, press **⏻** *BUTTON* to exit the setting mode.

9

## Illustration For Use

### ► To measure ear temperature:

- To achieve accurate readings, make sure a new, clean probe cover is in place before each measurement. Place a new probe cover on the probe firmly; making sure the probe and the inner portion of the probe cover fit snugly.
- Press the **⏻** *BUTTON*, The display is activated to show all segments. After self-checking Figure 2 appears on the display screen with beeps, so you can start a new measurement. If error messages appears, it means the thermometer is not ready for measurement.

Remarks:

- DET-103 must be used with probe cover, if not installed with probe cover, then "CAP" appears on the display. (as shown in Figure 3)
- To avoid the risk of cross contamination, it is recommended that you always measure the ear temperature with probe cover.



Figure 2




Figure 3

10

## Illustration For Use

3. Insert the probe along the ear canal as deep as possible into the ear as depicted in Figure 4. Then press the *TEST BUTTON* once to start the measurement. A long beep will sound approximately 1 second after the *TEST BUTTON* is activated. This signals the end of the process as results are displayed on the LCD screen.
4. The thermometer is ready for testing again once an ear mark displays on the LCD. A waiting period of 20 seconds between testing is recommended to avoid excessive cooling of the skin.
5. The thermometer will shut off automatically after 60 seconds of inactivity. To prolong battery life, press the *ON BUTTON* to turn the unit off.




Figure 4

11

## Illustration For Use

### ► To measure object temperature:

1. First eject the probe cover.
2. Aim the thermometer at the center of the object you want to measure with a distance of 1 to 2 cm.
3. Press the *ON BUTTON* to turn on the thermometer, you can take the object temperature after hearing two beep sounds. (see figure 5)

Remarks:

This mode shows the actual, unadjusted surface temperatures, which is different from the body temperature.



Figure 5

### ► How to change the ear mode and object mode:

You can press and hold *ON BUTTON* and then press and hold the *TEST BUTTON* to switch the mode between ear mode and object mode.

12

## Illustration For Use

### ► After measurement:

1. Power off: Device will automatically shut off if left idle for more than 1 minute to extend battery life.
2. Clean the probe after each use to ensure an accurate reading and avoid cross contamination.  
(See the section of Care and Cleaning for details.)

### ► BackLight:

In Ear mode:

1. The display will be lighted GREEN for 3 seconds with a happy face when the unit is ready for measurement and a measurement is completed with a reading less than 37.3 (99.1).
2. The display will be lighted YELLOW for 3 seconds with a happy face when a measurement is completed with a reading less than 37.8°C(100.0).
3. The display will be lighted RED for 3 seconds with a bad face when a measurement is completed with a reading equal to or higher than 37.8°C(100.0).

In Object mode:

The display will only be lighted GREEN for 3 seconds when the unit is ready for measurement and a measurement is completed

13

## Temperature Taking Guidance

1. The right ear reading may differ from the reading taken at the left ear. Therefore, always take the temperature in the same ear.
2. The ear must be free from obstructions or excess earwax buildup to take an accurate reading.
3. External factors may influence ear temperatures such as:
  - lying on one ear or the other
  - their ears covered
  - exposed to very hot or very cold temperatures
  - recently swimming or bathing

In these cases, remove the individual from the situation and wait 20 minutes prior to taking a temperature.
4. For persons wearing hearing aids or ear plugs, remove the device and wait 20 minutes prior to taking a temperature.
5. When using the thermometer on infants under age 1, pull the ear back making sure the sensor faces the eardrum.  
(See Figure 6)

14

## Temperature Taking Guidance

- When using the thermometer on individuals over the age of 1, pull the ear up making sure the sensor faces the eardrum. (See Figure 7)
- Do not touch the probe window. A cotton swab and rubbing alcohol may be used to clean the sensor window.
- If the thermometer is stored in a significantly different environment than testing location, place it in the testing location for approximately 30 minutes prior to use.




Figure 6



Figure 7

15

## Memory Mode

- The Memory Mode can be accessed either in ear mode or object mode:  
When the thermometer has been turned on and followed by Figure 2/5 or finished testing, press and hold the **MEM** Button for three seconds. The letter MEM will appear in the center left corner of the display. (See Figure 8)
- The thermometer will automatically memorize the last 10 temperature readings. Each memory also records the measurement date/time/mode icon. Each time the **MEM** Button is pressed, the screen displays past readings that correspond with a number 1-10. The number 1 reflects the most recent reading, while the number 10 reveals the oldest reading stored in memory. (See Figure 9)
- In the memory mode, ear mark or scan mark always exist. The user can press the **TEST BUTTON** to take new measurements.




Figure 8



Figure 9

16

## Selecting A Temperature Scale

- Temperature readings are available in the Celsius ( $^{\circ}\text{C}$ ) or Fahrenheit ( $^{\circ}\text{F}$ ) scale.
- With the unit off, press and hold the **TEST BUTTON** for 3 seconds to enter into unit changing mode.
- Press and release **TEST BUTTON** to select the unit.
- When the preferred unit on the display, press **MEM** BUTTON to exit the unit changing mode.




Figure 10

17

## Care And Cleaning

- Probe covers are intended for single use only!  
Absent, damaged, perforated, or soiled probe covers can lead to inaccurate temperature readings.
- The probe window must be kept clean, dry, and undamaged at all times to ensure accurate readings. The accuracy of temperature readings can be affected by damage to the probe window, or the presence of dirt and ear wax on the probe window.
- Fingerprints, earwax, dust and other soiling compounds reduce transparency of the window and result in lower temperature readings. If the thermometer is accidentally placed in the ear without a probe cover installed, the window must be cleaned immediately.
- The probe window is the most delicate part of the thermometer.  
To safely clean the window, gently wipe its surface with a cotton swab slightly moistened with isopropyl alcohol and immediately wipe dry with a clean cotton swab. After cleaning, allow at least 5 minutes drying time before taking temperatures.  
Note: Do not use any chemical other than isopropyl alcohol to clean the probe window.
- Use a soft, dry cloth to clean the thermometer display and exterior.
- The thermometer is not waterproof. Do not submerge the unit in water when cleaning.

18

## Care And Cleaning

7. Store the thermometer and probe cover in a dry location, free from dust and contamination and away from direct sunlight.
8. Periodic cleaning and disinfection of the device following use to prevent patient cross infection.  
-Use a soft cloth slightly moistened with a 75% isopropyl alcohol solution to disinfect the thermometer and probe. Do not use abrasive cleaners.
9. Ensure that children do not use the instrument unsupervised; some parts are small enough to be swallowed.
10. Do not remove or modify the equipment without permission.
11. Strong electromagnetic fields may interfere with the proper operation of the thermometer. The device needs special pre-cautions regarding EMC according to the EMC information.
12. Put the thermometer back to the original packaging after using.
13. It is not intended for use in the oxygen rich environment and presence of flammable anesthetic mixture with air, oxygen or nitrous oxide.

19

## Battery Replacement


1. Replace battery when "  " appears in the upper right corner of LCD display. (See Figure 11)
2. Slide battery cover down as shown in Figure 12.
3. Remove battery and install 2 new AAA alkaline batteries as shown in Figure 13 .
4. Slide battery cover back on.



Figure 11




Figure 12




Figure 13








20

## Specifications

Measuring range	Ear mode: 32.0°C~43.0°C(89.6°F~109.4°F) Object mode: 0°C~100°C(32°F~212°F)
Measuring site	Ear canal(Ear Mode)
Operation mode	Ear mode(Adjust mode)
Laboratory accuracy	Ear mode: ±0.2°C (0.4°F) during 35.5°C~42.0°C (95.9°F~107.6°F) at 15°C~35°C (59.0°F~95.0°F) operating temperature range ±0.3°C (0.5°F) for other measuring and operating temperature range Object mode: ±4% or ±2°C(4°F) whichever is greater
Display resolution	0.1°C or 0.1°F
Measure time	Approximately one second
Operating temperature range:	Ear/Object mode: 10°C~40°C(50°F~104°F), 15%~85%RH, non-condensing Atmospheric Pressure : 700hPa ~ 1060hPa
Storage and transport temperature range	-25°C~55°C (-13°F~131°F), 15%~95%RH, non-condensing Atmospheric Pressure : 700hPa ~ 1060hPa
Clinical accuracy	Clinical bias: 0.08°C (0.14°F) Clinical repeatability: 0.13°C (0.23°F) Limits of agreement:0.73°C (1.31°F)
Shock	withstands drop of 3 feet
Dimension	155 x 38x 47mm
Weight	Approx.83grams(with batteries)
Battery	DC3V(2×AAA battery)
Battery life	Approx.6000 readings
Ingress protecting rating	IP22

21

## Troubleshooting

Error message	Problem	Solution
	Measurement before thermometer is ready	Take a measurement until  or  appears on the display
	The ambient temperature is not within the range between 10°C and 40°C (50°F~104°F)	Place the thermometer in a room for at least 30 minutes at room temperature between 10°C and 40°C (50°F~104°F)
	The thermometer is placed incorrectly or unsteady	Read Illustration For Use thoroughly and take a new temperature measurement
	The thermometer showing a rapid ambient temperature change	Allow the thermometer to rest in a room for at least 30 minutes at room temperature: between 10°C and 40°C (50°F~104°F)
	The thermometer is not functioning properly	Unload the battery, wait for 1 minute and repower it. If the message reappears, contact the retailer for service

22

## Troubleshooting

Error message	Problem	Solution
	In object temperature mode, measurement with probe cover	Remove probe cover in object temperature mode
	In Ear mode: Temperature taken is higher than 43.0 °C (109.4°F) In Object mode: Temperature taken is higher than 100 °C (212°F)	Read Temperature Taking Hints Thoroughly, then check the integrity of the probe cover and take a new temperature measurement
	In Ear mode: Temperature taken is lower than 32.0 °C (89.6°F) In Object mode: Temperature taken is lower than 0 °C (32°F)	Read Temperature Taking Hints thoroughly, then make sure the probe cover and lens filter are clean, then take a new temperature measurement
	The thermometer works properly	Use the thermometer normally
	When battery outline flashes, it indicates that the power is low, but you can continue to measure	The thermometer will take a proper measurement but batteries must be replaced soon
	The thermometer could not work due to low battery	Replace two new alkaline batteries size AAA
	Probe cover was not used when taking ear temperature	Please use probe cover when measuring in ear temperature mode

23

## Calibration

The user must always comply with legal requirements for the control of the measurement, functionality, and accuracy of the device which are required by the scope of relevant laws, directives or ordinances where the device is used.



Type BF applied part



Direct Current



Disposal of this product and used batteries should be carried out in accordance with the national regulations for the disposal of electronic products



Consult Accompanying Documents



Batch Code

55°C Storage and Transportation Temperature Limit:  
-25°C -13°F ~ 55°C ~ 131°F (-25°C ~ 55°C)

24

## Service

The thermometer has a limited two years warranty. Do not attempt to disassemble or repair the thermometer by yourself. Should service be required during the warranty period you must contact the manufacturer. Repackage the thermometer carefully in its original packaging or securely pack to avoid damage during shipping. Include the original sales slip indicating the date of purchase, a note describing the problem, and your return address. Send the thermometer prepaid and insured. The lay operator or lay responsible organization should contact the manufacturer or the manufacturer's representative:

- for assistance, if needed, in setting up, using or maintaining the thermometer; or
- to report unexpected operation or events.



Manufacturer:  
JOYTECH HEALTHCARE Co., Ltd.  
No.365, Wuzhou Road, Yuhang Economic Development Zone, Hangzhou city,  
311100 Zhejiang, China

Telephone: +86-571-81957767  
Fax: +86-571-81957750

Importer:  
Seagull Healthcare Aps, Industrivej 1, DK-4200 Slagelse.



The product is in compliance with the requirements of MDR(EU) 2017/745, "0123" is the identification number of notify body



European Authorized Representative:  
Shanghai International Holding Corp. GmbH (Europe)  
Eiffestrasse 80, 20537 Hamburg, Germany

25

## Warranty

This appliance conforms to the following standards:  
EN 12470-5 Clinical thermometers —Part 5:Performance of infrared ear thermometers(with maximum device),  
ISO 80601-2-56 Medical electrical equipment —Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement,  
IEC 60601-1-11 Medical electrical equipment —Part 1-11: General requirements for basic safety and essential performance –Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment and complies with the requirements of IEC 60601-1-2(EMC), IEC 60601-1(Safety) standards. And the manufacturer is ISO 13485 certified.

Thermometer is warranted by manufacture to be free from defects in material and workmanship under normal use and service for a period of one year from the date of delivery to the first user who purchases the instrument. This warranty does not cover batteries, damage to the probe window, or damage to the instrument caused by misuse, negligence or accident, and extends to only to the first purchaser of the product. Additionally this warranty becomes void, if the thermometer is operated with anything other than this brand thermometer probe covers.

26

## Electromagnetic Compatibility Information

The device satisfies the EMC requirements of the international standard IEC 60601-1-2. The requirements are satisfied under the conditions described in the table below. The device is an electrical medical product and is subject to special precautionary measures with regard to EMC which must be published in the instructions for use. Portable and mobile HF communications equipment can affect the device. Use of the unit in conjunction with non-approved accessories can affect the device negatively and alter the electromagnetic compatibility. The device should not be used directly adjacent to or between other electrical equipment.

27

## Electromagnetic Compatibility Information

Table 1

Guidance and manufacturer's declaration – electromagnetic emission		
The device is intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment		
Emissions test	Compliance	Electromagnetic environment – guidance
RF emissions CISPR 11	Group 1	The device uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment
RF emissions CISPR 11	Class B	The device is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes
Harmonic emissions IEC 61000-3-2	Not applicable	
Voltage fluctuations / flicker emissions IEC 61000-3-3	Not applicable	

28

## Electromagnetic Compatibility Information

Table 2


Guidance and manufacturer's declaration – electromagnetic immunity			
The device is intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Electrostatic discharge (ESD) IEC 61000-4-2	± 8 kV contact ±2 kV, ±4 kV, ±8 kV, ±15 kV air	± 8 kV contact ±2 kV, ±4 kV, ±8 kV, ±15 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %
Electrostatic transient / burst IEC 61000-4-4	± 2 kV for power supply lines 100 kHz repetition frequency ± 1 kV for input/output lines	N/A	N/A
Surge IEC 61000-4-5	± 0.5 kV, ± 1 kV differential mode line-line	N/A	N/A
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	0 % UT (100 % dip in UT) for 0.5 cycle at 0°, 45°, 90°, 135°, 180°, 225°, 270°, and 315° 0 % UT (100 % dip in UT) for 1 cycle at 0° 70 % UT (30 % dip in UT) for 25/30 cycles at 0° 0 % UT (100 % dip in UT) for 250/300 cycle at 0°	N/A	N/A
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	30 A/m, 50/60Hz	30 A/m, 50/60Hz	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment

NOTE: UT is the a. c. mains voltage prior to application of the test level

29

## Electromagnetic Compatibility Information

Table 3

Guidance and manufacturer's declaration – electromagnetic immunity			
The device is intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz 6 Vrms 150 kHz to 80 MHz outside ISM bands	N/A	Portable and mobile RF communications equipment should be used no closer to any part of the device, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter <b>Recommended separation distance</b> $d = \left[ \frac{3.5}{V_1} \right] \sqrt{P}$
Radiated RF IEC 61000-4-3	10 V/m 80 MHz to 2.7 GHz	10 V/m	$d = \left[ \frac{3.5}{E_1} \right] \sqrt{P} \quad 80\text{MHz to } 800\text{MHz}$ $d = \left[ \frac{7}{E_1} \right] \sqrt{P} \quad 800\text{MHz to } 2.7\text{GHz}$ where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres(m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, <sup>a</sup> should be less than the compliance level in each frequency range <sup>b</sup> Interference may occur in the vicinity of equipment marked with the following symbol: 

30

# Electromagnetic Compatibility Information

Table 3 continued

NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies. NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people

a The ISM (industrial, scientific and medical) bands between 0,15 MHz and 80 MHz are 6,765 MHz to 6,795 MHz; 13,553 MHz to 13,567 MHz; 26,957 MHz to 27,283 MHz; and 40,66 MHz to 40,70 MHz. The amateur radio bands between 0,15 MHz and 80 MHz are 1,8 MHz to 2,0 MHz; 3,5 MHz to 4,0 MHz; 5,3 MHz to 5,4 MHz; 7 MHz to 7,3 MHz; 10,1 MHz to 10,15 MHz; 14 MHz to 14,2 MHz; 18,07 MHz to 18,17 MHz; 21,0 MHz to 21,4 MHz; 24,89 MHz to 24,99 MHz; 28,0 MHz to 29,7 MHz and 50,0 MHz to 54,0 MHz

b The compliance levels in the ISM frequency bands between 150 kHz and 80 MHz and in the frequency range 80 MHz to 2,7 GHz are intended to decrease the likelihood that mobile/portable communications equipment could cause interference if it is inadvertently brought into patient areas. For this reason, an additional factor of 10/3 has been incorporated into the formulae used in calculating the recommended separation distance for transmitters in these frequency ranges

c Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the device is used exceeds the applicable RF compliance level above, the device should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the device

d Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m

# Electromagnetic Compatibility Information

Table 4

Recommended separation distances between portable and mobile RF communications equipment and the device			
The device is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the device can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the device as recommended below, according to the maximum output power of the communications equipment			
Rated maximum output of transmitter W	Separation distance according to frequency of transmitter m		
	150 kHz to 80 MHz $d = \left[ \frac{3,5}{V_1} \right] \sqrt{P}$	80 MHz to 800 MHz $d = \left[ \frac{3,5}{E_1} \right] \sqrt{P}$	800 MHz to 2.7 GHz $d = \left[ \frac{7}{E_1} \right] \sqrt{P}$
0.01	0.12	0.04	0.07
0.1	0.37	0.12	0.23
1	1.17	0.35	0.7
10	3.7	1.11	2.22
100	11.7	3.5	7.0


For transmitters rated at a maximum output power not listed above the recommended separation distance d in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer

NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people

# Electromagnetic Compatibility Information

Table 5

Recommended separation distances between RF wireless communications equipment					
The device is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the device can help prevent electromagnetic interference by maintaining a minimum distance between RF wireless communications equipment and the device as recommended below, according to the maximum output power of the communications equipment					
Frequency MHz	Maximum Power W	Distance	IEC 60601 Test Level	Compliance Level	Electromagnetic Environment - Guidance
385	1.8	0.3	27	27	RF wireless communications equipment should be used no closer to any part of the device, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter <b>Recommended separation distance</b> $E = \frac{6}{d} \sqrt{P}$ Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitter, as determined by an electromagnetic site survey, should be less than the compliance level in each frequency range. Interference may occur in the vicinity of equipment marked with the following symbol: 
450	2	0.3	28	28	
710	0.2	0.3	9	9	
745					
780					
810					
870	2	0.3	28	28	
930					
1720					
1845	2	0.3	28	28	
1970					
2450	2	0.3	28	28	
5240					
5500					
5785					

Note 1: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people

# Electromagnetic Compatibility Information

## WARNINGS!

- This device should not be used in the vicinity or on the top of other electronic equipment such as cell phone, transceiver or radio control products. If you have to do so, the device should be observed to verify normal operation.
- The use of accessories and power cord other than those specified, with the exception of cables sold by the manufacturer of the equipment or system as replacement parts for internal components, may result in increased emissions or decreased immunity of the equipment or system.