

MRA Head W & WO		
Technical Parameters		
Anatomic Coverage:	Vertex to skull base; frontal through occipital, skin to skin	
Coil:	Head	
Scanner Preference:	1.5T or 3T	
Contrast:	Pre-injection:	Post-injection:
Scanning Plane:	COW	COW
Slice Thickness:	1mm	1mm
Gap:	0mm	0mm
Weighting:	3D TOF	3D TOF
Mode:	3D FFE	3D FFE
FAT SAT:	NO	NO
FOV:	20cm	20cm
MPR:	NO	NO
Notes:	MIP - COW, Right, Left and Posterior	