MRI Brain & Orbits WO							
Technical Parameters							
Anatomic Coverage:	Skull base to vertex						
Coil:	Head						
Scanner Preference:							
Scanning Plane:	Axial	Sagittal	Sagittal	Axial	Axial	Coronal	Coronal
Slice Thickness:	3mm	0.7-1.0mm	1mm	1mm	2mm	2mm	2mm
Gap:	0.3mm	0mm	0mm	0mm	-1mm	0.25mm	0.25mm
Weighting:	DWI	T1	T2 FLAIR	T2	SWI	STIR	T1
Mode:	EPI	3D TSE	3D IR-TSE	3D TSE	3D GRE	2D STIR	2D TSE
FAT SAT:	YES	NO	YES	NO	NO	STIR	DIXON: IN PHASE
FOV:	23cm	23cm	23cm	23cm	23cm	17cm	17cm
MPR:	NO	NO	NO	NO	NO	NO	NO
Notes:							