

MRI Brain MS WO					
Technical Parameters					
Anatomic Coverage:	Skull base to vertex				
Coil:	Head				
Scanner Preference:	1.5T or 3T				
Scanning Plane:	Sagittal	Sagittal	Axial	Axial	Axial
Slice Thickness:	1.2mm	1mm	4mm	3mm	2mm
Gap:	0mm	0mm	1mm	0.3mm	-1mm
Weighting:	T1	T2 FLAIR	T2	DWI	SWI
Mode:	3D TFE	3D IR-TSE	2D TSE	2D EPI	3D GRE
FAT SAT:	NO	YES	NO	YES	NO
FOV:	24-25.6cm	23cm	23cm	23cm	23cm
MPR:	NO	AXIAL, COR	NO	NO	NO
Notes:					