

MRI Brain Seizure WO

Technical Parameters

Anatomic Coverage:	Skull base to vertex					
Coil:	Head					
Scanner Preference:	1.5 T or 3T					
Scanning Plane:	Axial	Sagittal	Coronal Oblique	Coronal Oblique	Sagittal	Axial
Slice Thickness:	3mm	1.2mm	2mm	2.5mm	1mm	2mm
Gap:	0.3mm	0mm	0.2mm	0.25mm	0mm	-1mm
Weighting:	DWI	T1	T2 STIR	T2	T2 FLAIR	SWI
Mode:	EPI	3D TFE	2D STIR TSE	2D TSE	3D IR-TSE	3D GRE
FAT SAT:	YES	NO	YES	NO	YES	NO
FOV:	23cm	24-25.6cm	23cm	15cm	23cm	23cm
MPR:	NO	NO	NO	NO	AXIAL, COR OBLIQ	NO
Notes:						