MRI Brain Stroke WO								
Technical Parameters								
Anatomic Coverage:	Skull base to vertex							
Coil:	Head							
Scanner Preference:	1.5T or 3T							
Scanning Plane:	Axial	Coronal	Axial	Sagittal	Axial	Sagittal	Axial	Axial
Slice Thickness:	3mm	3mm	4mm	4mm	4mm	1mm	6mm	6mm
Gap:	0.3mm	0.3mm	1mm	1mm	1mm	0	1mm	1mm
Weighting:	DWI	DWI	T2	T1	T2	T2 FLAIR	ASL Single Phase	ASL Multi Phase
Mode:	EPI	EPI	GRE	TSE	TSE	3D IR-TSE	FFE EPI	FFE EPI
FAT SAT:	YES	YES	YES	NO	NO	YES	NO	NO
FOV:	23cm	23cm	23cm	24-25.6cm	23cm	23cm	23cm	23cm
MPR	NO	NO	NO	NO	NO	AX, COR, OBLIQ	NO	NO
Notes:								