MRI Brain WO							
Technical Parameters							
Anatomic Coverage:	Skull base to vertex						
Coil:	Head						
Scanner Preference:	1.5T or 3T						
Scanning Plane:	Axial	Axial	Axial	Axial	Axial	Coronal	Sagittal
Slice Thickness:	3mm	1mm	1mm	2mm	1mm	1mm	1mm
Gap:	0.3mm	0mm	0mm	-1mm	0mm	0mm	0mm
Weighting:	DWI (ADC & EADC maps)	T1	T2	SWI	T2	T2	FLAIR
Mode:	EPI	3D TFE	3D TSE	3D GRE	FLAIR	TSE	3D IR-TSE
FAT SAT:	YES	NO	NO	NO	NO	NO	YES
FOV:	23cm	23cm	23cm	23cm	23cm	23cm	23cm
MPR	NO	NO	NO	NO	NO	NO	NO
Notes:							