

MRI L-Spine WO				
Technical Parameters				
Anatomic Coverage:	T11-T12 to S2-S3			
Coil:	Torso and/or NV			
Scanner Preference:	1.5T or 3T			
Scanning Plane:	Sagittal	Sagittal	Sagittal	Axial
Slice Thickness:	1mm	3mm	3mm	4mm
Gap:	0mm	0.3mm	0.3mm	1mm
Weighting:	T2	T2	T1 FLAIR	T2
Mode:	3D TSE	2D TSE	2D IR-TSE	2D TSE
FAT SAT:	NO	DIXON: IN PHASE	NO	NO
FOV:				20x20cm
MPR:	AXIAL 1/0	NO	NO	NO
Notes:				