

MRI Neck Soft Tissue WO				
Technical Parameters				
Anatomic Coverage:	Mid orbits to mid aortic arch			
Coil:	NV or Spine			
Scanner Preference:				
Scanning Plane:	Sagittal	Axial	Coronal	Axial
Slice Thickness:	0.7mm	3mm	2.5mm	3mm
Gap:	0mm	0.3mm	0.25mm	0.3mm
Weighting:	T1	DWI	T2	T2
Mode:	3D TSE	EPI	2D TSE	2D TSE
FAT SAT:	NO	YES	DIXON: IN PHASE	DIXON: IN PHASE
FOV:	21cm	22cm	22cm	21cm
MPR:	NO	NO	NO	NO
Notes:				