

MRV Brain WO		
Technical Parameters		
Anatomic Coverage:	Vertex to skull base; frontal through occipital, skin to skin	
Coil:	Head	
Scanner Preference:	1.5T or 3T	
Scanning Plane:	Axial	Coronal
Slice Thickness:	2mm	2mm
Gap:	-0.5mm	-0.5mm
Weighting:	T1	T1
Mode:	M2D FFE	M2D FFE
FAT SAT:	NO	NO
FOV:	23cm	23cm
MPR:	NO	NO
Notes:		