

MRI Elbow WO				
Technical Parameters				
Anatomic Coverage:	4cm above and below elbow joint			
Coil:	Knee/Flex			
Scanner Preference:	1.5T or 3T			
Scanning Plane:	Axial	Axial	Coronal	Sagittal
Slice Thickness:	3mm	3mm	3mm	3mm
Gap:	0.5mm	0.5mm	0.5mm	0.5mm
Weighting:	PD FS	T1	PD FS	PD FS
Mode:	TSE	TSE	TSE	TSE
FAT SAT:	YES	NO	YES	YES
FOV:	14cm	14cm	14cm	14cm
MPR:	NO	NO	NO	NO
Notes:				

