

MRI Femur Right/Left WO					
Technical Parameters					
Anatomic Coverage:	Femoral head to knee				
Coil:	Torso				
Scanner Preference:	1.5T or 3T				
Scanning Plane:	Axial	Axial	Coronal	Coronal	Sagittal
Slice Thickness:	6mm	6mm	5mm	5mm	5mm
Gap:	1mm	1mm	1mm	1mm	1mm
Weighting:	T1	PD FS	T1	STIR	STIR
Mode:	TSE	TSE	TSE	TSE	TSE
FAT SAT:	NO	YES	NO	YES	YES
FOV:	40cm	40cm	40cm	40cm	40cm
MPR:	NO	NO	NO	NO	NO
Notes:					

