

MRI Finger (Not Thumb) WO				
Technical Parameters				
Anatomic Coverage:	Tip of finger soft tissue to distal metacarpal			
Coil:	Flex or Head			
Scanner Preference:	1.5T or 3T			
Scanning Plane:	Axial	Axial	Coronal	Sagittal
Slice Thickness:	3mm	3mm	3mm	3mm
Gap:	0.5mm	0.5mm	0.5mm	0.5mm
Weighting:	PD	PD FS	STIR	PD FS
Mode:	TSE	TSE	TSE	TSE
FAT SAT:	NO	YES	YES	YES
FOV:	12cm	12cm	12cm	12cm
MPR:	NO	NO	NO	NO
Notes:				

