

MRI Forearm WO					
Technical Parameters					
Anatomic Coverage:	Elbow through wrist joints				
Coil:	XL Torso				
Scanner Preference:	1.5T				
Scanning Plane:	Axial	Axial	Coronal	Coronal	Sagittal
Slice Thickness:	4mm	4mm	4mm	4mm	4mm
Gap:	1mm	1mm	1mm	1mm	1mm
Weighting:	T1	PD FS	T1	STIR	T2
Mode:	TSE	TSE	TSE	TSE	TSE
FAT SAT:	NO	YES	NO	YES	NO
FOV:	14cm	14cm	14cm	14cm	14cm
MPR:	NO	NO	NO	NO	NO
Notes:					

