

MRI Hand WO						
Technical Parameters						
Anatomic Coverage:	Area of interest					
Coil:	Flex or Head					
Scanner Preference:	1.5T or 3T					
Scanning Plane:	Axial	Axial	Coronal	Coronal	Sagittal	Sagittal
Slice Thickness:	2mm	2mm	2mm	2mm	2mm	2mm
Gap:	0.3mm	0.3mm	0.3mm	0.3mm	0.3mm	0.3mm
Weighting:	PD	PD FS	T1	STIR	PD	PD FS
Mode:	TSE	TSE	TSE	TSE	TSE	TSE
FAT SAT:	NO	YES	NO	YES	NO	YES
FOV:	12cm	12cm	12cm	12cm	12cm	12cm
MPR:	NO	NO	NO	NO	NO	NO
Notes:						