

MRI Thumb W & WO				
Technical Parameters				
Anatomic Coverage:	Area of contrast			
Coil:	Flex or Head			
Scanner Preference:	1.5T or 3T			
Contrast:	Pre-injection:	Post-injection:		
Scanning Plane:	Axial	Axial	Coronal	Sagittal
Slice Thickness:	3mm	3mm	3mm	3mm
Gap:	0.5mm	0.5mm	0.5mm	0.5mm
Weighting:	T1 FS	T1 FS	T1 FS	T1 FS
Mode:	TSE	TSE	TSE	TSE
FAT SAT:	YES	YES	YES	YES
FOV:	12cm	12cm	12cm	12cm
MPR:	NO	NO	NO	NO
Notes:				

