

MRI Thumb WO					
Technical Parameters					
Anatomic Coverage:	Area of contrast				
Coil:	Flex or Head				
Scanner Preference:	1.5T or 3T				
Scanning Plane:	Axial	Axial	Coronal	Coronal	Sagittal
Slice Thickness:	3mm	3mm	3mm	3mm	3mm
Gap:	0.5mm	0.5mm	0.5mm	0.5mm	0.5mm
Weighting:	PD	PD FS	PD	PD FS	STIR
Mode:	TSE	TSE	TSE	TSE	TSE
FAT SAT:	NO	YES	NO	YES	YES
FOV:	12cm	12cm	12cm	12cm	12cm
MPR:	NO	NO	NO	NO	NO
Notes:					

