

MRI Wrist WO					
Technical Parameters					
Anatomic Coverage:	3cm above and below the wrist				
Coil:	Wrist/Flex				
Scanner Preference:	1.5T or 3T				
Scanning Plane:	Axial	Axial	Coronal	Coronal	Sagittal
Slice Thickness:	2-3mm	2-3mm	2-3mm	2-3mm	2-3mm
Gap:	0.5mm	0.5mm	0.5mm	0.5mm	0.5mm
Weighting:	PD	PD FS	PD	PD FS	PD FS
Mode:	TSE	TSE	TSE	TSE	TSE
FAT SAT:	NO	YES	NO	YES	YES
FOV:	12cm	12cm	12cm	12cm	12cm
MPR:	NO	NO	NO	NO	NO
Notes:					

