

MRI Breast Bilateral WO				
Technical Parameters				
Anatomic Coverage:	Cover entire breast tissue and lymph nodes only			
Coil:	Breast coil			
Scanner Preference:				
Scanning Plane:	Axial	Sag (silicone implants)	Axial (silicone implants)	Axial
Weighting:	T2	T2	T2	T1
FAT SAT:	YES	YES	YES	NO
FOV:	ONLY breast tissue and lymph nodes	ONLY breast tissue and lymph nodes	ONLY breast tissue and lymph nodes	ONLY breast tissue and lymph nodes
MPR:	NO	NO	NO	NO
Notes:				