MRA Renal Arteries									
Technical Parameters									
Anatomic Coverage:	AP coverage from behind aorta through entire kidneys								
Coil:	None								
Scanner Preference:									
Contrast:	Pre-injection:					Post-injection:			
Scanning Plane:	Coronal	Coronal	Axial	Axial	Coronal	Coronal	Coronal	Coronal	Axial
Slice Thickness:	4mm	4mm	6mm	3mm	1.1mm	1.1mm	1.1mm	1.1mm	3mm
Gap:	0mm	0mm	0mm	0mm	0mm	0mm	0mm	0mm	0mm
Weighting:	T2	T2	T2	T1	FLASH	FLASH	FLASH	FLASH	T1
Mode:	HASTE BH	TRUFI BH	TRUFI BH	VIBE DIXON	3D BH	3D DYNAMIC BH PHASE 1	3D DYNAMIC BH PHASE 2	3D DYNAMIC BH PHASE 3	VIBE DIXON
FAT SAT:	NO	NO	NO	YES	YES	YES	YES	YES	YES
FOV:	Fit to Patient	Fit to Patient	Fit to Patient	Fit to Patient	Fit to Patient	Fit to Patient	Fit to Patient	Fit to Patient	Fit to Patient
MPR:	NO	NO	NO	NO	NO	NO	NO	NO	NO
Notes:									