

MRI Abdomen WO						
Technical Parameters						
Anatomic Coverage:	Fit to Patient					
Coil:	Torso					
Scanner Preference:	1.5T					
Scanning Plane:	Cor Abd	Axial Abd	Axial Abd	Axial Abd	Axial Abd	Axial Abd
Slice Thickness:	7mm	7mm	7mm	5mm	7mm	5mm
Gap:	1mm	1mm	1mm	1mm	1mm	1mm
Weighting:	T2	T1	T2	T2	T2	T1
Mode:	SS BH	Dual Echo SPGR BH	DWI	S5	TSE RT	3D mDixon pre BH
FAT SAT:	NO	NO	SPIR	NO	SPAIR	YES
FOV:	Fit to Patient	Fit to Patient	Fit to Patient	20-24mm	Fit to Patient	Fit to Patient
MPR:	NO	NO	NO	NO	NO	NO
Notes:						