

MRI Abdomen/Pelvis WO								
Technical Parameters								
Anatomic Coverage:	Entire liver through perineum							
Coil:	Torso							
Scanner Preference:	1.5T							
Scanning Plane:	Cor Abd/Pel	Ax Abd/Pel	Ax Abd	Ax Pel	Ax Abd/Pel	Ax Abd/Pel	Ax Pel	Ax Abd
Slice Thickness:	7mm	7mm	7mm	7mm	5mm	7mm	7mm	7mm
Gap:	1mm	1mm	1mm	1mm	1mm	1mm	1mm	1mm
Weighting:	T2	T1	T2	T2	T2	T2	T1	T1
Mode:	SS BH	Dual Echo SPGR BH	DWI	DWI	SS	TSE RT	3D THRIVE BH	3D mDixon PRE BH
FAT SAT:	NO	NO	SPIR	SPIR	NO	SPAIR	YES	YES
FOV:	Fit to patient	Fit to patient	Fit to patient	Fit to patient	20-24mm	Fit to patient	Fit to patient	Fit to patient
MPR:	NO	NO	NO	NO	NO	NO	NO	NO
Notes:								