MRI Adrenal W & WO								
Technical Parameters								
Anatomic Coverage:	Diaphragm to Aortic bifurcation							
Coil:	Torso							
Scanner Preference:	1.5T							
Contrast:	Pre-injection:				Post-injection: trigger when bolus reaches SMA			
Scanning Plane:	Coronal	Axial	Coronal	Axial	Axial	Coronal	Axial	Axial
Slice Thickness:	5mm	3mm	3mm	3mm	5mm	5mm	5mm	7mm
Gap:	0.5mm	0.5mm	0.5mm	0.5mm	1mm	1mm	1mm	1mm
Weighting:	T2	T1	T1	T1	T1	T1	T2	T2
Mode:	SSFSE BH	Dual Echo SPGR BH	Dual Echo SPGR BH	3D mDixon pre BH	3D mDixon x 3 phases BH	3D mDixon post	TSE FS	DWI
FAT SAT:	NO	NO	NO	YES	YES	YES	YES	SPIR
FOV:	Diaphragm to Aortic bifurcation	Fit to Patient	Diaphragm to Aortic bifurcation	Fit to Patient	Fit to Patient	Diaphragm to Aortic Bifurcation	Fit to Patient	Fit to Patient
MPR:	NO	NO	NO	NO	NO	NO	NO	NO
Notes:								