MRI Liver Mass W & WO									
Technical Parameters									
Anatomic Coverage:	Diaphragm to aortic bifurcation								
Coil:	Torso								
Scanner Preference:	1.5T or 3T								
Contrast:	Pre-injection:					Post-injection: trigger when bolus reaches SMA			
Scanning Plane:	Coronal	Axial	Axial	Axial	Axial	Axial	Axial	Coronal	Axial
Slice Thickness:	7mm	5mm	5mm	7mm	5mm	5mm	5mm	5mm	5mm
Gap:	1mm	0.5mm	0.5mm	1mm	0.5mm	0.5mm	0.5mm	0.5mm	0.5mm
Weighting:	T2	T1	T1	T2	T2	T2	T1	T1	T1
Mode:	SSFSE BH	Dual Echo SPGH BH	3D mDixon pre BH	DWI	SSFSE RT	TSE RT	3D mDixon x3 phases BH	3D mDixon 3 min post	3D mDixon BH
FAT SAT:	NO	NO	YES	NO	YES	SPAIR	YES	YES	YES
FOV:	Liver through bottom of kidneys	Entire Liver	Entire Liver	Liver through bottom of kidneys	Entire Liver	Entire Liver	Entire Liver	Diaphragm to Aortic bifurcation	Entire Liver
MPR:	NO	NO	NO	NO	NO	NO	NO	NO	NO
Notes:									